



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**



**SUPPLEMENTAL/ BID BULLETIN NO. 2**

**IB No. 2025-077**

**Supply and Delivery of Multi-use Temperature Data Logger with Display (Rebid)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Query during Pre-bidding Conference:		
Technical Specification	Query	Response of the End User Unit
Sensor measurement range: -40 °C to + 99 °C (-40 °F to 210 °F)	Sensor measurement range: - 30°C to + 60 °C (-22 °F to 140 °F)	Sensor measurement range: - 30 °C to + 60 °C (-22 °F to 140 °F) GRANTED
Delivery Period: 30 calendar days	Delivery Period: 30 -60 calendar days	Delivery Period: 30 -60 calendar days GRANTED

Bidders are advised to use the **attached forms and submit them together with all required documents for the submission of bids on the 20<sup>th</sup> day of May 2025, 9:00 AM, Amphitheater.**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 14<sup>th</sup> day of May 2025 in MMCHD

Approved by:

SGD.  
**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
Licensing Officer V / BAC Chairperson

<p style="text-align: center;">           Republic of the Philippines            Department of Health            Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b> </p>			
Item No. 1	Multi-use Temperature Data Logger with Display	Qty./Unit	341 pcs
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
<b>ABC:</b> P 1,534,500.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><b>A. Specifications:</b></p> <p>Battery: Manufacturer's standard with minimum operating life of 2 years with or without detachable probe connected to data logger</p> <p><i>Sensor measurement range: -30<sup>0</sup>C to + 60<sup>0</sup>C (-22 <sup>0</sup>F to 140 <sup>0</sup>F)</i></p> <p>Temperature reading accuracy: Better than ±0.5°C</p> <p>Rated Temperature Reading Resolution:0.1C</p> <p>Recording Capacity: Minimum 16,000 temperature readings</p> <p>Sampling Interval: Configurable from 30 seconds to 18 hours</p> <p>Connection Interface: USB 2.0, A-type plug</p> <p>Alarms: User Programmable upper and lower alarm temperatures with built-in-real time clock on display</p> <p>Warranty: Minimum 1 year</p> <p><i>Certifications: Calibration certificate and/or WHO-PQS Certification</i></p> <p>Provide software on CD or USB Stick:</p> <p>Delivery Period: 30 calendar days</p> <p>Place of Delivery: DOH MMCHD Pasig Warehouse</p>			
<p>B. Upon delivery, the following shall be complied with:</p> <p>1. Labeling instructions:</p> <p>1. On each small and bigger box/carton, the following should be legible imprinted or stickered with non-removable sticker or permanent sticker or label that is binding and will leave residue and ripping if removed.</p>			

<p align="center"> “Philippine Government Property-Department of Health” NOT FOR SALE </p>	
<p><b>D. Additional requirement by the Lowest/Single Calculated Bid (L/SCB) as part of post qualification:</b></p> <ol style="list-style-type: none"> <li>1. LCB / SCB must provide two (2) samples of actual unit for approval by the End-user as part of the post-qualification and prior to award. If approved by End-user, one (1) sample will be retained for reference of the units to be delivered.</li> <li>2. L/SCRB shall pick up the Contract and Notice to Proceed issued in its favor within three (3) calendar days from receipt of the notice. An electronic mail shall constitute an official notice to the Bidder.</li> <li>3. Refusal to sign and accept the Award or enter into a contract without justifiable reason may be grounds for imposing administrative sanctions under Rule XXIII of the Revised IRR of RA 9184.</li> <li>4. The registered company name and email address must be consistent and should reflect on all documents to be submitted.</li> <li>5. Request for extension should be submitted before the lapse of the original delivery date. The maximum allowable extension shall not be longer than the Original Delivery term.</li> <li>6. Delivery through courier service is not allowed.</li> </ol>	
<p><b>E. Product Recall &amp; Replacement:</b></p> <ol style="list-style-type: none"> <li>1. The supplier must ensure the quality of products. If there are problems in the quality, the Supplier will recall and replace the products distributed in the regions hospitals/treatment hubs/RHU/HC/BHSs based on Guidelines on Product Recall, FDA Circular No. 2016-012.</li> <li>2. In case of the product recalls, damage or expired medicines for replacement, the costs associated with the proper handling or pull out from health facilities where the medicines have already been distributed shall be borne by the Supplier.</li> </ol>	