



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 026E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of Brand New Various Medical Equipment for the CITY OF PARANAQUE.**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:


Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>ITEM 1</b>	
<b>2 Unit CRYOTHERAPY MACHINE</b>	
<b>FROM</b>	<b>TO</b>
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 2</b>	
<b>42 Unit DIGITAL STETHOSCOPE</b>	
<b>*No changes as stated in technical specifications</b>	
<b>ITEM 3</b>	
<b>31 Unit FETAL DOPPLER</b>	
<b>FROM</b>	<b>TO</b>
Power Supply: 220 VAC, 60Hz	Power Supply: Manufacturer's Standard
<b>ITEM 4</b>	
<b>1 Unit NEWBORN HEARING SCREENING MACHINE</b>	
<b>*No changes as stated in technical specifications</b>	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23 August 2021, 9:00 AM:


This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.  
For guidance and information of all concerned

Issued this 13<sup>th</sup> day of August, 2021 in MMCHD.

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
BAC Chairperson





**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 027E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of Brand New Various Medical Equipment for JUSTICE JOSE ABAD SANTOS GENERAL HOSPITAL**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and Clarification to provisions/specifications in the Bidding Documents:	
<b>ITEM 1</b>	
<b>1 Unit PHOTOTHERAPY LAMP</b>	
<b>FROM</b>	<b>TO</b>
Power supply: 230VAC, 50-60 Hz	Power Supply: 220V+-10
<b>ITEM 2</b>	
<b>2 Units WEIGHING SCALE WITH HEIGHT AND WEIGHT</b>	
<b>*No changes as stated in technical specifications</b>	
<b>ITEM 3</b>	
<b>4 Units INFANT INCUBATOR</b>	
<b>FROM</b>	<b>TO</b>
Air temperature control range: 23.0~37.0 ± 0.3 °C (Override >39.0 °C)	Air temperature control range equivalent to Set Point Range Set Point Range: 22.0 ~ 38.0 ±
Skin temperature control range: 35.0~37.0 ± 0.3 °C (Override >39.0 °C)	Skin temperature control riange equivalent to Set Point Range
Control system: Ultrasonic & steam	Control system: Ultrasonic & steam or equivalent
<b>ITEM 4</b>	
<b>2 Units OPERATING ROOM TABLE</b>	
<b>FROM</b>	<b>TO</b>
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30 – 60 Calendar Days
<b>ITEM 5</b>	
<b>2 Units BLOOD GAS ANALYZER</b>	
<b>*No changes as stated in technical specifications</b>	
<b>ITEM 6</b>	
<b>2 Units MOBILE OR LIGHT</b>	
<b>*No changes as stated in technical specifications</b>	
<b>ITEM 7</b>	
<b>2 Units AUTOCLAVE STERILIZING MACHINE</b>	
<b>*No changes as stated in technical specifications</b>	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23 August 2021, 9:00 AM:


This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned



Issued this 13<sup>th</sup> day of August, 2021 in MMCHD.

Approved by:



**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
BAC Chairperson





Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 028E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of Brand New Various Medical Equipment for the use of Health Facilities in MARIKINA CITY**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revisions and clarification to provisions/specifications in the Bidding Documents:	
<b>ITEM 1</b> <b>1 Unit Bed, Plain with Mattress and IV Stand</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 2</b> <b>1 Unit REAGENT REFRIGERATOR</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 3</b> <b>1 Unit X RAY VERTICAL BUCKY STAND</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 4</b> <b>1 Unit MINOR SURGERY SET FOR NSV</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 5</b> <b>1 Unit X RAY FILM PROCESSOR</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 6</b> <b>2 Units WEIGHING SCALE WITH HEIGHT MEASURING STICK</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 7</b> <b>2 Units MEDICAL INSTRUMENT CABINET</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 8</b> <b>1 Unit INSTRUMENT TABLE</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 9</b> <b>2 Unit EXAMINING TABLE</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 10</b> <b>1 Unit DIGITAL ECG WITH INTERPRETATION</b>	
<b>FROM</b>	<b>TO</b>
<i>ECG roll paper in 63 millimeter width</i>	<i>ECG roll paper: manufacturer's standard</i>
<i>Frequency response: 0.05 to 150 Hz within -3dB to +0.5 Db</i>	<i>Frequency response: Manufacturer's standard</i>
<i>Operating Pressure: 700 to 1060 hPa</i>	<i>Operating Pressure: Manufacturer's Standard</i>
<b>ITEM 11</b>	



<b>1 Unit CASSETTE WITH INTENSIFYING SCREEN 14x14</b>	
<b>FROM</b>	<b>TO</b>
<b>Standard features:</b>	<del>Standard features:</del> (Removed)
<b>Shock proof</b>	<del>Shock proof</del> (Removed)
<b>Light proof</b>	<del>Light proof</del> (Removed)
<b>Easy to load and unload</b>	<del>Easy to load and unload</del> (Removed)
<b>ITEM 12</b>	
<b>1 Unit CASSETTE WITHINTENSIFYING SCREEN 14x17</b>	
<b>FROM</b>	<b>TO</b>
<b>Standard features:</b>	<del>Standard features:</del> (Removed)
<b>Shock proof</b>	<del>Shock proof</del> (Removed)
<b>Light proof</b>	<del>Light proof</del> (Removed)
<b>Easy to load and unload</b>	<del>Easy to load and unload</del> (Removed)
<b>ITEM 13</b>	
<b>1 Unit CASSETTE WITH INTENSIFYING SCREEN 11x14</b>	
<b>FROM</b>	<b>TO</b>
<b>Standard features:</b>	<del>Standard features:</del> (Removed)
<b>Shock proof</b>	<del>Shock proof</del> (Removed)
<b>Light proof</b>	<del>Light proof</del> (Removed)
<b>Easy to load and unload</b>	<del>Easy to load and unload</del> (Removed)
<b>ITEM 14</b>	
<b>1 Unit BIOSAFETY CABINET</b>	
<b>*No changes as stated in technical specifications</b>	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23 August 2021, 9:00 AM:  
This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.  
For guidance and information of all concerned

Issued this 13<sup>th</sup> day of August, 2021 in MMCHD.

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
BAC Chairperson





Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 029E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of Brand New Various Medical Equipment for the use of Health Facilities in CITY OF VALENZUELA**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>ITEM 1</b>	
<b>1 Unit CERVICAL AND LUMBAR TRACTION WITH BED</b>	
<b>FROM</b>	<b>TO</b>
Traction Force: 2.5kg/5.5lb (minimum) 91kg/200lb (maximum)	To: Traction Force: Maximum of 100kg
Traction Speed: Stepless, continuous, and adjustable	Traction Speed: Manufacturer's standard
Therapy Mode: Continuous, Harmonic, Intermittent, Progressive	Therapy Mode: Manufacturer's standards
Power Consumption: 150VA max	Power Consumption: Manufacturer's standard
<b>ITEM 2</b>	
<b>10 Set EENT DIAGNOSTIC SET</b>	
<b>*No changes as stated in technical specifications</b>	
<b>ITEM 3</b>	
<b>7 Units PULSE OXIMETER WITH ADULT AND PEDIA PROBE</b>	
<b>FROM</b>	<b>TO</b>
Battery: Type 7.2V, 2,000Mah, Ni-MH Operation time: 4-8 hours Charging Time: 200 minutes	Battery Type: Manufacturer's standard Operation time: Minimum of 4 hours Charging Time: Manufacturer's standard
Safety Standard: EN60601-1	Safety Standard: EN60601-1 or its equivalent
Output Interface: Serial RS-232	Output interface: Manufacturer's standard
<b>ITEM 4</b>	
<b>1 Unit THERAPEUTIC ULTRASOUND MACHINE</b>	
<b>*No changes as stated in technical specifications</b>	
<b>ITEM 5</b>	
<b>4 Units WHEELED DETACHABLE STRETCHER</b>	
<b>*No changes as stated in technical specifications</b>	
<b>ITEM 6</b>	
<b>1 Unit DIGITAL HOT PACK HEATER</b>	
<b>*No changes as stated in technical specifications</b>	
<b>ITEM 7</b>	
<b>1 Unit DIGITAL PARAFFIN WAX BATH</b>	
<b>FROM</b>	<b>TO</b>
Indication of Button for time: Control the time (1-24)	Indication of Button for time: Manufacturer's Standard



Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23 August 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 13<sup>th</sup> day of August, 2021 in MMCHD.

Approved by:

A handwritten signature in blue ink, appearing to read 'J. Chan', is written above the printed name.

**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
BAC Chairperson





Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 030E**

**Procurement of Proposed Repair and Maintenance of HFEPMU Office and LHSD Proper Office  
(Lot Bidding) (Rebid)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Clarification to provisions/specifications in the Bidding Documents:
--

<b>*No changes as stated in technical specifications</b>
--


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23 August 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 13<sup>th</sup> day of August, 2021 in MMCHD.

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
BAC Chairperson





Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 - 031E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of Brand New 10 Set  
Emergency Cart (E-Cart) for the NOVALICHES DISTRICT HOSPITAL**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Clarification to provisions/specifications in the Bidding Documents:
--

<b>*No changes as stated in technical specifications</b>
--


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23 August 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.


For guidance and information of all concerned

Issued this 13<sup>th</sup> day of August, 2021 in MMCHD.

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
BAC Chairperson





Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 032E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipments for CALOOCAN CITY MEDICAL CENTER CCMC Under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for Health Development(Rebid)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and Clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
<b>ITEM A</b> 1 UNIT DRY BATH	<b>ITEM 4</b> 1 UNIT DRY BATH
<b>ITEM 4</b> 1 UNIT AGGLUTINATION VIEWER	<b>ITEM 5</b> 1 UNIT AGGLUTINATION VIEWER
<b>ITEM 1</b> <b>1 UNIT COLONOSCOPE</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 2</b> <b>1 UNIT BLOOD BANK REFRIGERATOR</b>	
FROM	TO
Precise temperature control system: LED screen with 4°C (±1°C) adjustable platinum resistance temperature sensor Read Ability: 0.1°C	Precise temperature control system: in range of measurement
<b>ITEM 3</b> <b>3 UNITS RADIANT WARMER WITH PHOTOTHERAPY</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 4</b> <b>1 UNIT DRY BATH</b> <b>*No changes as stated in technical specifications</b>	
<b>ITEM 5</b> <b>1 UNIT AGGLUTINATION VIEWER</b> <b>*No changes as stated in technical specifications</b>	


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23 August 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 13<sup>th</sup> day of August, 2021 in MMCHD.

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
BAC Chairperson





Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 033E**

**Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipments for VALENZUELA CITY EMERGENCY HOSPITAL Under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for Health Development (Rebid)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Clarification and Revision to provisions/specifications in the Bidding Documents:	
<b>ITEM 1</b>	
<b>1 UNIT HEPA FILTER MACHINE</b>	
<b>FROM</b>	<b>TO</b>
Schedule of Requirements: Delivery Schedule: 15-30 Calendar Days	Schedule of Requirements: Delivery Schedule: 30-60 Calendar Days
<b>ITEM 2</b>	
<b>1 UNIT MORTUARY REFRIGERATOR (4 CORPSE)</b>	
<b>FROM</b>	<b>TO</b>
Schedule of Requirements: Delivery Schedule: 15-30 Calendar Days	Schedule of Requirements: Delivery Schedule: 30-60 Calendar Days
<b>ITEM 3</b>	
<b>1 UNIT SLIT LAMP</b>	
<b>FROM</b>	<b>TO</b>
Schedule of Requirements: Delivery Schedule: 15-30 Calendar Days	Schedule of Requirements: Delivery Schedule: 30-60 Calendar Days
<b>ITEM 4</b>	
<b>3 UNITS PORTABLE TYPE STERILZER</b>	
<b>FROM</b>	<b>TO</b>
Schedule of Requirements: Delivery Schedule: 15-30 Calendar Days	Schedule of Requirements: Delivery Schedule: 30-60 Calendar Days


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23 August 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 13<sup>th</sup> day of August, 2021 in MMCHD.

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
BAC Chairperson





Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 034E**

**Procurement of various Medical Equipments for use of OSPITAL NG TONDO under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for Health Development (Rebid)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and Clarification to provisions/specifications in the Bidding Documents:	
<b>ITEM 1</b> <b>2 Units AUTOCLAVE</b>	
<b>FROM</b>	<b>TO</b>
Schedule of Requirements:	Schedule of Requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 2</b> <b>1 Unit BINOCULAR MICROSCOPE</b>	
<b>FROM</b>	<b>TO</b>
Schedule of Requirements:	Schedule of Requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 3</b> <b>4 UNITS FETAL DOPPLER</b>	
<b>FROM</b>	<b>TO</b>
Schedule of requirements:	Schedule of requirements:
4 UNITS FETAL DOPPLER ABC: P 45,000.00	5 UNITS FETAL DOPPLER ABC: P 225,000.00
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 4</b> <b>1 UNIT FETAL MONITOR</b>	
<b>FROM</b>	<b>TO</b>
Schedule of requirements:	Schedule of requirements:
5 UNITS FETAL MONITOR ABC: 225,000.00	1 UNIT FETAL MONITOR ABC: P 160,000.00
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 5</b> <b>1 SET DENTAL INSTRUMENT SET</b>	
<b>FROM</b>	<b>TO</b>
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 6</b> <b>4 UNIT INFANT WEIGHING SCALE, DIGITAL</b>	
<b>FROM</b>	<b>TO</b>



Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 7</b> <b>10 UNITS MINOR SURGICAL SET</b>	
<b>FROM</b>	<b>TO</b>
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 8</b> <b>2 UNITS PULSE OXIMETER</b>	
<b>FROM</b>	<b>TO</b>
Battery: Type 7.2V. 2,000Mah, Ni-MH Operation time: 4-8 hours Charging Time: 200 minutes	Battery Type: Manufacturer's standard Operation time: Minimum of 4 hours Charging Time: Manufacturer's standard
Safety Standard: EN60601-1	Safety Standard: EN60601-1 or its equivalent
Output Interface: Serial RS-232	Output interface: Manufacturer's standard
Conformity: MDD 93/42/EEC, CE0470	Conformity: Manufacturer's standard
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 9</b> <b>1 UNIT REAGENT REFRIGERATOR</b>	
<b>FROM</b>	<b>TO</b>
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 10</b> <b>5 UNITS SUCTION MACHINE HEAVYDUTY</b>	
<b>FROM</b>	<b>TO</b>
Noise: not more than 55 dB	Noise: Manufacturer's standard
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 11</b> <b>8 UNITS WEIGHING SCALE, ADULT WITH HEIGHT MEASURING STICK</b>	
<b>FROM</b>	<b>TO</b>
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 12</b> <b>3 UNITS ECG WITH ANALYSIS</b>	
<b>FROM</b>	<b>TO</b>
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
Communication: PC communication with RS232 interface and LAN	Communication: Manufacturer's standard
<b>ITEM 13</b> <b>10 UNITS PORTABLE UV LIGHT STERILIZER</b>	



FROM	TO
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 14</b> <b>12 UNIT STERILIZING UNIT</b>	
FROM	TO
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 15</b> <b>1 UNIT DENTAL UNIT MOTORIZED WITH ACCESSORIES</b>	
FROM	TO
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 16</b> <b>1 UNIT DENTAL AEROSOL SUCTION MACHINE</b>	
FROM	TO
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 17</b> <b>1 UNIT DENTAL INSTRUMENT CABINET</b>	
FROM	TO
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 18</b> <b>1 UNIT DENTAL ULTRASONIC SCALER MACHINE</b>	
FROM	TO
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days
<b>ITEM 19</b> <b>6 UNITS POWERED AIR PURIFYING RESPIRATOR</b>	
FROM	TO
Schedule of requirements:	Schedule of requirements:
Delivery Schedule: 15-30 Calendar Days	Delivery Schedule: 30-60 Calendar Days

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23 August 2021, 9:00 AM:


This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned  
Issued this 13<sup>th</sup> day of August, 2021 in MMCHD.

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
BAC Chairperson





Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2021 – 035E**

**Procurement of various Medical Equipments for GAT ANDRES BONIFACIO MEMORIAL  
MEDICAL CENTER Under the Health Facilities Enhancement Program (HFEP) of Metro  
Manila Center for Health Development (Rebid)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>ITEM 1</b>	
<b>1 UNIT AUTOCLAVE</b>	
<b>*No changes as stated in technical specifications</b>	
<b>ITEM 2</b>	
<b>1 UNIT ECG MACHINE</b>	
<b>FROM</b>	<b>TO</b>
Power Supply: Internal Rechargeable Ni-Mh batteries, 2Ah	Power Supply: Manufacturer's standard
Conformity Standard: 93/42/CEE; EN60601-1; EN60601-1-2; IEC601-2-25	Conformity Standard: 93/42/CEE; EN60601-1; EN60601-1-2; IEC601-2-25 or its equivalent
<b>ITEM 3</b>	
<b>1 UNIT FETAL DOPPLER</b>	
<b>*No changes as stated in technical specifications</b>	
<b>ITEM 4</b>	
<b>3 UNITS INCUBATOR</b>	
<b>FROM</b>	<b>TO</b>
Air Temperature control range: 23.0~37.0 ± 0.3°C (Override >39.0 ± 0.3°C)	Air temperature control range equivalent to Set Point Range Set Point Range: 22.0 ~ 38.0 ±
Display panel: Minimum of At least 7 inches TFT Color LCD	Display panel: 4 inches to 7 inches TFT Color LCD


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23 August 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 13<sup>th</sup> day of August, 2021 in MMCHD.

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
BAC Chairperson