# AATMENT SHIP

### Republic of the Philippines Department of Health

### METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

### IB 2022 – 038E PROCUREMENT OF VARIOUS LABORATORY EQUIPMENT (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 1 – 1 UNIT MICROHEMATOCRIT CENTRIFUGE	
*No changes as stated in technical specifications.	
Item No. 2 – 1 UNIT URINE ANALYZER	
*No changes as stated in technical specifications.	
Item No. 3 + 1 UNIT DIFFERENTIAL COUNTER	
*No changes as stated in technical specifications.	
Item No. 4 – 2 UNITS CENTRIFUGE	
*No changes as stated in technical specifications.	
Item No. 5 – 10 SETS CLINICAL CENTRIFUGE 12-PLACER	
*No changes as stated in technical specifications.	
Item No. 6 – 1 UNIT CLINICAL ROTATOR	
Recipient:	
Health Facilities in Paranaque City	
Item No. 7 – 2 SETS COLORIMETER	
*No changes as stated in technical specifications.	
Item No. 8 – 1 UNIT MICROSCOPE BINOCULAR TYPE	
ANNUAL TEXT OF STATE STA	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 18th day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

\*No changes as stated in technical specifications.

Issued this 9th day of April, 2022 in MMCHD

Approved by:

ALELI ANNE GRACE P. SUDIACAL, MD, MPH

Director III / BAC Chairperson



Republic of the Philippines				
Department of Health				
Metro Manila Center for Health Development				
	TECHNICAL SPECIFI	CATIONS		
Item No. 1	MICROHEMATOCRIT CENTRIFUGE	Qty./Unit	1 Unit	
Name of Ma	nufacturer:		Country of Origin	
Brand:			Model: (if applicable)	
ABC: <b>48,00</b> 0	0.00			
	PURCHASER'S SPECIFICATION	1	STATEMENT OF	
			COMPLIANCE	
A. Technical	Specifications:			
• Design: T	able Top			
• Speed: 12	2,000 rpm			
• RCF: At le	east 14,000 x g			
• Radius of	Rotor: At least 9 cm			
• Timer: 0-15 mins				
• Rotor: Universal Rotor				
• Capacity: 24 capillary tubes (1.5x75mm)				
• Power Su	pply: 220V, 60Hz			
• Standard	Accessories			
- Percenta	ge reader			
- Disc read	er			
- Capillary	tubes			
<del>-</del>	EMENTS IF AWARDED TH	<u>E</u>		
CONTRACT	<u>:</u>			
1. <b>Completion Period</b> : The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.				
2. <b>Testing</b> : Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and				



WENT	
must have no physical damage and defect.	
3. <b>Training</b> : The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment	
4. <b>Warranty</b> : Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.	
5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.	
6. <b>Manuals</b> : The supplier must provide the enduser one (1) hard and one (1) soft copy of the following:	
a) Service manual in English language	
b) Operation manual in English language	
7. With "DOH-MMCHD HFEP"(Government	
Property not for sale) sticker in each unit.	
X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-	
Recipient:	
Health Facilities of Paranaque City	
C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:  1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.	



	Republic of the Philippines			
Department of Health				
Metro Manila Center for Health Development				
	TECHNICAL SPECIFICA	ATIONS		
Item No. 2	URINE ANALYZER	Qty./Unit	1 Unit	
Name of Ma	nufacturer:		Country of Origin	
Brand:			Model: (if applicable)	
ABC: <b>50,00</b> 0	0.00			
	PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technica	l Specifications:			
Type				
- Bench	top			
Features				
- Employs reflectance photometry				
- Automatic self-calibration				
- Compatible with various urine strips in 4, 5 or 10 parameters available in the market.				
- Strip i	ncubation time: 1 min			
- LCD D	isplay or better			
- Built-i	n printer			
Test Mode	s			
- Routir	ne			
- STAT				
- QC				
Throughp	ut			
- 60-12	0 tests/hr			
Memory				
- can ho	old at least 1,000 results			
Power Sup	pply			
- 220 VAC,	60 Hz			
			i .	



MENT OF THE	
Accessories	
- Automatic Voltage Regulator	
- Ten parameter urine strips as starter kits	
- At least 2 rolls of thermal paper	
- Dust cover	
B. REQUIREMENTS IF AWARDED THE CONTRACT:	
1. <b>Completion Period</b> : The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.	
2. <b>Testing</b> : Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.	
3. <b>Training</b> : The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment	
4. <b>Warranty</b> : Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.	
5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.	
6. <b>Manuals</b> : The supplier must provide the enduser one (1) hard and one (1) soft copy of the following:	
a) Service manual in English language	



b) Operation manual in English language	
7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.	
X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X	
Recipient:	
Health Facilities of Paranaque City	
c. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:	
1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.	



	Republic of the Phi	lippines		
Department of Health				
Metro Manila Center for Health Development				
	TECHNICAL SPECIF	ICATIONS		
Item No. 3	DIFFERENTIAL COUNTER	Qty./Unit	1 Unit	
Name of Ma	nufacturer:		Country of Origin	
Brand:			Model: (if	
ADC: 15 000	0.00		applicable)	
ABC: <b>15,00</b> 0				
	PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technica	l Specifications:			
<ul> <li>9 3-digitation</li> <li>total amount</li> </ul>	t LCD windows (8 for cells a nt)	nd 1 for		
• 4 function	n buttons (+/-, %, RESET, FU	N)		
- +/- bւ	utton to increase or decrease	the count		
- Percentage button for any-time percentage calculations				
- RESET to clear all counts				
- FUN to adjust the contrast and monitor battery energy				
• Power: 220VAC, 50/60Hz				
• Frequen	cy: 28kHz ± 3kHz			
Can oper	rate on battery power (2 AA	batteries)		
•	EMENTS IF AWARDED THE	<u> </u>		
CONTRACT	<u>':</u>			
testing and its accessor and mainte	cion Period: The delivery, in commissioning of the equip ries, including the training of mance staff must be complet r days upon receipt of Notice	ment and f end-users red within		
2. <b>Testing</b> : Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and				





Republic of the Philippines				
Department of Health				
Metro Manila Center for Health Development				
	TECHNICAL SPECIF	ICATIONS		
Item No. 4	CENTRIFUGE	Qty./Unit	2 Units	
Name of Ma	nufacturer:		Country of Origin	
Brand:			Model: (if applicable)	
ABC: <b>280,0</b> 0	00.00			
	PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technical	Specifications:			
• Speed (R	RPM): 500-4000 RPM			
<ul><li>Maximum</li></ul>	m RCF(g): 2500 x g			
• Rotor: 24	4 place fixed			
• Tube size	e: 2-15ml			
● Motor: DC				
● Power Supply: 220V, 60Hz				
● Size: Manufacturer's Standard				
• Weight:	Manufacturer's Standard			
<ul><li>Variable</li></ul>	Speed: 1000-3400 RPM			
• 24 Place	r fixed angle motor			
● 60 min d	igital timer			
<ul><li>Capacity</li></ul>	24x10ml tubes or 16x15ml	tubes		
B. <u>REQUIR</u> CONTRACT	EMENTS IF AWARDED THE	]		
	_			
testing and its accessor and mainte	cion Period: The delivery, in commissioning of the equip ries, including the training of mance staff must be complet r days upon receipt of Notice	ment and f end-users red within		
2. <b>Testing</b> : Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.				



3. <b>Training</b> : The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment	
4. <b>Warranty</b> : Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.	
5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.	
6. <b>Manuals</b> : The supplier must provide the enduser one (1) hard and one (1) soft copy of the following:	
a) Service manual in English language	
b) Operation manual in English language	
7. With " <b>DOH-MMCHD HFEP</b> " (Government Property not for sale) sticker in each unit.	
X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-	
Recipient:	
Health Facilities of Paranaque City	
c. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:  1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.	



Republic of the Philippines				
Department of Health				
Metro Manila Center for Health Development				
	TECHNICAL SPECIF	ICATIONS		
Item No. 5	CLINICAL CENTRIFUGE 12-PLACER	Qty./Unit	10 Sets	
Name of Mar	nufacturer:	1	Country of Origin	
Brand:			Model: (if applicable)	
ABC: <b>440,0</b> 0	00.00			
	PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technical	Specifications:			
Rotor Capa	city: 8 x 15 mL, 12 x 3-12 m	L		
Rotor Type	: Fixed Angle			
Max Rotor	Speed: 15,000 rpm			
Max RCF: 2490 x g				
Run Time: 30 sec to 99 min / Continuous				
Driving Motor: Brushless DC Motor				
Display: Digital, LCD or better				
Noise Level: ≤ 60 dB				
Dimensions	s: Manufacturer's Standard			
Power Supp	oly: 220 VAC, 60 Hz			
Safety Feat	ures:			
Inclusions:				
- Rotor for	8 x 15 mL			
- Adapters for all tube sizes 3-12 mL				
B. <u>REQUIRI</u> <u>CONTRACT</u>	EMENTS IF AWARDED THI :	<u>:</u>		
testing and its accessor and mainte	ion Period: The delivery, in commissioning of the equipries, including the training on nance staff must be completed and suppose the completer days upon receipt of Notice.	oment and f end-users ted within		



2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.	
3. <b>Warranty</b> : Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.	
4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.	
5. <b>Manuals</b> : The supplier must provide the enduser one (1) hard and one (1) soft copy of the following:	
a) Service manual in English language	
b) Operation manual in English language	
6. With " <b>DOH-MMCHD HFEP</b> " (Government Property not for sale) sticker in each unit.	
X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-	
Recipient:	
Health Facilities of Paranaque City	
B. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:	
1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.	



	Republic of the Ph	ilippines		
Department of Health				
Metro Manila Center for Health Developm			ient	
	TECHNICAL SPECIF	ICATIONS		
Item No. 6	CLINICAL ROTATOR	Qty./Unit	1 Unit	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: <b>180,0</b> 0	00.00		аррисавіе	
	PURCHASER'S SPECIFICATION		STATEMENT OF	
	PURCHASER 3 SPECIFICATION		COMPLIANCE	
A. Technical	Specifications:			
• Load Capa	acity: At least 10 lbs			
• Controls:	Digital			
• Display: L	.CD or better			
• Dimensio	ns: Manufacturer's Standard	l		
• Min Speed	d: 30 RPM			
• Max Spee	d: 220 RPM			
• Orbital Di	ameter: At least 19 mm			
• Power Su	pply: 220V, 50/60Hz			
• Material:	Stainless Steel exterior			
B. <u>REQUIRE</u>	MENTS IF AWARDED THE CO	NTRACT:		
1. Complet	ion Period: The delivery, in	stallation,		
testing and	commissioning of the equip	ment and		
	ries, including the training o			
	nance staff must be complet r days upon receipt of Notic			
Proceed.	i days upon receipt of Notic			
2 Tacting	Prior to accentance the end	lucar chall		
2. <b>Testing</b> ; Prior to acceptance, the end user shall conduct a physical inspection and functionality				
test. The equipment must be functioning and				
must have	no physical damage and defe	ect.		
3. Training	<b>g</b> : The supplier shall provide	e a training		



on the proper use and maintenance of the equipment to the end-users and to the hospital	
maintenance staff	
4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.	
5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.	
6. <b>Manuals</b> : The supplier must provide the enduser one (1) hard and one (1) soft copy of the following:	
a) Service manual in English language	
b) Operation manual in English language	
7. With " <b>DOH-MMCHD HFEP</b> " (Government Property not for sale) sticker in each unit.	
<b>Recipient:</b> Health Facilities in Paranaque City	
d. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:	
1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.	



Republic of the Philippines				
Department of Health				
Metro Manila Center for Health Developn			nent	
TECHNICAL SPECIFICATIONS				
Item No. 7	COLORIMETER	Qty./Unit	2 Sets	
Name of Manufacturer:			Country of Origin	
Brand:		Model: (if applicable)		
ABC: <b>360,0</b> 0	00.00			
	PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technica	l Specifications:			
• Detector:	Silicon Photodiode			
• Data Logg	ger: At least 500 measured v	alues		
• Display: L	ED			
• Dimensio	ns: Manufacturer's Standard	ł		
• Enclosure	e Rating: At least IP67			
• Optical Sy	stem: 0/180 deg transmitta	nce		
• Photomet	ric Accuracy: At least ± 0.00	5 Abs		
• Photomet	ric Linearity: At least ± 0.00	2 Abs		
• Photomet	cric Measuring Range: At lea	st 0-2 Abs		
• Sample Co	ell Compatibility: 1 inch rou	nd		
• Power Su charger	pply: Rechargeable Battery	with		
• Inclusions	s:			
- Recharge	able batteries with charger			
- At least 2 10, 20, 25 r	glass sample cells 1-inch ro nL marks	und with		
- At least 2	plastic sample cells 1cm/10	)mL		
- Sample cell adapter (if necessary)				
B. REQUIRE	EMENTS IF AWARDED THE	<u> </u>		

# TATMENT OF HER

### Republic of the Philippines Department of Health

### METRO MANILA CENTER FOR HEALTH DEVELOPMENT

- 1. **Completion Period**: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within **30** calendar days upon receipt of Notice to Proceed.
- 2. **Testing**; Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
- 3. **Training**: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff
- **4. Warranty**: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
- 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
- 6. **Manuals**: The supplier must provide the enduser one (1) hard and one (1) soft copy of the following:
  - a) Service manual in English language
  - b) Operation manual in English language
- 7. With "**DOH-MMCHD HFEP**" (Government Property not for sale) sticker in each unit.

### Recipient:

Health Facilities of Paranaque City



C. Additional Requirement to be submitted by the	
Single/Lowest Calculated Bidder (SCB/LCB) as part	
of post qualification:	
1. One (1) original sample of manufacturer's product to	
be submitted and returned after evaluation. The	
sample submitted and approved during the evaluation	
shall be the same item to be delivered upon award of	
contract. Prototype of the labelling instruction must be	
part of the sample submitted however, the technical	
specifications of the labelling instruction of the product	
must be complied upon delivery.	



Republic of the Philippines				
Department of Health				
Metro Manila Center for Health Development			nent	
	TECHN	IICAL SPECIF	CATIONS	
Item No. 8	MICROSCOPE BINOCULAR TYPE	Qty./Unit	1 Unit	
Name of Manufacturer:		Country of Origin		
		M - 1 -1 C:C		
Brand:		Model: (if applicable)		
ABC: <b>88,000</b>	0.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE		
A. Technica	l Specifications:			
Illuminatio	on			
- Built-in L	ED illumination system			
Focusing				
- Coaxial coarse and fine adjustments with 200 graduations		vith 200		
- Precision	of at least 2 microns			
Nosepiece				
- Quadrupl	e revolving nosepiece			
Eyepieces				
- Wide field	d 10x/20mm eyepieces			
Head				
- 30° inclin	ed			
- 360° rota	table			
- Interpupillary Distance: Adjustable between 55-75 mm				
Stage				
- Single specimen holder (1 pc Double specimen holder and 1 pc sheet holder)				
Condenser	•			
- must be adjustable				
- achromatic objectives 4x, 10x, 40x and 100x				



MENT OF THE STATE	
- all anti-fungus treated	
Power	
- 220 VAC, 50/60 Hz	
Accessories	
- Power Cord	
- LED Bulb	
- Dust Cover	
B. REQUIREMENTS IF AWARDED THE CONTRACT:	
1. <b>Completion Period</b> : The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.	
2. <b>Testing</b> ; Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.	
3. <b>Training</b> : The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff	
4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.	
5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.	
6. <b>Manuals</b> : The supplier must provide the enduser one (1) hard and one (1) soft copy of the	



following:	
a) Service manual in English language	
b) Operation manual in English language	
7. With "DOH-MMCHD HFEP" (Government	
Property not for sale) sticker in each unit.	
X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-	
Recipient:	
Health Facilities of Paranaque City	
C. Additional Requirement to be submitted by the	
Single/Lowest Calculated Bidder (SCB/LCB) as part	
Single/Lowest Calculated Bidder (SCB/LCB) as part	
Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:	
Single/Lowest Calculated Bidder (SCB/LCB) as part	
Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:  1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation	
Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:  1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of	
Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:  1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be	
Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:  1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical	
Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:  1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be	