



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 040E

PROCUREMENT OF VARIOUS MEDICAL INSTRUMENT (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
<p>Item No. 1 – 24 SETS MINOR SURGICAL SET</p> <p>3. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p>	<p>3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p>
<p>4. Manual:</p> <p>a) Service manual in English language</p>	<p>a) Manual for proper sterilization and cleaning.</p>
<p>Item No. 2 – 3 UNITS ADULT LARYNGOSCOPE</p> <p>3. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p>	<p>3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p>
<p>Item No. 3: 3 SETS MAJOR SURGERY INSTRUMENT SET WITH STERILIZING TRAY</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period</p>	<p>4. Notarized undertaking that the supplier shall replace the instrument with defect within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the instrument is unusable due to defective material or workmanship, shall be added to the warranty period.</p>
<p>5. Manual:</p> <p>a) Operation manual in English language</p>	<p>a) Manual for proper sterilization and cleaning.</p>
<p>Item No. 4: 4 UNITS PEDIA/NEONATE LARYNGOSCOPE</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment</p>	<p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of</p>



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

	the equipment - removed
<p>Item No. 5: 5 SETS BILATERAL TUBAL LIGATION SET</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p>	<p>4. Notarized undertaking that the supplier shall replace the instrument with defect within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the instrument is unusable due to defective material or workmanship, shall be added to the warranty period.</p>
<p>5. Manual:</p> <p>a) Service manual in English language</p>	<p>a) Manual for proper sterilization and cleaning.</p>
Item No. 6: 2 SETS CERVICAL INSPECTION SET	
Vaginal Speculum (S, M, L, size)	Vaginal Speculum (1pc S, 2pcs M, 1 pc L, size)
3 Ring Forceps – Long	3 Ring Forceps – Long: 3 pcs
2 Vaginal Retractor	Vaginal Retractor: 2 pcs
1 Needle holder long 8"	Needle holder long 8": 1 pc
1 Tissue Forceps Long 9"	Tissue Forceps Long 9": 1 pc
Rust proof, Stainless Steel	Stainless Steel
Revision and clarification to provisions/specifications in the Bidding Documents:	
<p>Item No. 7: 3 SETS DRESSING SET</p> <p>Additional:</p> <p>Kidney basin, stainless steel, 8"</p> <p>Instrument tray with lid, stainless steel, 12" x 8" x 2 1/2"</p>	
<p>REQUIREMENTS IF AWARDED THE CONTRACT:</p> <p>Additional:</p> <p>5. Manual:</p> <p>a) Manual for proper sterilization and cleaning.</p>	
FROM	TO
Item No. 7: 3 SETS DRESSING SET	
Surgical scissors, straight, Stainless Steel	Surgical scissors, straight, Stainless Steel 5 1/2"
Surgical scissors, curved, Stainless Steel	Surgical scissors, curved, Stainless Steel 5 1/2"
Bandage scissors, Stainless Steel	Bandage scissors, Stainless Steel 5 1/2"
Pick-up or ovum forceps, Stainless Steel	Pick-up or ovum forceps, Stainless Steel 7"
Mosquito forceps, Stainless Steel, 2pcs	Mosquito forceps, Stainless Steel, 5", 1 pc straight, 1 pc curved
Tissue forceps with teeth, Stainless Steel, 2pcs	Tissue forceps with teeth, Stainless Steel, 5 1/2"



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

	2pcs
Tissue forceps without teeth, Stainless Steel, 2pcs	Tissue forceps without teeth, Stainless Steel, 5 ½", 2pcs
Suture removal scissors, Stainless Steel	Suture removal scissors, Stainless Steel 4 ½"
3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff.	3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff. - Removed



Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 18th day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 9th day of April, 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III/ BAC Chairperson 



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 1	MINOR SURGICAL SET	Qty./Unit	24 Sets	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 480,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: Instrument tray with cover 18/8, stainless steel (410mm x 254mm x 64mm) Mayo Scissors 5-3/4", straight (14.5 cm) Mosquito hemostatic forceps straight 5" (12.5 cm) 20mm jaw with full serrations 2pcs Mosquito hemostatic forceps straight 4-3/4" (12 cm) 20mm jaw with full serrations 2pcs Needle holder, bulldog jaw, 16 cm (6-1/4") 1 pc Surgical knife handle #3 (scalpel) 1 pc Surgical knife handle #4 (scalpel) 1 pc Tissue forceps 1 x 2 teeth, 5-3/4" (14.5 cm) 2 pcs Bandage scissors With Laser Mark "DOH-MMCHD" in each instrument.				
<u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p style="padding-left: 40px;">a) Manual for proper sterilization and cleaning.</p> <p>X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-</p> <p>Recipients: Ospital ng Paranaque I - 2 sets Ospital ng Paranaque II – 4 sets</p>		
<p><u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer’s product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 2	ADULT LARYNGOSCOPE	Qty./Unit	3 Units	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 60,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> All instruments must be made of Stainless Steel. Fiber Optic Handle: Hollow, Cylindrical, Ribbed Blades: <ul style="list-style-type: none"> - Macintosh Blade # 1 - Macintosh Blade # 2 - Macintosh Blade # 3 - Macintosh Blade # 4 Laser mark the word "DOH-MMCHD" in each instrument. 				
<u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Testing; Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>3. Warranty: Warranty certificate for one (1) year on craftsmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"> a) Service manual in English language b) Operation manual in English language <p>7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.</p> <p>X-X</p> <p>Recipient: Ospital ng Paranaque I - 3 sets</p>		
<p><u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 3	MAJOR SURGERY INSTRUMENT SET WITH STERILIZING TRAY	Qty./Unit	3 sets	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 225,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> - 6 pcs. Sponge Holding Forceps 9 1/2" straight - 12 pcs. Backhaus Towel Clamp 5 1/4" - 1 pc. Knife Handle #3 - 1 pc. Knife Handle #4 - 1 pc. Knife Handle #7 - 1 pc. Mayo Dissecting Scissor 6 3/4" straight - 1 pc. Mayo Dissecting Scissor 6 3/4" curved - 1 pc. Metzenbaum Scissor 7" curved - 1 pc. Operating Scissor S/B 5 1/2" straight - 1 pc. Suture Wire Scissor Angled to side (One Serrated Blade) - 2 pcs. Thumb Forcep 5 1/2" Serrated - 1 pc. Thumb Forcep 8" Serrated - 2 pcs. Tissue Forcep (1x2T) 5 1/2" - 1 pc. Tissue Forcep (1x2T) 8" - 6 pcs. Allis Tissue Forcep (5x6T) 6" - 2 pcs. Babcock Tissue Forcep 6 1/4" - 12 pcs. Halsted Mosquito Forcep 5" straight - 12 pcs. Halsted Mosquito Forcep 5" curved - 6 pcs. Crile Forcep 6 1/4" straight 				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<ul style="list-style-type: none"> - 6 pcs. Crile Forcep 6 1/4" curved - 12 pcs. Rochester-Pean Forcep 6 1/4" curved - 6 pcs. Ochsner Forcep 6 1/4" straight - 2 pcs. Kelly Retractor Hollow Grip Handle (small) - 2 pcs. Kelly Retractor Hollow Grip Handle (medium) - 2 pcs. Kelly Retractor Hollow Grip Handle (large) - 2 pcs. Deaver Retractor 2×12" - 2 pcs. Volkman Retractor 4 prong 9" Blunt - 2 pcs. Volkman Retractor 6 prong 9" Blunt - 2 pcs. Cushing Vein Retractor (10mm) 8" - 1 set US Army Retractor (set of 2 pcs.) - 1 pc. Malleable Retractable (Ribbon) 1×13" - 2 pcs. Malleable Retractable (Ribbon) 1 1/2×13" - 1 pc. Balfour Abdominal Retractable 8" spread - 1 pc. Groove Director and Tongue tie 5" - 1 pc. Probe with Eye 5" - 1 pc. Deschamps Ligature Carrier Blunt Left - 1 pc. Deschamps Ligature Carrier Blunt Right - 1 pc. Poole Suction Tube straight - 2 pcs. Mayo-Hegar Needle Holder 6" - 2 pcs. Mayo-Hegar Needle Holder 7" - 1 pc. Sterilizing tray (Perforated) • Laser mark the word "DOH-MMCHD" in each instrument. 		
<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2.; Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The</p>		



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>equipment must be functioning and must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for one (1) year on workmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall replace the instrument with defect within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the instrument is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Manual for proper sterilization and cleaning.</p> <p>x-x</p> <p>Recipient:</p> <p>Ospital ng Paranaque I - 3 sets</p>		
<p><u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 4	PEDIA/NEONATE LARYNGOSCOPE	Qty./Unit	4 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 72,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> • All instruments must be made of Stainless Steel. • Fiber Optic • Handle: Hollow, Cylindrical, Ribbed • Blades: <ul style="list-style-type: none"> - Macintosh Blade # 00 - Macintosh Blade # 0 - Macintosh Blade # 1 • Laser mark the word "DOH-MMCHD" in each instrument . 			
<u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2.; Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>physical damage and defect.</p> <p>3.Warranty: Warranty certificate for one (1) year on workmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Service manual in English language</p> <p>x-x</p> <p>Recipient:</p> <p>Ospital ng Paranaque I - 4 sets</p>		
<p><u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 5	BILATERAL TUBAL LIGATION SET	Qty./Unit	5 Sets	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 150,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> • All instruments must be made of stainless steel. - 1 Crile Hemostatic Forceps 5-1/2" (13.8cm), curved - 1 Crile Hemostatic Forceps 5-1/2" (13.8cm), straight - 1 Babcock Forceps 8-1/4" (20.6cm), 10mm wide jaws - 1 Metzenbaum Scissors 7" (17.5cm), tungsten carbide, curved - 1 Mayo Scissors 6-3/4" (16.9cm), tungsten carbide, straight, beveled blades - 1 Tissue Forceps 6" (15cm), 1x2 teeth - 1 Dressing Forceps 6" (15cm), serrated tips - 1 DeBakey Crile-Wood Needle Holder 7" (17.5cm), serrated jaws, tungsten carbide - 1 Mayo-Hegar Needle Holder 7" (17.5cm), serrated, tungsten carbide - 2 Richardson Retractor w/ Loop Handle, small 1" (2.5cm) x 3/4" (1.9cm), 9-1/2"(23.8cm - 2 Richardson Retractor w/ Loop Handle, medium 1-1/4" (3.2cm) x 1" (2.5cm), 9-1/2"(23.8cm) 				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>- 2 Scalpel Handle #3 Graduated CM's - for blade sizes #10-15, extra fine 5" (12.5cm)</p> <p>- 1 Sterilization Container - Small w/ Perforated Lid - Silver 11-1/4"W x 11-3/8"L x 4"H</p> <p>- 1 Wire Basket for Sterilization Container - Small 9-7/8"W x 10-1/4"L x 2"H w/ inside handle</p> <p>• Laser mark the work "DOH-MMCHD" in each equipment .</p>		
<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2.; Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3.Warranty: Warranty certificate for one (1) year on workmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Notarized undertaking that the supplier shall replace the instrument with defect within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the instrument is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Manual for proper sterilization and cleaning.</p> <p>X-X</p> <p>Recipients:</p>		



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Osipital ng Paranaque II - 4 sets Health Facility in Paranaque City – 1 set		
<u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u> 1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.		



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 6	CERVICAL INSPECTION SET	Qty./Unit	2 Sets	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 40,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> ● Vaginal Speculum (1pc S, 2pcs M, 1 pc L, size) ● 3 Ring Forceps – Long: 3 pcs ● Vaginal Retractor: 2 pcs ● Needle holder long 8": 1 pc ● Tissue Forceps Long 9": 1 pc ● Stainless Steel ● Laser mark the word "DOH - MMCHD" in each instrument. 				
<u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2.; Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3.Warranty: Warranty certificate for one (1) year on workmanship. The supplier shall either repair				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Service manual in English language / Sterilization manual in English</p> <p>X-X</p> <p>Recipient:</p> <p>Ospital ng Paranaque II - 2 sets</p>		
<p><u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 7	DRESSING SET	Qty./Unit	3 Sets	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 60,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> ● Surgical scissors, straight, Stainless Steel 5 ½" ● Surgical scissors, curved, Stainless Steel 5 ½" ● Bandage scissors, Stainless Steel 5 ½" ● Pick-up or ovum forceps, Stainless Steel 7" ● Mosquito forceps, Stainless Steel, 5", 1 pc straight, 1 pc curved ● Tissue forceps with teeth, Stainless Steel, 5 ½" 2pcs ● Tissue forceps without teeth, Stainless Steel, 5 ½", 2pcs ● Suture removal scissors, Stainless Steel 4 ½" Additional: Kidney basin, stainless steel, 8" Instrument tray with lid, stainless steel, 12" x 8" x 2 ½" <ul style="list-style-type: none"> ● Laser mark the word "DOH - MMCHD " in each instrument. 				
<u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2; Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for one (1) year on workmanship. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>4. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Manual for proper sterilization and cleaning.</p> <p>X-X</p> <p>Recipient:</p> <p>Health Facilities in Paranaque City – 3 set</p>		
<p><u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		