



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB 2022 – 051E**

**PROCUREMENT OF 18 UNITS AUTOMATED EXTERNAL DEFIBRILLATOR (AED)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
Recipient: Health Facilities in Paranaque City	
<b>From</b>	<b>To</b>
• Includes five (5) scenarios, volume control and language button	• <del>Includes five (5) scenarios, volume control and language button/settings</del> – REMOVED
Pads can be used at least 25 times during CPR/AED classes	• <del>Pads can be used at least 25 times during CPR/AED classes</del> – REMOVED
1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.	1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30-60</b> calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 18<sup>th</sup> day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 9<sup>th</sup> day of April, 2022 in MMCHD

Approved by:

  
**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson *yg*




# METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>				
Item No. <b>1</b>	<b>AUTOMATED EXTERNAL DEFIBRILLATOR (AED)</b>	Qty./Unit	<b>18 Units</b>	
Name of Manufacturer:		Country of Origin		
Brand:		Model: (if applicable)		
ABC: <b>2,700,000.00</b>				
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE		
<b>A. Technical Specifications:</b>  <ul style="list-style-type: none"> <li>• Two-button Operation: On/Off, Discharge (shock button)</li> <li>• Voice Prompts: Voice Command that gives instruction to the user in operating the device</li> <li>• Automatically evaluates patient impedance for proper pad contact.</li> <li>• Energy output accuracy: <math>\pm 15\%</math></li> <li>• Pulse shape: Biphasic</li> <li>• Charge time: maximum of 15 seconds</li> <li>• Battery Capacity: Capable of providing at least 200 discharges at maximum energy</li> <li>• Replaceable battery</li> <li>• Visible and Audible Indicators:               <ul style="list-style-type: none"> <li>- Low battery</li> <li>- Pad skin contact/disconnection</li> <li>- AED Status</li> </ul> </li> </ul>				



## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<ul style="list-style-type: none"> <li>• Defibrillator Pads: self-adhesive, disposable and non-polarized</li> <li>- at least five (5) pairs for adult</li> <li>- at least five (5) pairs for pedia</li> </ul>		
<p><b><u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u></b></p> <p><b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30 – 60</b> calendar days upon receipt of Notice to Proceed.</p> <p><b>2; Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p><b>3. Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p><b>4.Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period</p> <p><b>6. Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p>		

[illegible]