



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 020E

PROCUREMENT OF VARIOUS HOSPITAL EQUIPMENTS (LOT BIDDING) (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 1	
No Changes	
FROM	TO
IB 2022-003E	IB 2022-020E
Item No. 2 Table with drawers	Table with at least 3 drawers and 1 cabinet Drawer & cabinet dimensions: Manufacturer's Standard
Item No. 3 ER BED / TRANSFER STRETCHER	TRANSFER STRETCHER
Weight Capacity: At least 750 lbs	Weight Capacity: At least 660 lbs (300 kg)
Mattress: At least 6"	Mattress: At least 4"
Height: Adjustable 23-26"	Height: Adjustable 23-36"

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 4th day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 25th day of March, 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson

Technical Specifications

Item	Specification	Statement of Compliance
		<p><i>[Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>
<p>IB 2022- 020E</p>	<p>PROCUREMENT OF VARIOUS HOSPITAL EQUIPMENTS (LOT BIDDING) (REBID) ABC: P 990,000.00</p> <p>ITEM 1: 6 UNITS WHEELED DETACHABLE STRETCHER ABC: P 420,000.000</p> <p>Technical Specifications: Features: Detached, Automatic Loading, Aluminum Alloy & Stainless Steel Material: Aluminum Alloy & Stainless Steel or equivalent Maximum Back Angle:65° Has a separable stretcher with the main stretcher frame which could be removed from trolley and used as a separate stretcher Easy locking/unlocking system to lock the separable stretcher into the trolley Two sets loading wheels to fit different height of ambulance board The backrest can be adjusted from 0 - 65 deg by gas spring Material: Aluminum Alloy & Stainless Steel or equivalent The mechanical legs are controlled by the handles in both sides, There is a safety lever to control front and hind legs to fold while loading and prevent collapse while standing</p>	

The stretcher can be locked by the fixing device when it is in the ambulance

Made of high-strength aluminum alloy or equivalent.

Wide rubber castors, two with brakes

With "**DOH-MMCHD HFEP**"(Government Property not for sale) sticker in each unit.

DOCUMENTARY REQUIREMENTS:

1. Product brochure or technical data sheet(s) of the equipment showing the technical specifications in English Language.
2. Valid and current Certificate of Compliance of manufacturer of the equipment with the latest version of ISO 13485: Quality Management System – Requirements for regulatory purposes in the name of the manufacturer. The Certificates must be issued by an independent Certifying Body/Agency.
3. Valid Certificate of Distributorship (as first Tier Distributor) issued by the Manufacturer of each equipment authorizing the bidder to sell/distribute the offered equipment.
4. Proof (such as sales invoice) that the Brand of the equipment has been sold to other health facilities in the Philippines.
5. Notarized Certificate of the bidder:
 - a. That the brand of the equipment has been in the local and/or international market for at least ten (10) years.
 - b. That the equipment and its accessories are brand new, unused, not discontinued models and were not subjected to any product recall.
6. Bidder’s valid and Current License to Operate (LTO) as a medical device distributor issued by the Philippine Food and Drug Administration. In case of expired LTO, the following must be submitted: i) Copy of expired LTO, ii) Application for renewal, iii) Official Receipt as proof of payment for the renewal of LTO.

REQUIREMENTS IF AWARDED THE CONTRACT:

1. **Completion Period:** The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed **within 30 calendar days** upon receipt of Notice to Proceed.
2. **Testing:** Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
3. **Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar

days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

5. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

X-

Recipients:

Muntinlupa - 3 units

Municipality of Pateros - 3 units

ITEM 2

20 UNITS EXAMINATION TABLE

ABC: 500,000.00

TECHNICAL SPECIFICATIONS:

Steel top 3 section adjustable

With stainless steel heel detachable stirrups

With mattress and leatherette cover at least 4-inch

With stainless steel drain pan

Table with at least 3 drawers and 1 cabinet

Drawer & cabinet dimensions: Manufacturer’s Standard

With retractable rubber covered footstep

Table Top Dimension: Manufacturer’s Standard

DOCUMENTARY REQUIREMENTS:

1. Product brochure or technical data sheet(s) of the equipment showing the technical specifications in English Language.
2. Valid and current Certificate of Compliance of manufacturer of the equipment with the latest version of ISO 13485: Quality Management System – Requirements for regulatory purposes in the name of the manufacturer. The Certificates must be issued by an independent Certifying Body/Agency.
3. Valid Certificate of Distributorship (as first Tier Distributor) issued by the Manufacturer of each equipment authorizing the bidder to sell/distribute the offered equipment.
4. Notarized Certificate of the bidder:
 - a. That the brand of the equipment has been in the local and/or international market for at least ten (10) years.
 - b. That the equipment and its accessories are brand new, unused, not discontinued models and were not subjected to any product recall.
5. Bidder’s valid and Current License to Operate (LTO) as a medical device distributor issued by the Philippine Food and Drug Administration. In case of expired LTO, the following must be submitted: i) Copy of expired LTO, ii) Application for renewal, iii) Official Receipt as proof of payment for the renewal of LTO.

REQUIREMENTS IF AWARDED THE CONTRACT:

1. **Completion Period:** The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with **30 calendar days** upon receipt of Notice to Proceed.
2. **Testing:** Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
3. **Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
5. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
 - a) Service manual in English language
 - b) Operation manual in English language
6. With **"DOH-MMCHD HFEP"(Government Property not for sale)** sticker in each unit

Recipients:

Marikina City - 8 units
Muntinlupa City - 12 units

ITEM 3

1 TRANSFER STRETCHER

ABC: 70,000.00

TECHNICAL SPECIFICATIONS:

Weight Capacity: At least 660 lbs (300 kg)
Dimensions: 83" x 35" x 23-36" (LxWxH)
Adjustable pneumatic back rest
With Oxygen tank holder
2-hook IV pole and pole socket
Mattress: At least 4"
Material: Metal
Height: Adjustable 23-36"

DOCUMENTARY REQUIREMENTS:

1. Product brochure or technical data sheet(s) of the equipment showing the technical specifications in English Language.
2. Valid and current Certificate of Compliance of manufacturer of the equipment with the latest version of ISO 13485: Quality Management System – Requirements for

regulatory purposes in the name of the manufacturer. The Certificates must be issued by an independent Certifying Body/Agency.

3. Valid Certificate of Distributorship (as first Tier Distributor) issued by the Manufacturer of each equipment authorizing the bidder to sell/distribute the offered equipment.

4. Notarized Certificate of the bidder:

a. That the brand of the equipment has been in the local and/or international market for at least ten (10) years.

b. That the equipment and its accessories are brand new, unused, not discontinued models and were not subjected to any product recall.

5. Bidder's valid and Current License to Operate (LTO) as a medical device distributor issued by the Philippine Food and Drug Administration. In case of expired LTO, the following must be submitted: i) Copy of expired LTO, ii) Application for renewal, iii) Official Receipt as proof of payment for the renewal of LTO.

REQUIREMENTS IF AWARDED THE CONTRACT:

1. **Completion Period:** The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed **with 30 calendar days** upon receipt of Notice to Proceed.

2. **Testing:** Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.

3. **Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

5. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

a) Service manual in English language

b) Operation manual in English language

With "**DOH-MMCHD HFEP**" (**Government Property not for sale**) sticker in each unit

Recipients:

Muntinlupa City - 1 unit