



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 117E

#### PROCUREMENT OF 1 UNIT HYPERTHERMIA BLOOD FLUID WARMING (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
PROCUREMENT OF 1 UNIT HYPERTHERMIA BLOOD FUILD WARMING (REBID)	PROCUREMENT OF 1 UNIT HYPERTHERMIA BLOOD FLUID WARMING (REBID)

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 30<sup>th</sup> day of August 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 22<sup>nd</sup> day of August 2022 in MMCHD

Approved by:

SGD.

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. <b>1</b>	<b>HYPERTHERMIA BLOOD FLUID WARMING</b>	Qty./Unit	<b>1 Unit</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>270,000.00</b>			
<b>PURCHASER'S SPECIFICATION</b>			<b>STATEMENT OF COMPLIANCE</b>
<b>Technical Specifications:</b> <ul style="list-style-type: none"> <li>• With two airflow settings and hose-end temperature sensor</li> <li>• Capable of recording machine errors and therapy reports</li> <li>• With built-in hour meter or service meter</li> <li>• Dimensions: Manufacturer's Standard</li> <li>• Noise Level: ≤ 55 dBA</li> <li>• Hose: Flexible</li> <li>• Filtration System: At least 0.2 µm air filter</li> <li>• Mounting: Can be clamped on to an IV pole, placed on a hard surface or mounted on a rolling stand accessory.</li> <li>• Temperature Control: Electronically controlled.</li> <li>• Alarm System: With audible and visual alarm</li> <li>• Blower Motor               <ul style="list-style-type: none"> <li>- Speed: Max of at least 4500 rpm for high-fan setting</li> <li>- Airflow: At least up to 45 cfm</li> </ul> </li> <li>• Power Supply: 220V, 60 Hz</li> <li>• Inclusion:               <ul style="list-style-type: none"> <li>- Rolling Cart</li> </ul> </li> </ul>			
<b><u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></b>			
<b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed			



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

<p>within 30 calendar days upon receipt of Notice to Proceed.</p> <p><b>2. Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p><b>3. Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within (3) days upon delivery of the equipment.</p> <p><b>4. Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p><b>5. Notarized undertaking</b> that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p><b>6. Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"><li>a) Service manual in English language</li><li>b) Operation manual in English language</li></ul> <p><b>7. With "DOH-MMCHD HFEP"</b> (Government Property not for sale) sticker in each unit.</p> <p>X-X</p> <p>Recipient Gat Andres Bonifacio Memorial Medical Center - 1 unit</p>	
<p><b><u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></b></p> <p>1. One (1) original sample of manufacturer’s product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	