



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 065E

PROCUREMENT OF 3 UNITS RADIANT WARMER WITH PHOTOTHERAPY

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
• Air mode temperature range: 25.0°C~37.0°C	• Air mode temperature range: 34.0°C~37.5°C
• Available adjustment of radiator and mattress tilting	• Available adjustment of radiator and mattress tilting or its equivalent
• With phototherapy	With integrated phototherapy
1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed.	1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 90 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 23rd day of May 2022, 9:00 AM:


This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 15th day of May 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	RADIANT WARMER WITH PHOTOTHERAPY	Qty./Unit	3 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,335,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> Power Supply: 220V, 50/60Hz Air mode temperature range: 34.0°C~37.5°C Over temperature alarm: 38.0°C Temperature Deviation alarm: ±1.0°C Skin sensor accuracy: ≤±0.3°C Mattress temperature uniformity: ≤2.0°C Heating rate: ≤30minutes APGAR timing: 1min, 5min, 10min Available adjustment of radiator and mattress tilting or its equivalent Skin temperature and manual control With Illumination Light With integrated phototherapy			
<u>REQUIREMENTS IF AWARDED THE CONTRACT</u> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 90 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment 			



1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.