



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 115E

PROCUREMENT OF VARIOUS BIOMEDICAL REFRIGERATOR.(LOT BIDDING) (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

FROM	TO
Item No. 1: BIOREFRIGERATOR <ul style="list-style-type: none">• Capacity: At least 340 L• Temperature Range: -60 ~ -86°C• Display: Digital, touch screen LCD or better• Material: Stainless Steel 304- Low battery	Item No. 1: BIOREFRIGERATOR <ul style="list-style-type: none">• Capacity: At least 335L• Temperature Range: - 40 to -60 ~ -80°C• Display: Digital, touch screen LCD or LED or better• Material: Cold rolled steel or better, powder coatedLow battery or Low voltage
Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 2: REAGENT REFRIGERATOR With LED lamp	
FROM	TO
<ul style="list-style-type: none">• Stainless steel interior• Cyclical defrosting and evaporator temperature detection system	<ul style="list-style-type: none">• Stainless steel interior• Interior Material: Cold rolled steel, power coated• Cyclical defrosting and evaporator temperature detection system or auto defrost
1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 19th day of July 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 12th day of July 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	BIOREFRIGERATOR	Qty./Unit	21 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 3,780,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>A. Technical Specifications:</p> <ul style="list-style-type: none">• Capacity: At least 335L• Cabinet Type: Vertical• Cooling System: Direct Cooling• Temperature Range: - 40 to -60 ~ -80°C• Humidity Range: 30~60% RH• Display: Digital, touch screen LCD or LED or better• Material: Cold rolled steel or better, powder coated• Insulation: Polyurethane (PU) foam• Power Supply: 220V, 50/60Hz• Dimensions: Manufacturer's Standard• Alarms:<ul style="list-style-type: none">- High/Low Temperature- Power Failure- Abnormal Voltage- Sensor Error- Filter alarm- Low battery or Low voltage- Door ajar• Accessories:<ul style="list-style-type: none">- Temperature recorder			
B. REQUIREMENTS IF AWARDED THE CONTRACT:			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

1. **Completion Period:** The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within **60** calendar days upon receipt of Notice to Proceed.
2. **Testing:** Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
3. **Training:** The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment
4. **Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
 - a) Service manual in English language
 - b) Operation manual in English language
7. With "**DOH-MMCHD HFEP**"(Government Property not for sale) sticker in each unit.

X-X

Recipient: Health Facilities of Paranaque

C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted *however, the technical specifications of the labelling instruction of the product must be complied upon delivery.*



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 2	REAGENT REFRIGERATOR	Qty./Unit	2 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 70,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
A. Technical Specifications: <ul style="list-style-type: none"> Power Supply Input: 220V, 50/60Hz Effective Capacity: at least 300 Liters Polyurethane foam insulation Hinged door, double-paned glass door Compressor, at least 160 W Interior Material: Cold rolled steel, power coated or better. Cyclical defrosting and evaporator temperature detection system or auto defrost Digital display temperature Temperature range of 2°C to 14°C With fluorescent lamp With 4 caster wheel swiveling with brakes Safety Features: <ul style="list-style-type: none"> - High and Low Temperature alarm system with audio-visual indicator - Temperature recorder - Door ajar alarm With LED lamp			
B. DOCUMENTARY REQUIREMENTS: <ol style="list-style-type: none"> 1. Product brochure or technical data sheet(s) of the equipment showing the technical specifications in English Language. 2. Valid and current Certificate of Compliance of manufacturer of the equipment with the latest version of ISO 13485: Quality Management System – Requirements for regulatory purposes in the name of the manufacturer. The 			



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>Certificates must be issued by an independent Certifying Body/Agency.</p> <p>3. Notarized Certificate of the bidder:</p> <p>a. That the brand of the equipment has been in the local and/or international market for at least ten (10) years.</p> <p>b. That the equipment and its accessories are brand new, unused, not discontinued models and were not subjected to any product recall.</p> <p>4. Bidder's valid and Current License to Operate (LTO) as a medical device distributor issued by the Philippine Food and Drug Administration. In case of expired LTO, the following must be submitted: i) Copy of expired LTO, ii) Application for renewal, iii) Official Receipt as proof of payment for the renewal of LTO.</p>	
<p><u>C. REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Service manual in English language</p> <p>b) Operation manual in English language</p> <p>7. With "DOH-MMCHD HFEP" (Government Property not for</p>	

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>(sale) sticker in each unit.</p> <p>X-X</p> <p>Recipient:</p> <p>Health Facilities of Paranaque 1 unit</p> <p>Health Facilities of Marikina - 1 unit</p>	
<p>C. <u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <ol style="list-style-type: none"> 1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery. 	