



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 032E

PROCUREMENT OF VARIOUS MEDICAL FURNITURE (LINE BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
From	To
Bidding Document Title: PROCUREMENT OF VARIOUS MEDICAL FURNITURE (LOT BIDDING)	Bidding Document Title: PROCUREMENT OF VARIOUS MEDICAL FURNITURE (LINE BIDDING)
Item No. 1 - 11 Units DENTAL INSTRUMENT CABINET Size: Manufacturer's Standard	Remove: Size: Manufacturer's Standard
Silent anti-skid wheel with brakes	Caster wheel with brakes
Anti-corrosion, fireproof, high hardness, tempered glass top with baffle.	Tempered glass top with baffle. Remove: Anti-corrosion, fireproof, high hardness
Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 1 - 11 Units DENTAL INSTRUMENT CABINET Dimension: Height: At least 820 mm Length: At least 500 mm Depth: At least 500 mm	
Item No. 2 - 3 Units INSTRUMENT CABINET Dimension: Height: At least 1800 mm Length: At least 800 mm Depth: At least 400 mm Stainless steel: at least grade 304	
Item No. 3 - 7 Units MEDICINE CABINET *No changes as stated in technical specifications.	
From	To
Item No. 4 - 7 Units EXAMINATION TABLE With stainless steel heel detachable stirrups	With stainless steel heel detachable stirrups, Stainless steel: at least grade 304
With stainless steel drain pan	With stainless steel drain pan, Stainless steel: at least grade 304
Table with drawers	Table with drawers: Drawer: at least 3 drawers and 1 cabinet Dimension: Manufacturer's Standard Material: Powder coated or better
Item No. 5 Title- 1 Unit ER BED / TRANSFER STRETCHER	Item No. 5 Title- 1 Unit TRANSFER STRETCHER
Weight Capacity: At least 750 lbs	Weight Capacity: at least 660lbs (300kg)
Dimensions: 83" x 35" x 23-36" (LxWxH)	Dimensions: at least 75-83x23-35x23-36in (LxWxH)



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Mattress: At least 6"	Mattress: at least 4 in
Height: Adjustable 23-26"	Height: Adjustable 23-36 in
Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 6 - 38 Units INFANT WEIGHING SCALE, DIGITAL *No changes as stated in technical specifications.	
Item No. 7 - 23 Units WEIGHING SCALE WITH HEIGHT MEASURING STICK, HEAVY DUTY, ADULT *No changes as stated in technical specifications.	


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 11th day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 4th day of April, 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson

8