



# METRO MANILA CENTER FOR HEALTH DEVELOPMENT

## SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 051

### PROCUREMENT OF 800 BOXES MUTI-USE TEMPERATURE DATA LOGGER WITH DISPLAY (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
Alarm: preferably visible and audible	
The calibration certificate can be provided upon provision of the sample item for the use of BAC TWG evaluation.	
FROM	TO
Unit: Box	Unit: Piece
Sensor measurement range: -40 C to + 99 C (-40 F to 210 F)	Sensor measurement range Upper limit: + 55 C (131 F) or higher Lower limit: -30 C (-22 F) or lower


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12<sup>th</sup> day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 5<sup>th</sup> day of April 2022 in MMCHD

Approved by:

  
**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson