



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2024-001E
PROCUREMENT OF 1 UNIT 2D ECHOCARDIOGRAM/VASCULAR MACHINE
(SHORT OF AWARD)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
• Digital Video Recorder - USB Type	• Digital Video Recorder - USB Type with Storage USB 500GB Capacity
• Type: Portable Ultrasound, Laptop style with trolley	• Type: Portable Ultrasound, Laptop or console style with trolley - RETAIN
• Support 8 hours continuous scanning	• Support 8 hours continuous scanning – RETAIN THE ORIGINAL SPECIFICATION
• Motion Mode: Not less than 3 Lines	• Motion Mode: Not less than 3 Lines or Manufacturer Standard– RETAIN THE ORIGINAL SPECIFICATION
• Phased Array Adult: 1.5 - 4.5 MHz, up to 64 elements	• Phased Array Adult: 1.5 - 4.5 MHz, up to 64 elements – RETAIN THE ORIGINAL SPECIFICATION
• Phased Array Pedia: 2.0 - 8.0 MHz, up to 96 elements	• Phased Array Pedia: 2.0 - 8.0 MHz, up to 96 elements – RETAIN THE ORIGINAL SPECIFICATION
• Linear Transducer: 3.2 - 12.3 MHz, up to 128 elements	• Linear Transducer: 3.2 – 12.3 MHz, up to 128 elements– RETAIN THE ORIGINAL SPECIFICATION

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on November 20, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 14th day of November, 2023 in MMCHD.

Approved by:

SGD
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No. 1	2D ECHOCARDIOGRAM/VASCULAR MACHINE	Qty./Unit	1 Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 4,500,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
TECHNICAL SPECIFICATIONS: Physical Specification <ul style="list-style-type: none"> • Type: Portable Ultrasound, Laptop style with trolley • Main Display: at least 15" high resolution color LED monitor • Main Display Resolution: at least 1920*1080 • Touch screen panel size: at least 12" high sensitivity anti-glare • Touch screen panel Resolution: at least 1920*720 • With at least 3 active probe ports • With at least 4 USB ports • Power supply: 220V, 50/60 Hz • With replaceable and rechargeable lithium battery (Built-in Battery for main unit) • Support 8 hours continuous scanning • Storage: at least 128 GB Solid State Drive <ul style="list-style-type: none"> - at least 45 GB internal hard drive reserved for patient data storage - Capable of storage up to approximately 173,242 single frames • Connectivity: Cable and Wireless/WiFi • Abdomen/General, Obstetrics, Gynecology, Cardiac, Small Parts, Urology, Vascular, Pediatrics and Nerve Imaging Parameters and features: <ul style="list-style-type: none"> • Digital Beamformer • Spatial compound imaging • Speckle suppression imaging • Full screen zoom • High resolution flow • Motion Mode: Not less than 3 Lines • Tissue Tracking with Quantitative Analysis 			

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
2D Echocardiogram/Vascular Machine	1 unit	Health Facilities in The City of Marikina	90 calendar days upon receipt of the Notice to Proceed

Name of Company: _____

Address: _____

Signature Over Printed Name : _____

Telephone/Fax Number : _____

Email: _____