



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB NO. 2023-117E  
PROCUREMENT OF 13 UNITS AUTOCLAVE 20L**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
No changes as stipulated in technical specification	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on November 6, 2023, 9:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 27th day of October, 2023 in MMCHD.

**PRETCHELL P. TOLENTINO, MD, MCHM**  
Director III / BAC Chairperson

*Section VII. Technical Specifications*

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No. 1</b>	<b>AUTOCLAVE 20L</b>	Qty./Unit	<b>13UNITS</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>780,000.00</b>			
<b>PURCHASER'S SPECIFICATION</b>			<b>STATEMENT OF COMPLIANCE</b>
<b>TECHNICAL SPECIFICATIONS:</b> <ul style="list-style-type: none"> <li>• Chamber: Stainless Steel, Horizontal</li> <li>• Chamber Capacity: At least 20 Liters</li> <li>• Power Supply: 220V, 60Hz, Single-phase</li> <li>• Heater: At least 1kW</li> <li>• Safety Devices:                             <ul style="list-style-type: none"> <li>-Low Water Cut-off Switch</li> <li>-Safety Release Valve</li> <li>-Emergency Exhaust Valve</li> <li>-Door Safety</li> <li>-Pressure Gauge with Pressure Control Switch</li> <li>-Steam Trap</li> <li>-Timer with Alarm</li> <li>-Standard Accessories: Stainless Steel Tray</li> </ul> </li> </ul>			
<b>Requirements if awarded the Contract</b> <ol style="list-style-type: none"> <li>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.</li> <li>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and</li> </ol>			

