



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2023-119E
PROCUREMENT OF 16 UNITS
DENTAL UNIT, MOTORIZED WITH ACCESSORIES

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
No changes as stipulated in technical specification	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on November 6, 2023, 9:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 27th day of October, 2023 in MMCHD.

SGD
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	DENTAL UNIT, MOTORIZED With ACCESSORIES	Qty./Unit	16 UNITS
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 2,880,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
TECHNICAL SPECIFICATIONS: <ul style="list-style-type: none"> · Fully motorized with pre-set functions · Touch control chair · Slim ergonomic design upholstery · LED light with sensor switch · Double side armrest · Colored glass/ ceramic bowl, rotating detachable cuspidor bowl (with plastic strainer) · Two (2) pcs. 3 way syringe · Three (3) pcs hand piece attachments (2 hi-speed adaptors & 1 low-speed adaptor) · Auto-cup-filler & bowl flush (water flushing control valve, cup-filler with control valve) · Saliva ejector-built-in motor pump type, does not use air or water · Clean water system · LED X-ray film viewer (Peri-apical X-ray viewer) · Multi-functional foot control · Water tank for distilled water for hand piece · Two (2) hi-speed hand piece · One (1) low-speed complete with straight hand piece & Contra-angle · LED light cure unit (external) · Tissue box with tray and cup dispenser · Silent type and oil-free dental air compressor · Saddle-type with backrest dentist stool 			
Requirements if awarded the Contract <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p>			

