



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2023-125E
PROCUREMENT OF 27 UNITS WHEELED DETACHABLE STRETCHER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
•Castor wheels: At least 100 mm diameter castors	>Castor wheels: At least 100 mm diameter castors (4 castor wheels).
• Dimensions: L x W x H = 2000mm x 600mm x 850mm, ± 50mm in every dimension, adjustable height must be at least in range or bigger	Dimensions: L x W x H = 2000mm x 600mm x 850mm, ±100mm in every dimension, adjustable height must be at least in range or bigger
Additional: >with side rail	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on November 6, 2023, 9:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 27th day of October, 2023 in MMCHD.

SGD
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	Wheeled Detachable Stretcher	Qty./Unit	27 UNITS
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,890,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Frame: Aluminum Alloy • Type: Detachable/Removable • Mattress: At least 3 cm thick, water proof seamless mattress • Has 3 belts • Castor wheels: At least 100 mm diameter castors • Max. Load: At least 160kg • Dimensions: L x W x H = 2000mm x 600mm x 850mm, ± 50mm in every dimension, adjustable height must be at least in range or bigger • Adjustable height: 300mm to 850mm 			
Requirements if awarded the Contract			
1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30-60			

