



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2023-131E
PROCUREMENT OF 1 UNIT AUTOREFRACTOR/KERATOMETER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
No changes as stipulated in technical specification	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on November 6, 2023, 9:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 27th day of October, 2023 in MMCHD.

PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	AUTOREFRACTOR/KERATOMETER	Qty./Unit	1 UNIT
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 480,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Measurement modes <ul style="list-style-type: none"> - Continuous Keratometry and Refractometry - Refractometry - Keratometry • Has Color view and Contact lens fitting assistance (white and blue LED light) • Refractometry <ul style="list-style-type: none"> - Vertex Distance: 0.0, 12.0, 13.75, 15.0 - Sphere: -30.00~+25.00D (VD= at least 12mm)(increments: 0.01, 0.12, 0.25D) - Cylinder: 0.00~±12.00D (increments: 0.01, 0.12, 0.25D) - Axis: 0~180° (1° unit) - Astigmatism Indication: -, +, ± (mixed) - Pupil Distance: 10~85mm - Minimum Pupil Diameter(Ø): 2.0mm • Keratometry <ul style="list-style-type: none"> - Radius of Curvature: 5.0~13.0mm (Increments: 0.01mm) - Cornea Power: 25.96D~67.50D (Increments: 0.05, 0.12, 0.25D) 			

Source of Fund: SAA 2022-02-0531 (HFEP 2022 ConAp)	
--	--

Recipients:

Ospital ng Paranaque II - 1 unit