



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2023-138E
PROCUREMENT OF 1 SET
COMPLETE TRACTION SET WITH DEVICE (CERVICAL AND LUMBAR)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
-Pelvic Belt (locally manufactured)	- Pelvic Belt

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on November 6, 2023, 9:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 27th day of October, 2023 in MMCHD.

SGD
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	COMPLETE TRACTION SET WITH DEVICE (CERVICAL AND LUMBAR)	Qty./Unit	1 SET
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 365,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
TECHNICAL SPECIFICATIONS: <ul style="list-style-type: none"> • Traction Force: 2.5kg/5.5lb (minimum) 91kg/200lb (maximum) • Traction Force Compensation: Computerized Automatic Compensation • Force Display: Actual force/preset force-automatic switching display • Traction Speed: Stepless, continuous, and adjustable • Therapy Mode: Continuous, Harmonic, Intermittent, Progressive • Treatment Time: 1 to 99 minutes ($\pm 2\%$) • Power Consumption: 150VA max • Power Supply: 100/120/220/240V, 50/60Hz • Current: 1.2A (110V, 120V), 0.6A (220V,240V) • Traction Bed and accessories: <ul style="list-style-type: none"> - Flexion Stool - Pelvic Belt 			
Requirements if awarded the Contract 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and			

