



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2023-151E
PROCUREMENT OF 1 UNIT DIAGNOSTIC AUDIOMETER WITH SPEECH TEST
AND FREEFIELD TEST (ADULT)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
<ul style="list-style-type: none"> • Display: At least 5-inch Color Display • Thermal printer size: At least 3" 	<ul style="list-style-type: none"> • Display: At least 4-inch Color Display • Thermal Printer size: At least 3" thermal printer or laptop (i5, windows 11, 16GB RAM) with mouse, keyboard and external printer
<p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30-60 calendar days upon receipt of Notice to Proceed.</p>	<p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60-90 calendar days upon receipt of Notice to Proceed.</p>
<p>4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p>	<p>4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. - laptop and printer not covered</p>

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on December 11, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 1st day of December, 2023 in MMCHD.

Approved by:

SGD.
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
TECHNICAL SPECIFICATIONS			
Item No.	PROCUREMENT OF 1 UNIT DIAGNOSTIC AUDIOMETER WITH SPEECH TEST AND FREEFIELD TEST (ADULT)	Qty./Unit	1 UNIT
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 850,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
TECHNICAL SPECIFICATIONS: <ul style="list-style-type: none"> • Type: Air, Bone and Speech Audiometer • Display: At least 4-inch Color Display • Resolution: At least 480 x 272 pixels • Weight: Manufacturer's Standard • Dimension: Manufacturer's Standard • Has Speech Test • Has Freefield Test • Has Comfortable Ergonomic Design • Printer: At least 3" thermal printer or laptop (i5, windows 11, 16GB RAM) with mouse, keyboard and external printer • Power Supply: 220V, 50/60Hz 			

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
PROCUREMENT OF 1 UNIT DIAGNOSTIC AUDIOMETER WITH SPEECH TEST AND FREEFIELD TEST (ADULT)	1 unit	Mandaluyong City Medical Center	30-60 calendar days upon receipt of Notice to Proceed.

Name of Company: _____

Address: _____

Signature Over Printed Name : _____

Telephone/Fax Number : _____

Email: _____