



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2024-002E
PROCUREMENT OF 1 UNIT AIR SPRAY PULVERIZATOR MACHINE
(SHORT OF AWARD)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
• 5 liters Disinfection Solution	5 liters Disinfection Solution (Packaging in 1 liter opaque bottle)

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on November 20, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 14th day of November, 2023 in MMCHD.

Approved by:

SGD
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	AIR SPRAY PULVERIZATOR MACHINE	Qty./Unit	1 UNIT
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 480,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> • Spray by pulverization composition of Hydrogen Peroxide and silver in the form of particles less than 5 micron • No irritation, no odor, no residue, high material compatibility, environmental friendly, safe for use • Air test strip for disinfection validation • Provides high level disinfection, kills spores at surfaces. • Capable to disinfect room volume from 1 to 999 cubic meters • Disinfection time not more than 3 minutes for 50 cubic meters • The treated room can be re-used in 30-60 minutes • Programmable time to start and time to stop • Automatic printing of disinfection data through printer • Stored in memory up to 100 room data cycle and printable if desire • With RF remote control to start and paused a disinfection cycle • The operation can be paused and resumed • Adjustable screen brightness • Password protected • Visible and audible warning system for start and end of operation. • Portable and transferable with plastic grip • Particle filter keeps the air access system clean • ABS body with 304 grade stainless steel metal parts • Conforms to LVD and EMC directives, 93/42 EEC Medical Device Directive Class I • 220-240 VAC, 50/60Hz, 1100 W with 2 meter long power supply cable <p>Disinfectant Solution Specification</p> <ul style="list-style-type: none"> • Disinfectant solution composed of 6% Hydrogen Peroxide with .003% colloidal silver and other auxiliary substances • Aldehyde and phenol Free • 50 pcs monitoring strips • 5 rolls thermal printer paper • 5 liters Disinfection Solution (Packaging in 1 liter opaque bottle) • Packaging in 1 liter opaque bottle 			
<p>Requirements if awarded the Contract</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service.</p>			

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
AIR SPRAY PULVERIZATOR MACHINE	1 unit	Health Facilities in The City of Marikina	60 calendar days upon receipt of the Notice to Proceed

Name of Company: _____

Address: _____

Signature Over Printed Name : _____

Telephone/Fax Number : _____

Email: _____