



**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB2024 – 052E**

**PROCUREMENT OF DENTAL X-RAY, WALL MOUNTED (REBID)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
No changes stipulated in the Technical Specifications	

The MMCHD will hold an official activity on July 9, 2024. Suppliers who would want to get the bidding documents are welcome to visit our BAC Secretariat Office until July 8, 2024.

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 10<sup>th</sup> day of July 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 2<sup>nd</sup> day of July 2024 in MMCHD

Approved by:

**JEREMIAS FRANCIS Y. CHAN, MD**  
Licensing Officer V / BAC Chairperson





**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB2024 – 052E**

**PROCUREMENT OF DENTAL X-RAY, WALL MOUNTED (REBID)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>ORIGINAL TECHNICAL SPECIFICATIONS</b>	<b>AMENDED</b>
No changes stipulated in the Technical Specifications	

The MMCHD will hold an official activity on July 9, 2024. Suppliers who would want to get the bidding documents are welcome to visit our BAC Secretariat Office until July 8, 2024.

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 10<sup>th</sup> day of July 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 2<sup>nd</sup> day of July 2024 in MMCHD

Approved by:

**JEREMIAS FRANCIS Y. CHAN, MD**  
Licensing Officer V / BAC Chairperson



TECHNICAL SPECIFICATIONS			
Item No. 1	DENTAL X-RAY, WALL MOUNTED	Qty./Unit	2 UNITS
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 300,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<b>TECHNICAL SPECIFICATIONS:</b> <ul style="list-style-type: none"> <li>• X-ray Plug: Standard 3 Electrode</li> <li>• Power: 220V, 50/60 Hz</li> <li>• Focus: 0.7mm or better</li> <li>• Anodic Angle: 19°</li> <li>• Tube Voltage: 70kv±10%</li> <li>• Anodic Electric Current: 7mA±20%</li> <li>• Total Filtration: ≥2.1mm AL</li> <li>• Connection: Coaxial Plug</li> </ul>			
Requirements if awarded the Contract <ol style="list-style-type: none"> <li>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.</li> <li>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. <b>Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. <b>Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>5. <b>Notarized</b> undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</li> <li>6. <b>Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:               <ol style="list-style-type: none"> <li>a) Service manual in English language</li> <li>b) Operation manual in English language</li> </ol> </li> <li>7. With "<b>DOH-MMCHD HFEP</b>" (Government Property not for sale) sticker in each unit</li> </ol>			
Source of Fund: SAA 2024-02-001002 (HFEP 2024) Recipient: Health Facilities in Marikina City			