



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB2024 – 056E
PROCUREMENT OF HOSPITAL CART (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
No changes in the Technical Specifications		

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

TECHNICAL SPECIFICATIONS			
Item No. 1	EMERGENCY CART WITH CARDIAC BOARD	Qty./Unit	7 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 840,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> • EG steel plate construction with scratch & chip resistant epoxy powder coating • With transparent top allowing visuals into the first drawer • With at least five (5) drawers with seal lock system and ABS molded bins in 3 drawers • With cardiac board located at back of cart • Equipped with protective bumper all around • Has provision for oxygen tank and an IV pole • With an external tray for defibrillator • Castors: four 125mm swivel castors with all brakes • Dimensions: Manufacturer's Standard 			
<p>REQUIREMENTS, if awarded the contract</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p>4. Warranty:</p> <p>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</p> <p>c) Corrective maintenance within five (5) calendar days upon</p>			

