



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB2024 – 057E PROCUREMENT OF 14 UNITS WHEELED STRETCHER, HEAVY DUTY

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

I. Query during Pre-bidding Conference:

Technical Specifications	Query	Response of the End User Unit
Backrest adjustable by gas spring	Backrest adjustable by gas spring or hydraulic – for clarification with end user	Backrest adjustable by gas spring or hydraulic – GRANTED
Two ways trendelenburg / tilting assisted by gas spring mechanism	Two ways trendelenburg / tilting assisted by gas spring or hydraulic mechanism	Two ways trendelenburg / tilting assisted by gas spring or hydraulic mechanism GRANTED
PVC bumper guard rail	PVC or ABS bumper guard rail – for clarification with end user	PVC or ABS bumper guard rail – GRANTED
Collapsible side rails	Collapsible or glide away side rails	Collapsible or glide away side rails GRANTED
Material: Mild steel framework	Material: Mild steel framework or better	Material: Mild steel framework or better GRANTED
Finishing: Epoxy powder coating	Finishing: Epoxy powder coating or better	Finishing: Epoxy powder coating or better GRANTED
Backrest: 0° - 90°	Backrest: 0° - 70° to 90°	Backrest: 0° - 70° or better
Trendelenburg: 0° - 12° Reverse Trendelenburg: 0° -6°	Trendelenburg: 0° - 12° or better Reverse Trendelenburg: 0° -6° or better	Trendelenburg: 0° - 12° or better GRANTED Reverse Trendelenburg: 0° -6° or better GRANTED
Approx dimension (Length): 2100mm (± 50mm)	Approx dimension (Length): 2100mm (± 70mm)	Approx dimension (Length): 2100mm (± 70mm) GRANTED


Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD
 Licensing Officer V / BAC Chairperson



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB2024 – 057E PROCUREMENT OF 14 UNITS WHEELED STRETCHER, HEAVY DUTY

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Backrest adjustable by gas spring	Backrest adjustable by gas spring or hydraulic – for clarification with end user	Backrest adjustable by gas spring or hydraulic – GRANTED
Two ways trendelenburg / tilting assisted by gas spring mechanism	Two ways trendelenburg / tilting assisted by gas spring or hydraulic mechanism	Two ways trendelenburg / tilting assisted by gas spring or hydraulic mechanism GRANTED
PVC bumper guard rail	PVC or ABS bumper guard rail – for clarification with end user	PVC or ABS bumper guard rail – GRANTED
Collapsible side rails	Collapsible or glide away side rails	Collapsible or glide away side rails GRANTED
Material: Mild steel framework	Material: Mild steel framework or better	Material: Mild steel framework or better GRANTED
Finishing: Epoxy powder coating	Finishing: Epoxy powder coating or better	Finishing: Epoxy powder coating or better GRANTED
Backrest: 0° - 90°	Backrest: 0° - 70° to 90°	Backrest: 0° - 70° or better
Trendelenburg: 0° - 12° Reverse Trendelenburg: 0° -6°	Trendelenburg: 0° - 12° or better Reverse Trendelenburg: 0° -6° or better	Trendelenburg: 0° - 12° or better GRANTED Reverse Trendelenburg: 0° -6° or better GRANTED
Approx dimension (Length): 2100mm (± 50mm)	Approx dimension (Length): 2100mm (± 70mm)	Approx dimension (Length): 2100mm (± 70mm) GRANTED


Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer / BAC Chairperson

TECHNICAL SPECIFICATIONS			
Item No. 1	WHEELED STRETCHER, HEAVY DUTY	Qty./Unit	14 UNITS
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 4,200,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • Gas spring tilting • Hydraulic foot pedal-controlled Hi-Lo position • <i>Backrest adjustable by gas spring or hydraulic</i> • <i>Two ways trendelenburg / tilting assisted by gas spring or hydraulic mechanism</i> • <i>PVC or ABS bumper guard rail</i> • Push handle at head and both ends • <i>Collapsible or glide away side rails</i> • Two-hook stainless steel IV pole • Storage utility tray • Mattress must have cover with fire retardant, anti-bacterial, anti-static, stain resistant, urine resistant and waterproof • Oxygen tank holder • <i>Material: Mild steel framework or better</i> • <i>Finishing: Epoxy powder coating or better</i> • Safety load: 250kg or better • <i>Backrest: 0° - 70° or better</i> • <i>Trendelenburg: 0° - 12° or better</i> • <i>Reverse Trendelenburg: 0° -6° or better</i> • Castor Dim: at least 200mm central locking • Mattress Size: at least 50mm thick or better • <i>Approx dimension (Length): 2100mm (± 70mm)</i> • <i>Approx dimension (Width): 550-870mm</i> 			
Requirements if awarded the Contract <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical 			

TECHNICAL SPECIFICATIONS			
Item No. 1	WHEELED STRETCHER, HEAVY DUTY	Qty./Unit	14 UNITS
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 4,200,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
Technical Specifications: <ul style="list-style-type: none"> • Gas spring tilting • Hydraulic foot pedal-controlled Hi-Lo position • <i>Backrest adjustable by gas spring or hydraulic</i> • <i>Two ways trendelenburg / tilting assisted by gas spring or hydraulic mechanism</i> • <i>PVC or ABS bumper guard rail</i> • Push handle at head and both ends • <i>Collapsible or glide away side rails</i> • Two-hook stainless steel IV pole • Storage utility tray • Mattress must have cover with fire retardant, anti-bacterial, anti-static, stain resistant, urine resistant and waterproof • Oxygen tank holder • <i>Material: Mild steel framework or better</i> • <i>Finishing: Epoxy powder coating or better</i> • Safety load: 250kg or better • <i>Backrest: 0° - 70° or better</i> • <i>Trendelenburg: 0° - 12° or better</i> • <i>Reverse Trendelenburg: 0° -6° or better</i> • Castor Dim: at least 200mm central locking • Mattress Size: at least 50mm thick or better • <i>Approx dimension (Length): 2100mm (± 70mm)</i> • <i>Approx dimension (Width): 550-870mm</i> 			
Requirements if awarded the Contract <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical 			

