



**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB2024 – 059E**  
**PROCUREMENT OF SUCTION MACHINE, HIGH FLOW RATE**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:			
Technical Specifications	Query	Response of the End User Unit	
With anti-vibration system that allows no vibration and low noise level	With anti-vibration system that allows no vibration and low noise level 45 decibels below	With anti-vibration system that allows no vibration and low noise level 45 decibels below GRANTED	
Maximum vacuum pressure: 735 mmHg	Maximum vacuum pressure: 710 mmHg to 750 mmHg	Maximum vacuum pressure: 710 mmHg to 750 mmHg GRANTED	

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7<sup>th</sup> day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD**  
Licensing Officer V / BAC Chairperson



**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB2024 – 059E**

**PROCUREMENT OF SUCTION MACHINE, HIGH FLOW RATE**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:			
Technical Specifications	Query	Response of the End User Unit	
With anti-vibration system that allows no vibration and low noise level	With anti-vibration system that allows no vibration and low noise level 45 decibels below	With anti-vibration system that allows no vibration and low noise level 45 decibels below GRANTED	
Maximum vacuum pressure: 735 mmHg	Maximum vacuum pressure: 710 mmHg to 750 mmHg	Maximum vacuum pressure: 710 mmHg to 750 mmHg GRANTED	

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7<sup>th</sup> day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD**  
Licensing Officer V / BAC Chairperson

**TECHNICAL SPECIFICATIONS**

Item No. 1	<b>SUCTION MACHINE, HIGH FLOW RATE</b>	Qty./Unit	<b>13 UNITS</b>
------------	--	-----------	-----------------

Name of Manufacturer:	Country of Origin
-----------------------	-------------------

Brand:	Model: (if applicable)
--------	------------------------

ABC: <b>1,950,000.00</b>	<b>STATEMENT OF COMPLIANCE</b>
--------------------------	--------------------------------

<p align="center"><b>PURCHASER'S SPECIFICATION</b></p> <p>Technical Specifications:</p> <ul style="list-style-type: none"> <li>• Oil-less piston type vacuum pump</li> <li>• Suction body with ABS plastic material composition</li> <li>• <i>With anti-vibration system that allows no vibration and low noise level 45 decibels below</i></li> <li>• With two-step overflow prevention system</li> <li>• With silicon tube for outer and inner tubing</li> <li>• With footswitch control</li> <li>• With 2 units of autoclavable 3.0L polycarbonate bottle</li> <li>• <i>Maximum vacuum pressure: 710 mmHg to 750 mmHg</i></li> <li>• Flow rate: 50L/min</li> <li>• Dimensions: Manufacturer's Standard</li> <li>• Weight: Manufacturer's Standard</li> <li>• Supplied with extra 2 pcs filter and pcs silicone hose</li> </ul>	
---	--

<p><b>REQUIREMENTS</b>, if awarded the contract</p> <p><b>1. Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>60</b> calendar days upon receipt of Notice to Proceed.</p> <p><b>2. Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p><b>3. Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p><b>4. Warranty:</b></p> <p>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</p> <p>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</p> <p>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</p> <p>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the</p>	
---	--



**TECHNICAL SPECIFICATIONS**

Item No. 1	<b>SUCTION MACHINE, HIGH FLOW RATE</b>	Qty./Unit	<b>13 UNITS</b>
Name of Manufacturer:		Country of Origin	
Brand: ABC: <b>1,950,000.00</b>		Model: (if applicable)	
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> <li>• Oil-less piston type vacuum pump</li> <li>• Suction body with ABS plastic material composition</li> <li>• <i>With anti-vibration system that allows no vibration and low noise level 45 decibels below</i></li> <li>• With two-step overflow prevention system</li> <li>• With silicon tube for outer and inner tubing</li> <li>• With footswitch control</li> <li>• With 2 units of autoclavable 3.0L polycarbonate bottle</li> <li>• <i>Maximum vacuum pressure: 710 mmHg to 750 mmHg</i></li> <li>• Flow rate: 50L/min</li> <li>• Dimensions: Manufacturer's Standard</li> <li>• Weight: Manufacturer's Standard</li> <li>• Supplied with extra 2 pcs filter and pcs silicone hose</li> </ul>			
<p><b>REQUIREMENTS</b>, if awarded the contract</p> <p>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>60</b> calendar days upon receipt of Notice to Proceed.</p> <p>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p>4. <b>Warranty:</b></p> <p style="padding-left: 20px;">a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p style="padding-left: 20px;">b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</p> <p style="padding-left: 20px;">c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</p> <p style="padding-left: 20px;">d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</p> <p style="padding-left: 20px;">e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the</p>			

