



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB2024 – 063E
PROCUREMENT OF CAUTERY MACHINE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Maximum CUT output: 300 watts at 500 Ohms	Maximum CUT output: 300 – 400 watts at 300 - 500 Ohms	Maximum CUT output: 300 – 400 watts at 300 - 500 Ohms GRANTED
Maximum COAG output: up to 200 watts	Maximum COAG output: 120 - 200 watts	Maximum COAG output: 120 - 200 watts - GRANTED
Bipolar cut: up to 100W ±20% at 500Ω	Bipolar cut: up to 70W - 100W ±20% at 100- 500Ω	Bipolar cut: up to 70W - 100W ±20% at 100- 500Ω GRANTED
Bipolar coag: up to 120W ±20% at 75Ω	Bipolar coag: 70w - 120W ±20% at 75Ω - 100Ω	Bipolar coag: 70w - 120W ±20% at 75Ω - 100Ω GRANTED

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB2024 – 063E
PROCUREMENT OF CAUTERY MACHINE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Maximum CUT output: 300 watts at 500 Ohms	Maximum CUT output: 300 – 400 watts at 300 - 500 Ohms	Maximum CUT output: 300 – 400 watts at 300 - 500 Ohms GRANTED
Maximum COAG output: up to 200 watts	Maximum COAG output: 120 - 200 watts	Maximum COAG output: 120 - 200 watts - GRANTED
Bipolar cut: up to 100W ±20% at 500Ω	Bipolar cut: up to 70W - 100W ±20% at 100- 500Ω	Bipolar cut: up to 70W - 100W ±20% at 100- 500Ω GRANTED
Bipolar coag: up to 120W ±20% at 75Ω	Bipolar coag: 70w - 120W ±20% at 75Ω - 100Ω	Bipolar coag: 70w - 120W ±20% at 75Ω - 100Ω GRANTED

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

TECHNICAL SPECIFICATIONS			
Item No. 1	CAUTERY MACHINE	Qty./Unit	2 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 3,000,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • Automatic power adjustment: Voltage, power and arc control • Separate adjustment of power limitation and effects • Variable configuration with at least four (4) socket modules • Activation either via hand piece and also via foot switch • <i>Maximum CUT output: 300 – 400 watts at 300 - 500 Ohms</i> <ul style="list-style-type: none"> • <i>Maximum COAG output: 120 - 200 watts</i> • <i>Bipolar cut: up to 70W - 100W ±20% at 100- 500Ω</i> • <i>Bipolar coag: 70w - 120W ±20% at 75Ω - 100Ω</i> • Automatic stop function for bipolar coagulation • Automatic Start function for Bipolar Modes • Modes for cutting: Auto cut, Bipolar cut or equivalent • Modes for coag: Classic coag, Forced coag, Spray coag, bipolar soft coag or equivalent • With safety system for monopolar surgery using split plate with outer ring • Dimensions: Manufacturer's Standard • Accessories and inclusions: <ul style="list-style-type: none"> - Electrosurgical pencil with connecting cable - Patient plate: split, with contact surface 168cm² with connecting cable - Two-pedal foot switch with connecting cable at least 5 meters - One-pedal foot switch with connecting cable at least 5 meters - Universal cart with four 100mm castors with 2 front conductive fixing with brake 			
<p>REQUIREMENTS, if awarded the contract</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 90 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p>4. Warranty:</p> <p>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</p> <p>c) Corrective maintenance within five (5) calendar days upon</p>			

TECHNICAL SPECIFICATIONS

Item No. 1	CAUTERY MACHINE	Qty./Unit	2 Units
Name of Manufacturer:		Country of Origin	
Brand: ABC: 3,000,000.00		Model: (if applicable)	
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • Automatic power adjustment: Voltage, power and arc control • Separate adjustment of power limitation and effects • Variable configuration with at least four (4) socket modules • Activation either via hand piece and also via foot switch • <i>Maximum CUT output: 300 – 400 watts at 300 - 500 Ohms</i> <ul style="list-style-type: none"> • <i>Maximum COAG output: 120 - 200 watts</i> • <i>Bipolar cut: up to 70W - 100W ±20% at 100- 500Ω</i> • <i>Bipolar coag: 70w - 120W ±20% at 75Ω - 100Ω</i> • Automatic stop function for bipolar coagulation • Automatic Start function for Bipolar Modes • Modes for cutting: Auto cut, Bipolar cut or equivalent • Modes for coag: Classic coag, Forced coag, Spray coag, bipolar soft coag or equivalent • With safety system for monopolar surgery using split plate with outer ring • Dimensions: Manufacturer's Standard • Accessories and inclusions: <ul style="list-style-type: none"> - Electrosurgical pencil with connecting cable - Patient plate: split, with contact surface 168cm² with connecting cable - Two-pedal foot switch with connecting cable at least 5 meters - One-pedal foot switch with connecting cable at least 5 meters - Universal cart with four 100mm castors with 2 front conductive fixing with brake 			
<p>REQUIREMENTS, if awarded the contract</p> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 90 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment. 4. Warranty: <ol style="list-style-type: none"> a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations; c) Corrective maintenance within five (5) calendar days upon 			

notification from the end-user regarding equipment breakdown/defects.

d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.

e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.

5. **Notarized undertaking** that the supplier shall conduct the necessary corrective maintenance, replacements and repair within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

a) Service manual in English language

b) Operation manual in English language

7. With "**DOH-MMCHD HFEP (Government Property not for sale)**" sticker in each unit.

X-X

Source of Fund: SAA 2024-02-001002 (HFEP 2024)

Recipient: Taguig-Pateros District Hospital

