



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB No. 2024 – 071E
PROCUREMENT OF OPERATING TABLE (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:

Technical Specifications	Query	Response of the End User Unit
Lot Bidding	Itemized Bidding	Itemized Bidding GRANTED (Individual Bidding Documents should be submitted in each item)
ITEM NO 1: Operating Table Electric		
Title of Item no. 1 Patient Monitor Electric	Operating Table Electric	Operating Table Electric
Modularised radiolucent table top with a pre-release click mechanism	Modularised radiolucent table top with a pre-release click mechanism or its equivalent	Modularised radiolucent table top with a pre-release click mechanism or its equivalent GRANTED
Head rest with manual servo-assisted continuous double adjustment function	Head rest with manual servo-assisted continuous double adjustment function for clarification with end user	Head rest with manual servo-assisted continuous double adjustment function - Retained the original specs
detachable and interchangeable at the push of a button	detachable and interchangeable at the push of a button for clarification with end user	detachable and interchangeable at the push of a button Retained the original specs
Two-part leg section with individual manual swiveling - with electro-hydraulic and servo-assisted manual tilting	Two-part leg section with individual manual swiveling - with electro-hydraulic and servo-assisted manual tilting	Two-part leg section with individual manual swiveling - with electro-hydraulic and servo-assisted manual tilting - Retained the original specs
Tilting angle: at least +10°/ - 90° or better	Tilting angle: at least +10°/ - 90° or better	Tilting angle: at least +10°/ - 90° or better - Retained the original specs
Electropolished stainless steel 18/10 frame construction, disinfectant-proof	Electropolished stainless steel 18/10 frame construction, disinfectant-proof	Electropolished stainless steel 18/10 frame construction, disinfectant-proof Retained the original specs
Mattresses of viscoelastic foam	Mattresses of viscoelastic or polyurethane foam - for clarification with end user	Mattresses of viscoelastic or polyurethane foam - Retained the original specs
Must be equipped with the latest and updateable software	Must be equipped with the latest and updateable	Must be equipped with the latest and updateable



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB No. 2024 – 071E
PROCUREMENT OF OPERATING TABLE (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Lot Bidding	Itemized Bidding	Itemized Bidding GRANTED (Individual Bidding Documents should be submitted in each item)
ITEM NO 1: Operating Table Electric		
Title of Item no. 1 Patient Monitor Electric	Operating Table Electric	Operating Table Electric
Modularised radiolucent table top with a pre-release click mechanism	Modularised radiolucent table top with a pre-release click mechanism or its equivalent	Modularised radiolucent table top with a pre-release click mechanism or its equivalent GRANTED
Head rest with manual servo-assisted continuous double adjustment function	Head rest with manual servo-assisted continuous double adjustment function for clarification with end user	Head rest with manual servo-assisted continuous double adjustment function – Retained the original specs
detachable and interchangeable at the push of a button	detachable and interchangeable at the push of a button for clarification with end user	detachable and interchangeable at the push of a button Retained the original specs
Two-part leg section with individual manual swiveling - with electro-hydraulic and servo-assisted manual tilting	Two-part leg section with individual manual swiveling - with electro-hydraulic and servo-assisted manual tilting	Two-part leg section with individual manual swiveling - with electro-hydraulic and servo-assisted manual tilting - Retained the original specs
Tilting angle: at least +10°/ - 90° or better	Tilting angle: at least +10°/ - 90° or better	Tilting angle: at least +10°/ - 90° or better – Retained the original specs
Electropolished stainless steel 18/10 frame construction, disinfectant-proof	Electropolished stainless steel 18/10 frame construction, disinfectant-proof	Electropolished stainless steel 18/10 frame construction, disinfectant-proof Retained the original specs
Mattresses of viscoelastic foam	Mattresses of viscoelastic or polyurethane foam – for clarification with end user	Mattresses of viscoelastic or polyurethane foam – Retained the original specs
Must be equipped with the latest and updateable software	Must be equipped with the latest and updateable	Must be equipped with the latest and updateable

technology	software technology - for clarification with end user	software technology Retained the original specs
Cable-connected and also wireless foot controls for all electric table top movements	Cable-connected and also wireless foot controls for all electric table top movements	Cable-connected and also wireless foot controls for all electric table top movements - Retained the original specs
4 intragrated rechargeable batteries with a capacity of 100 surgical interventions	4 integrated rechargeable batteries with a capacity of 60 — 100 surgical interventions- for clarification with end user	4 integrated rechargeable batteries with a capacity of 60 - 100 surgical interventions- Retained the original specs
with central brake pedal	with central brake pedal or better	with central brake pedal or better GRANTED
- 1 unit leg section	1 unit pair section - for clarification with end user	1 unit pair section - Retained the original specs
ITEM NO2: Operating Table, Bariatric, Electrohydraulic		
Extension Device	Extension Device for clarification with end user	Extension Device Retained the original specs
Mounting arbor	Mounting arbor for clarification with end user	Mounting arbor - Retained the original specs
Transport trolley	Transport trolley for clarification with end user	Transport trolley Retained the original specs
Kirschner Wire device	Kirschner Wire device for clarification with end user	Kirschner Wire device Retained the original specs
Side rail extension (pair)	Side rail extension (pair) for clarification with end user	Side rail extension (pair) Retained the original specs
ITEM NO 3: Operating Table, Manual Hydraulic		
Headrest with servo-assisted continuous double adjustment function, detachable and interchangeable at the push of a button	Headrest with servo-assisted continuous double adjustment function, detachable and interchangeable at the push of a button for clarification with end user	Headrest with servo-assisted continuous double adjustment function, detachable and interchangeable at the push of a button Retained the original specs
Two-part leg section with individual manual swiveling and servo assisted tilting, detachable and interchangeable at the push of a button:	Two-part leg section with individual manual swiveling and servo assisted tilting, detachable and interchangeable at the push of a button: for clarification with end user	Two-part leg section with individual manual swiveling and servo assisted tilting, detachable and interchangeable at the push of a button: - Retained the original specs
Tilting angle: +20°/ -90° -	Tilting angle: +20°/ -90° - for clarification with end user	Tilting angle: +20°/ -90° - Retained the original specs
Electropolished stainless steel 18/10 frame construction, disinfectant-proof -	Electropolished stainless steel 18/10 frame construction, disinfectant-proof - for clarification with end user	Electropolished stainless steel 18/10 frame construction, disinfectant-proof - Retained the original specs
Tabletop dimension (Length):	Tabletop dimension (Length):	Tabletop dimension (Length):

at least 2020mm – 2100mm (± 50mm) -	at least 2020mm – 2100mm (± 50mm) – for clarification with end user	at least 2020mm – 2100mm (± 50mm) – Retained the original specs
Electropolished stainless steel 18/10 frame construction, disinfectant-proof -	Electropolished stainless steel 18/10 frame construction, disinfectant-proof – for clarification with end user	Electropolished stainless steel 18/10 frame construction, disinfectant-proof – Retained the original specs
Tabletop dimension (Length): at least 2020mm – 2100mm (± 50mm) -	Tabletop dimension (Length): at least 2020mm – 2100mm (± 50mm) – for clarification with end user	Tabletop dimension (Length): at least 2020mm – 2100mm (± 50mm) – retained the original specs

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD
 Licensing Officer V / BAC Chairperson *uf*

technology	software technology - for clarification with end user	software technology Retained the original specs
Cable-connected and also wireless foot controls for all electric table top movements	Cable-connected and also wireless foot controls for all electric table top movements	Cable-connected and also wireless foot controls for all electric table top movements - Retained the original specs
4 intragrated rechargeable batteries with a capacity of 100 surgical interventions	4 integrated rechargeable batteries with a capacity of 60— 100 surgical interventions- for clarification with end user	4 integrated rechargeable batteries with a capacity of 60 - 100 surgical interventions- Retained the original specs
with central brake pedal	with central brake pedal or better	with central brake pedal or better GRANTED
- 1 unit leg section	1 unit pair section - for clarification with end user	1 unit pair section - Retained the original specs
ITEM NO2: Operating Table, Bariatric, Electrohydraulic		
Extension Device	Extension Device for clarification with end user	Extension Device Retained the original specs
Mounting arbor	Mounting arbor for clarification with end user	Mounting arbor - Retained the original specs
Transport trolley	Transport trolley for clarification with end user	Transport trolley Retained the original specs
Kirschner Wire device	Kirschner Wire device for clarification with end user	Kirschner Wire device Retained the original specs
Side rail extension (pair)	Side rail extension (pair) for clarification with end user	Side rail extension (pair) Retained the original specs
ITEM NO 3: Operating Table, Manual Hydraulic		
Headrest with servo-assisted continuous double adjustment function, detachable and interchangeable at the push of a button	Headrest with servo-assisted continuous double adjustment function, detachable and interchangeable at the push of a button for clarification with end user	Headrest with servo-assisted continuous double adjustment function, detachable and interchangeable at the push of a button Retained the original specs
Two-part leg section with individual manual swiveling and servo assisted tilting, detachable and interchangeable at the push of a button:	Two-part leg section with individual manual swiveling and servo assisted tilting, detachable and interchangeable at the push of a button: for clarification with end user	Two-part leg section with individual manual swiveling and servo assisted tilting, detachable and interchangeable at the push of a button: - Retained the original specs
Tilting angle: +20°/ -90° -	Tilting angle: +20°/ -90° - for clarification with end user	Tilting angle: +20°/ -90° - Retained the original specs
Electropolished stainless steel 18/10 frame construction, disinfectant-proof -	Electropolished stainless steel 18/10 frame construction, disinfectant-proof - for clarification with end user	Electropolished stainless steel 18/10 frame construction, disinfectant-proof - Retained the original specs
Tabletop dimension (Length):	Tabletop dimension (Length):	Tabletop dimension (Length):

at least 2020mm – 2100mm (± 50mm) -	at least 2020mm – 2100mm (± 50mm) - for clarification with end user	at least 2020mm – 2100mm (± 50mm) - Retained the original specs
Electropolished stainless steel 18/10 frame construction, disinfectant-proof -	Electropolished stainless steel 18/10 frame construction, disinfectant-proof - for clarification with end user	Electropolished stainless steel 18/10 frame construction, disinfectant-proof - Retained the original specs
Tabletop dimension (Length): at least 2020mm – 2100mm (± 50mm) -	Tabletop dimension (Length): at least 2020mm – 2100mm (± 50mm) - for clarification with end user	Tabletop dimension (Length): at least 2020mm – 2100mm (± 50mm) - retained the original specs

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD
 Licensing Officer V / BAC Chairperson