



SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB2024 – 074E
 PROCUREMENT OF BILIRUBIN (LOT BIDDING)**

his Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Item No. 1 Bilirubin Blanket		
No changes stipulated in the Technical Specifications		
Item No. 2 Bilirubin Light		
with 2 levels adjustable intensity	with 2 levels adjustable intensity or better	with 2 levels adjustable intensity or better GRANTED
Effective Area: 40 x 20 cm	Effective Area: 40 x 20 cm or its equivalent	Effective Area: 40 x 20 cm or its equivalent
•LED lifetime: 20,000 hours	•LED lifetime: 20,000 hours or better	LED lifetime: 20,000 hours or better
Wavelength: peak between 450~470nm	Wavelength: peak between 450~470nm or better	Wavelength: peak between 450~470nm or better
Audible Noise: <60dB	Audible Noise: <60dB or better	Audible Noise: <60dB or better

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:

JEREMIAS FRANCIS Y. CHAN, MD
 Licensing Officer V / BAC Chairperson



SUPPLEMENTAL/ BID BULLETIN NO. 1

**IB2024 – 074E
 PROCUREMENT OF BILIRUBIN (LOT BIDDING)**

his Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:

Technical Specifications	Query	Response of the End User Unit
Item No. 1 Bilirubin Blanket		
No changes stipulated in the Technical Specifications		
Item No. 2 Bilirubin Light		
with 2 levels adjustable intensity	with 2 levels adjustable intensity or better	with 2 levels adjustable intensity or better GRANTED
Effective Area: 40 x 20 cm	Effective Area: 40 x 20 cm or its equivalent	Effective Area: 40 x 20 cm or its equivalent
•LED lifetime: 20,000 hours	•LED lifetime: 20,000 hours or better	LED lifetime: 20,000 hours or better
Wavelength: peak between 450~470nm	Wavelength: peak between 450~470nm or better	Wavelength: peak between 450~470nm or better
Audible Noise: <60dB	Audible Noise: <60dB or better	Audible Noise: <60dB or better

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:

JEREMIAS FRANCIS Y. CHAN, MD
 Licensing Officer V / BAC Chairperson

TECHNICAL SPECIFICATIONS

Item No. 1	BILIRUBIN BLANKET	Qty./Unit	10 Units
Name of Manufacturer:		Country of Origin	
Brand: ABC: 1,600,000.00		Model: (if applicable)	
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • Type: Blue LED • Wavelength: 455 to 465 nm • Effective surface area: at least 412 mm x 102 mm or better • Intensity control: <ul style="list-style-type: none"> • Low: 30 ± 10 µW/cm²/nm or better • High: 60 ± 10 µW/cm²/nm or better • Timer mode: 30 mins to 24 hrs • Display: treatment time, remaining time, total using time • Screen: TFT Color LCD • Power: 220V/ 60Hz • Battery operation time: 10 hrs • Battery charging time: 4 hrs • Dimensions: Manufacturer's Standard • Weight: Manufacturer's Standard • Accessories: <ul style="list-style-type: none"> - Eye protectors - Disposable covers (5 pieces) 			
<p>REQUIREMENTS, if awarded the contract</p> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment. 4. Warranty: <ol style="list-style-type: none"> a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations; c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects. d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period. e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance, replacements and repair within five (5) calendar days upon notification of the equipment breakdown from the end-user. 			

<p>The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"> a) Service manual in English language b) Operation manual in English language <p>7. With "DOH-MMCHD HFEP (Government Property not for sale)" sticker in each unit.</p> <p>X-X</p> <p>Source of Fund: SAA 2024-02-001002 (HFEP 2024)</p> <p>Recipient: Taguig-Pateros District Hospital</p>	
--	--

TECHNICAL SPECIFICATIONS			
Item No. 2	BILIRUBIN LIGHT	Qty./Unit	5 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 750,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • With colored TFT LCD display • Must have flexible neck • <i>with 2 levels adjustable intensity or better</i> • With timer function • Light Source: <ul style="list-style-type: none"> - Type: Blue LED - Variation in intensity over 6 hours: $\pm 10\%$ - <i>Effective Area: 40 x 20 cm or its equivalent</i> - <i>Wavelength: peak between 450~470nm or better</i> • Heat output: Over 6 hours: $< 10^{\circ}\text{C}$ • Dimension: Manufacturer's Standard • <i>Audible Noise: <60dB or better</i> • <i>LED lifetime: 20,000 hours or better</i> 			
<p>REQUIREMENTS, if awarded the contract</p> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment. 4. Warranty: <ol style="list-style-type: none"> a) Warranty certificate for two (2) years on parts and service. The 			

supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;

c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.

d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.

e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.

5. **Notarized undertaking** that the supplier shall conduct the necessary corrective maintenance, replacements and repair within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

a) Service manual in English language

b) Operation manual in English language

7. With "**DOH-MMCHD HFEP** (Government Property not for sale)" sticker in each unit.

X-X

Source of Fund: SAA 2024-02-001002 (HFEP 2024)

Recipient: Taguig-Pateros District Hospital

TECHNICAL SPECIFICATIONS			
Item No. 1	BILIRUBIN BLANKET	Qty./Unit	10 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,600,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
Technical Specifications: <ul style="list-style-type: none"> • Type: Blue LED • Wavelength: 455 to 465 nm • Effective surface area: at least 412 mm x 102 mm or better • Intensity control: <ul style="list-style-type: none"> • Low: 30 ± 10 µW/cm²/nm or better • High: 60 ± 10 µW/cm²/nm or better • Timer mode: 30 mins to 24 hrs • Display: treatment time, remaining time, total using time • Screen: TFT Color LCD • Power: 220V/ 60Hz • Battery operation time: 10 hrs • Battery charging time: 4 hrs • Dimensions: Manufacturer's Standard • Weight: Manufacturer's Standard • Accessories: <ul style="list-style-type: none"> - Eye protectors - Disposable covers (5 pieces) 			
REQUIREMENTS, if awarded the contract <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment. 4. Warranty: <ol style="list-style-type: none"> a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations; c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects. d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period. e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance, replacements and repair within five (5) calendar days upon notification of the equipment breakdown from the end-user. 			

supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;

c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.

d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.

e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.

5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance, replacements and repair within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

a) Service manual in English language

b) Operation manual in English language

7. With "**DOH-MMCHD HFEP** (Government Property not for sale)" sticker in each unit.

X-X

Source of Fund: SAA 2024-02-001002 (HFEP 2024)

Recipient: Taguig-Pateros District Hospital



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB2024 – 072E
PROCUREMENT OF ECG MACHINE (LOT BIDDING)

his Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Item no. 1 ECG Machine		
Sampling Frequency: 1 MHz	Sampling Frequency: 1 mHz	Sampling Frequency: 1 mHz
Item No. 2: Wireless Rest ECG		
Wireless Rest ECG	Wireless Rest ECG – Usual ECG but wireless	Wireless Rest ECG – Usual ECG but wireless GRANTED
• Low pass filter: 20, 35, 40 Hz	Low pass filter: 20-25, 35, 40-45 Hz	Low pass filter: 20-25, 35, 40-45 Hz

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:

JEREMIAS FRANCIS Y. CHAN, MD
 Licensing Officer V / BAC Chairperson

TECHNICAL SPECIFICATIONS			
Item No. 1	ECG MACHINE	Qty./Unit	7 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 1,400,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> • Standard 12 leads • Has color LCD touchscreen display • With full alphanumeric keyboard • Dimensions: Manufacturer's Standard • With internal storage of at least 200 ECGs • EMG Filter: 25 Hz/ 35 Hz/ 45 Hz/ off • LOWPass Filter: 150 Hz/ 100 Hz/ 75 Hz • With correct detection for failure electrodes • Supporting ECG waves displaying with grid • ECG data can be transmitted to the PC software thru serial cable, net cable or WIFI • With anti-noise technology • With defibrillation protection • <i>Sampling Frequency: 1 mHz</i> • With built-in rechargeable lithium battery • With internal thermal recorder • recording frequency response: $\leq 300\text{Hz}$ • Supplied with 10 pcs ECG paper 			
<p>REQUIREMENTS, if awarded the contract</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p>4. Warranty:</p> <p>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</p> <p>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</p> <p>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</p> <p>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</p>			

