

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB2024 – 075E
PROCUREMENT OF PATIENT WARMING SYSTEM

his Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:

Technical Specifications	Query	Response of the End User Unit
No changes stipulated in the Technical Specifications		

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB2024 – 075E
PROCUREMENT OF PATIENT WARMING SYSTEM

his Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:

Technical Specifications	Query	Response of the End User Unit
No changes stipulated in the Technical Specifications		


Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	PATIENT WARMING SYSTEM	Qty./Unit	1 Unit
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 700,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • Dimensions: Manufacturer's Standard • Weight: Manufacturer's Standard • Hose type: Flexible • Hose size: 80-inch length by 2-inches diameter • Electronically controlled • Time to reach operating temperature: 2 to 5 mins or better • Accessories and inclusions: <ul style="list-style-type: none"> - 20 pieces adult underbody - 20 pieces pediatric underbody - 20 pieces full body blanket - 20 pieces multiposition blanket 			
<p>REQUIREMENTS, if awarded the contract</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p>4. Warranty:</p> <p style="padding-left: 20px;">a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p style="padding-left: 20px;">b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</p> <p style="padding-left: 20px;">c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</p> <p style="padding-left: 20px;">d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</p> <p style="padding-left: 20px;">e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance, replacements and repair within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall</p>			

be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language.

7. With "**DOH-MMCHD HFEP** (Government Property not for sale)" sticker in each unit.

X-X

Source of Fund: SAA 2024-02-001002 (HFEP 2024)

Recipient: Taguig-Pateros District Hospital

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	PATIENT WARMING SYSTEM	Qty./Unit	1 Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 700,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • Dimensions: Manufacturer's Standard • Weight: Manufacturer's Standard • Hose type: Flexible • Hose size: 80-inch length by 2-inches diameter • Electronically controlled • Time to reach operating temperature: 2 to 5 mins or better • Accessories and inclusions: <ul style="list-style-type: none"> - 20 pieces adult underbody - 20 pieces pediatric underbody - 20 pieces full body blanket - 20 pieces multiposition blanket 			
<p>REQUIREMENTS, if awarded the contract</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 60 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p>4. Warranty:</p> <p style="padding-left: 20px;">a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p style="padding-left: 20px;">b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</p> <p style="padding-left: 20px;">c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</p> <p style="padding-left: 20px;">d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</p> <p style="padding-left: 20px;">e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance, replacements and repair within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall</p>			

be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

7. With "**DOH-MMCHD HFEP** (Government Property not for sale)" sticker in each unit.

X-X

Source of Fund: SAA 2024-02-001002 (HFEP 2024)

Recipient: Taguig-Pateros District Hospital