



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB2024 – 082E
PROCUREMENT OF 19 UNITS WHEELED DETACHABLE STRETCHER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
<p>Aluminum Alloy</p> <p>Adjustable height: 300mm to 850mm</p> <p>Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30-60 calendar days upon receipt of Notice to Proceed.</p> <p>Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment</p> <p>Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</p> <p>Manuals: The supplier</p>	<p>Aluminum Alloy or Stainless Steel or better</p> <p>Adjustable height: 300mm to 850mm or Manufacturer Standard</p> <p>Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed.</p> <p>3.Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</p> <p>6. Manuals: The supplier must provide the end user one (1)</p>	<p>Aluminum Alloy or Stainless Steel or better GRANTED</p> <p>Adjustable height: Manufacturer Standard GRANTED</p> <p>Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed. GRANTED</p> <p>Training: Removed - GRANTED</p> <p>Preventive maintenance - removed - GRANTED</p> <p>Preventive removed - GRANTED</p>

<p>must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Service manual in English language</p> <p>b) Operation manual in English language</p>	<p>hard and one (1) soft copy of the following:—</p> <p>— a) Service manual in English language—</p> <p>— b) Operation manual in English language</p>	
---	--	--


Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 24th day of September 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 16th day of September 2024 in MMCHD

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD
 Licensing Officer V / BAC Chairperson

Republic of the Philippines

Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	WHEELED DETACHABLE STRETCHER	Qty./Unit	19 UNITS
Name of Manufacturer:		Country of Origin:	
Brand:		Model: (if applicable)	
ABC: 1,330,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>TECHNICAL SPECIFICATIONS:</p> <p><i>Frame: Aluminum Alloy or Stainless Steel or better</i></p> <p>Type: Detachable/Removable</p> <p>Mattress: At least 4- inch thick, water proof seamless mattress</p> <p>Has 3 belts</p> <p>Castor wheels: At least 100 mm diameter</p> <p>Max Load: at least 160kg</p> <p>Dimensions (L x W x H): 2000mm x 600mm x 850mm, ± 100mm in every dimension, adjustable height must be at least in range or bigger</p> <p>Adjustable height: 300mm to 850mm</p>			
<p>Requirements if awarded the Contract</p> <p>1. <i>Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed.</i></p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p style="text-align: right;">4.</p> <p>Warranty</p> <p>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>b) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</p> <p>c) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</p> <p>d) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the</p>			

equipment for three (3) years after the warranty period.

3. **Notarized undertaking** that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

4. With "**DOH-MMCHD HFEP**" (Government Property not for sale) sticker in each unit.

x-x

Source of Fund: SAA 2023-02-000687 (HFEP 2023 ConAp)

Recipient:

Taguig City – 11 units

Makati – 8 units