



**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB2024 – 083E**  
**PROCUREMENT OF 3 UNITS FETAL MONITOR**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
<b>Technical Specifications</b>	<b>Query</b>	<b>Response of the End User Unit</b>
Ultrasound Frequency: 1 MHz,  At least 150 - 160 patient data saving  <del>power aid module: 10-12hours</del>  1 units ultrasound gel	Ultrasound Frequency: 1 MHz, $\pm$ .1 MHz  At least 150 - 160 patient data saving or better  <del>power aid module: 10-12hours</del>  2 gallons ultrasound gel	Ultrasound Frequency: 1 MHz, $\pm$ .1 MHz-GRANTED  At least 150 - 160 patient data saving or better-GRANTED  power aid module: removed -GRANTED  2 gallons ultrasound gel - GRANTED

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 24th day of September 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 16<sup>th</sup> day of September 2024 in MMCHD

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD**  
Licensing Officer V / BAC Chairperson

Republic of the Philippines  
Department of Health  
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	<b>FETAL MONITOR</b>	Qty./Unit	<b>3 UNITS</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
<b>ABC: 480,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> <li>• Dual pulsed Doppler FHR</li> <li>• <i>Ultrasound Frequency: 1 MHz, ±.1 MHz</i></li> <li>• Fetal Heart Rate Measurement Range: 30-240 bpm</li> <li>• With auto-detection of dual fetal movement</li> <li>• With array-type thermal printer</li> <li>• With automatic printing function</li> <li>• At least 7-inch TFT color LCD Display</li> <li>• 3-channel display: FHR I, FHR II, UC</li> <li>• With Clinical event marking</li> <li>• <i>At least 150 - 160 patient data saving or better</i></li> <li>• Dimensions: Manufacturer's Standard</li> <li>• Can be connected to a central monitoring system through PC interface</li> <li>• Has built-in battery: 4-8 hours</li> <li>• Accessories: <ul style="list-style-type: none"> <li>- 2 units Ultrasound Doppler probe</li> <li>- 1 unit event marker jack</li> <li>- 3 units probe belt</li> <li>- 1 unit UC probe</li> <li>- <i>2 gallons ultrasound gel</i></li> <li>- 2 units probe holder</li> </ul> </li> </ul>		•	
<p><b>Requirements</b> if awarded the Contract</p> <p>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with <b>30</b> calendar days upon receipt of Notice to Proceed.</p> <p>2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p> <p>4. <b>Warranty</b></p> <p>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by</p>			

