



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB#2024 – 084E
PROCUREMENT OF 9 UNITS DENTAL UNIT, MOTORIZED WITH
ACCESSORIES

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed.	Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed.	Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed - GRANTED

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 24th day of September 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 16th day of September 2024 in MMCHD

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	DENTAL UNIT, MOTORIZED WITH ACCESSORIES	Qty./Unit	9 UNITS
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: P1,620,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> · Fully motorized with pre-set functions · Touch control chair · Slim ergonomic design upholstery · LED light with sensor switch · Double side armrest · Colored glass/ ceramic bowl, rotating detachable cuspidor bowl (with plastic strainer) · Two (2) pcs. 3 way syringe · Three (3) pcs hand piece attachments (2 hi-speed adaptors & 1 low-speed adaptor) · Auto-cup-filler & bowl flush (water flushing control valve, cup-filler with control valve) · Saliva ejector-built-in motor pump type, does not use air or water · Clean water system · LED X-ray film viewer (Peri-apical X-ray viewer) · Multi-functional foot control · Water tank for distilled water for hand piece · Two (2) hi-speed hand piece · One (1) low-speed complete with straight hand piece & Contra-angle · LED light cure unit (external) · Tissue box with tray and cup dispenser · Silent type and oil-free dental air compressor · Saddle-type with backrest dentist stool 			
<p>Requirements if awarded the Contract</p> <p>1. <i>Completion Period:</i> <i>The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60 calendar days upon receipt of Notice to Proceed.</i></p> <p>2. <i>Testing:</i> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <i>Training:</i> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</p>			1.

