



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2024 – 094E

PROCUREMENT OF 2 UNITS INFANT INCUBATOR (REBID)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

1. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Has manual and automatic humidity adjustment	Has manual and /or automatic humidity adjustment	Has manual and/or automatic humidity adjustment NOT GRANTED
Storage Volume: not less than 80 L	Storage volume (cabinet (30 L): not less than 80 L	Storage volume (cabinet (30 L): not less than 80 L GRANTED
<ul style="list-style-type: none"> • Air temperature mode set point range 20.0°C (68.0°F) to 39.0°C (102.2°F) • Air temperature mode set point override temperature range 37.0°C (98.6°F) to 39.0°C (102.2°F) • Skin temperature mode set point range 34.0°C (93.2°F) to 37.0°C (98.6°F) • Skin temperature mode set point override temperature range 37.0°C (98.6°F) to 38°C (100.4°F) • Temperature rise time at 22°C (72°F) ambient <35 min 	<ul style="list-style-type: none"> • Air temperature mode set point range 20.0°C (68.0°F) to 39.0°C (102.2°F) • Air temperature mode set point override temperature range 37.0°C (98.6°F) to 39.0°C (102.2°F) • Skin temperature mode set point range 34.0°C (93.2°F) to 37.0°C (98.6°F) • Skin temperature mode set point override temperature range 37.0°C (98.6°F) to 38°C (100.4°F) • Temperature rise time at 22°C (72°F) ambient <35 min 	<ul style="list-style-type: none"> • Air temperature mode set point range 20.0°C (68.0°F) to 39.0°C (102.2°F) • Air temperature mode set point override temperature range 37.0°C (98.6°F) to 39.0°C (102.2°F) • Skin temperature mode set point range 34.0°C (93.2°F) to 37.0°C (98.6°F) • Skin temperature mode set point override temperature range 37.0°C (98.6°F) to 38°C (100.4°F) • Temperature rise time at 22°C (72°F) ambient <35 min NOT GRANTED
Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed	Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60-90 calendar days upon receipt of	Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60-90 calendar days upon receipt of Notice to

	Notice to Proceed.	Proceed. GRANTED
--	--------------------	------------------

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 20th day of November 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 12th day of November 2024 in MMCHD

Approved by:

JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

Republic of the Philippines
 Department of Health
 Metro Manila Center for Health Development
 TECHNICAL SPECIFICATIONS

Item No. 1	INFANT INCUBATOR	Qty./Unit	2 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 4,200,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> • Intensive care incubator for treatment of premature and newborn babies • Can accommodate premature and full-term babies up to a maximum of 10kg (22 lbs.) • Has manual and automatic humidity adjustment • Double-wall with automatic air curtain to support thermoregulation • <i>Storage volume (cabinet (30 L): not less than 80 L</i> • Has ventral skin-to-skin contact (SSC) mode: <ul style="list-style-type: none"> - maintains the ideal microclimate inside the incubator while the neone is undergoing skin-to-skin care - with display in elapsed time and central peripheral temperature for continuous monitoring of staff • Has its condensate from the incubator compartment isolated from the clean water supply of the humidity system <ul style="list-style-type: none"> - The entire humidity system can be easily removed for quick, convenient yet effective hygienic reprocessing after every patient • Has dual-skin temperature monitoring • With variable height adjustment • With keypad lock function • Has 7-day trend for weight gain and loss • With integrated xray • 24-hour trend • Air temperature and skin temperature 			

<ul style="list-style-type: none"> • Relative humidity, Oxygen Concentration, Heater Power • Has central temperature probe • Peripheral Temperature Probe • Temperature control modes Skin and Air Temperature Control Mode • Air temperature mode set point range 20.0°C (68.0°F) to 39.0°C (102.2°F) • Air temperature mode set point override temperature range 37.0°C (98.6°F) to 39.0°C (102.2°F) • Skin temperature mode set point range 34.0°C (93.2°F) to 37.0°C (98.6°F) • Skin temperature mode set point override temperature range 37.0°C (98.6°F) to 38°C (100.4°F) • Temperature rise time at 22°C (72°F) ambient <35 min • Operating noise level in hood <47 dBA (with servo Oxygen Control) • Humidity control ranges from 30 to 95% in 1% increments • Humidity control capacity 1,500 ml • Accessories: <ul style="list-style-type: none"> - Storage - Utility shelf - IV pole • Device Classification <ul style="list-style-type: none"> - Protection class: Class I, Type BF, Continuous Operation - Ingress of liquids IPX0 	
<p>Requirements if awarded the Contract</p> <ol style="list-style-type: none"> 1. Completion Period: <i>The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 60-90 calendar days upon receipt of Notice to Proceed..</i> 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment. 4. Warranty <ol style="list-style-type: none"> a) Warranty certificate for two (2) years on parts and service. The supplier 	<p>1.</p>

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 7	Essential Oils	Qty./Unit	3,000 pcs
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: P 360,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
Specifications: Assorted scents, 10ml bottle with dropper Alcohol free, no harsh chemicals, safely ingredient, cruelty free, long lasting fragrance			
B. Additional requirement by the Lowest/Single Calculated Bid (L/SCB) as part of post qualification: 1. One (1) original sample of manufacturer's product to			

be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon contract award. The prototype of the labeling instruction must be part of the sample submitted; however, the technical specifications of the labeling instruction of the product must be complied with upon delivery.

2. L/SCRB shall pick up the Contract and Notice to Proceed issued in its favor within three (3) calendar days from receipt of notice. An electronic mail shall constitute an official notice to the Bidder.

3. Refusal to sign and accept the Award or enter into a contract without justifiable reason may be grounds for imposing administrative sanctions under Rule XXIII of the Revised IRR of RA 9184.

4. The registered company name and email address must be consistent and should reflect on all documents to be submitted.

5. Request for extension should be submitted before the lapse of the original delivery date. The maximum allowable extension shall not be longer than the Original Delivery term.

6. Delivery through courier service is not allowed.