



SUPPLEMENTAL/ BID BULLETIN NO. 1

IB#2024-118

Procurement of Various Assistive Devices

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

I. Query during Pre-bidding Conference:		
Technical Specifications	Query	Response of the End User Unit
Item No. 1 Crutches		
Minimum Height: 114 cm	Minimum Height: 114 cm or Manufacturer Standard	Minimum Height: 114 cm or Manufacturer Standard - GRANTED
Maximum Height: 135 cm	Maximum Height: 135 cm or Manufacturer Standard	Maximum Height: 135 cm or Manufacturer Standard- GRANTED.
Item No. 2: Medical Walker		
Overall width: 21"	Overall width: 21" or Manufacturer Standard	Overall width: 21" or Manufacturer Standard - GRANTED
Height Adjustment: Minimum 32"; Maximum 39"	Height Adjustment: Minimum 32"; Maximum 39" or Manufacturer Standard	Height Adjustment: Minimum 32"; Maximum 39" or Manufacturer Standard - GRANTED
Item No. 3: Quad Cane		
Base Material: Stainless Steel	Base Material: Stainless Powder Coated Steel	Base Material: Powder Coated Steel - GRANTED
Adjustable Height handle; Minimum 32"; Maximum 39"	Adjustable Height handle; Minimum 32"; Maximum 39" or Manufacturer Standard	Adjustable Height handle; Minimum 32"; Maximum 39" or Manufacturer Standard- GRANTED
Item no. 4: Single Cane		
Base Material: Stainless Steel	Base Material: Stainless Steel	Base material: Aluminum - GRANTED

Additional Requirements to be attached to the Technical Specifications form arranged, numbered, and tabbed as enumerated below:

1. Valid and current Certificate Product Registration (CPR) or Certificate of Medical Device Registration (CMDR) or Certificate of Medical Device Notification (CMDN) issued by the Philippine Food and Drug Administration (PFDA); The CPR must be valid for the entire period of the award. If the CPR/CMDN/CMDR is about to expire, the supplier must have submitted a copy of an application

of renewal to the FDA at least three (3) months before the expiry date (a copy of the expiring CPR, which is stamped with an "Extension of Validity" shall be submitted as proof); [AO 2019-0041]

2. Valid and current License to Operate (LTO) for Medical Device Exporter / Trader /Importer / Distributor / Wholesaler issued by Philippine Food and Drugs Administration (PFDA)

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 7th day of August 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 31st day of July 2024 in MMCHD

Approved by:



JEREMIAS FRANCIS Y. CHAN, MD
Licensing Officer V / BAC Chairperson

TECHNICAL SPECIFICATIONS			
Item No. 1	Crutches	Qty./Unit	580 pcs
Name of Manufacturer:		Country of Origin (If applicable)	
Brand:		Model: (if applicable)	
ABC: P 464,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Specifications:</p> <p>Aluminum Material, Heavy Duty, Lightweight</p> <p>Portable, adjustable</p> <p>Comfortable Rubber Cushioning and Hand Grip</p> <p><i>Minimum Height: 114 cm or Manufacturer Standard</i></p> <p><i>Maximum Height: 135 cm or Manufacturer Standard</i></p> <p>Period: 30-60 calendar days after receipt of approved P.O./NTP</p> <p>Place: Pasig Warehouse</p>			
<p><i>B. Additional Requirements to be attached to the Technical Specifications form arranged, numbered, and tabbed as enumerated below:</i></p> <p><i>1. Valid and current Certificate Product Registration (CPR) or Certificate of Medical Device Registration (CMDR) or Certificate of Medical Device Notification (CMDN) issued by the Philippine Food and Drug Administration (PFDA):</i></p> <p><i>The CPR must be valid for the entire period of the award. If the CPR/CMDN/CMDR is about to expire, the supplier must have submitted a copy of an application of renewal to the FDA at least 3 months before the expiry date (a copy of the expiring CPR, which is stamped with an "Extension of Validity" shall be submitted as proof); [AO 2019-0041]</i></p> <p><i>2. Valid and current License to Operate (LTO) for Medical Device Exporter / Trader / Importer / Distributor / Wholesaler issued by Philippine Food and Drugs Administration (PFDA)</i></p>			

TECHNICAL SPECIFICATIONS			
Item No. 2	Adult Medical Walker	Qty./Unit	435 pcs
Name of Manufacturer:		Country of Origin (If applicable)	
Brand:		Model: (if applicable)	
ABC: P 565,500.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Specifications:</p> <p>Load Capacity : 60kg minimum Sturdy anodized aluminum frame that insures maximum safety and strength while remaining lightweight</p> <p>Padded Hand Grip: with soft and comfortable handle</p> <p>Shoes/Tips: Broad based PVC shoes</p> <p>Fordable Frame</p> <p><i>Overall width: 21" or Manufacturer Standard</i></p> <p>Two button push release is easy to extend/fold.</p> <p>Each side folds independently to allow easy movement through narrow spaces slip-resistant rubber tips.</p> <p><i>Height Adjustment: Minimum 32"; Maximum 39" or Manufacturer Standard</i></p> <p>Period: 30-60 calendar days after receipt of approved P.O./NTP</p> <p>Place: Pasig Warehouse</p>			
<p><i>B. Additional Requirements to be attached to the Technical Specifications form arranged, numbered, and tabbed as enumerated below:</i></p> <p><i>1. Valid and current Certificate Product Registration (CPR) or Certificate of Medical Device Registration (CMDR) or Certificate of Medical Device Notification (CMDN) issued by the Philippine Food and Drug Administration (PFDA):</i></p> <p><i>The CPR must be valid for the entire period of the award. If the CPR/CMDN/CMDR is about to expire, the supplier must have submitted a copy of an application of renewal to the FDA at least 3 months before the expiry date (a copy of the expiring CPR, which is stamped with an "Extension of Validity" shall be submitted as proof); [AO 2019-0041]</i></p> <p><i>2. Valid and current License to Operate (LTO) for Medical Device Exporter / Trader /Importer / Distributor</i></p>			

/ Wholesaler issued by Philippine Food and Drugs Administration (PFDA)	
--	--

TECHNICAL SPECIFICATIONS			
Item No. 3	Quad Cane	Qty./Unit	725 pcs
Name of Manufacturer:		Country of Origin (if applicable)	
Brand:		Model: (if applicable)	
ABC: P 580,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Specifications:</p> <p>Load Capacity: 60kg minimum</p> <p><i>Base Material: Powder Coated Steel</i></p> <p>4 point base, combined with offset handle that provides additional support and stability</p> <p>Aluminum tubing is durable and strong</p> <p>Soft and comfortable padded handgrip with standard wrist strap</p> <p>Secured with a Locking ring for added safety</p> <p>Easy Push Button</p> <p><i>Adjustable Height handle; Minimum 32"; Maximum 39" or Manufacturer Standard</i></p> <p>Period: 30-60 calendar days after receipt of approved P.O./NTP</p> <p>Place: Pasig Warehouse</p>			
<p><i>B. Additional Requirements to be attached to the Technical Specifications form arranged, numbered, and tabbed as enumerated below:</i></p> <p><i>1. Valid and current Certificate Product Registration (CPR) or Certificate of Medical Device Registration (CMDR) or Certificate of Medical Device Notification (CMDN) issued by the Philippine Food and Drug Administration (PFDA);</i></p> <p><i>The CPR must be valid for the entire period of the award. If the CPR/CMDN/CMDR is about to expire, the supplier must have submitted a copy of an application of renewal to the FDA at least 3 months before the expiry date (a copy of the expiring CPR, which is stamped with an "Extension of Validity" shall be submitted as proof); [AO 2019-0041]</i></p> <p><i>2. Valid and current License to Operate (LTO) for</i></p>			