

Republic of the Philippines Department of Health METRO MANILA CENTER FOR HEALTH DEVELOPMENT



## SUPPLEMENTAL/ BID BULLETIN NO. 1

# IB2025 – 018E

# PROCUREMENT OF 1 UNIT CARDIOTOCOGRAPHY MACHINE (SHORT OF AWARD)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Query during Pre-bidding Conference:		
Technical Specification	Query	<b>Response of the End User</b>
		Unit
No changes stipulated in the technical specifications		
The Bidder's SLCC as indicated in <b>ITB</b> Clause 5.3 should have been completed within five		
(5) years prior to the deadline for the submission and receipt of bids.		
The Bidder must have completed a two or more contract that is similar to this Project		
Use the MMCHD Template for Omnibus Sworn Statement Revised with Secretary		
Certificate, which is included below.		

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 17<sup>th</sup> day of December 2024, 9:00 AM:** 

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 10<sup>th</sup> day of December 2024 in MMCHD

Approved by: JEREMIAS FRANCIS Y. CHAN, MD Licensing Officer V / BAC Chairperson

Republic of the Philippines Department of Health Metro Manila Center for Health Development			
Item No. 1	TECHNICAL SPECIFICATI Cardiotocography Machine	Qty./Unit	1Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 700,000.00	PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE
<ul> <li>Power: 220V, 5</li> <li>Rechargeable L time</li> <li>Ultrasound Tra <ul> <li>12-Crystal T</li> <li>With color ii</li> </ul> </li> <li>Toco Transduce</li> <li>Recording: <ul> <li>Recorder: T</li> <li>Recorder Pa</li> <li>Paper Widt</li> </ul> </li> </ul>	CIFICATIONS: olor LCD Touchscreen 0/60Hz i-ion battery: 5000 mAh, with 2 hour continu nsducer: Fransducer with 2.5m cable identification er cable length: 2.5m Thermal Dot Matrix aper: Z-fold, Thermosensitive h: 152mm/ 150mm FData: ±5% (X-axis) and ±1% (Y-axis)	ous working	
- Measurem Pressure - Method: Os	ent: Systolic Pressure, Diastolic Pressure, I cillometric Method 1 mmHg (0.1 kPa)	Mean Artery	
• FHR:			
<ul> <li>Operating Mode: Ultrasound Pulse Doppler with autocorrelation</li> <li>FHR Measuring Range: 50 bpm ~ 240bpm</li> <li>Resolution: 1bpm</li> <li>Accuracy: ±2 bpm</li> <li>Alarm: FHR Alarm</li> <li>Temperature Rise: less than 2° above ambient temperature</li> <li>Dielectric Strength: 4000 Vrms</li> </ul>			
<ul> <li>Dielectric Si</li> <li>DECG <ul> <li>DFHR Meas</li> <li>Resolution:</li> <li>Accuracy: ±</li> <li>Alarm: DFH</li> </ul> </li> <li>IUP <ul> <li>Pressure Rational Pressure Pres</li></ul></li></ul>	Error: ±10% Automatic/ Manual trength: 4000 Vrms uring Range: 30 bpm ~ 240bpm 1bpm 1 bpm		

<ul> <li>HIR Measuring Range: 30 bpm ~ 240bpm</li> <li>Resolution: 1bpm MHR</li> <li>MIR Measuring Accuracy: ±2 bpm</li> <li>\$PO2</li> <li>Measurement Range: 50%~100%</li> <li>Resolution: 1%</li> <li>Accuracy: ±3 bpm</li> <li>TEMP</li> <li>Measurement Range: 0~50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: ±0.3°C</li> <li>Data Transmission</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Transmission</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Tedument of the delivery, installation, testing and</li> <li>commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Training: The supplier shall provide a training on the proper use and maintenance, the end users shall conduct a physical amaintenance to the equipment to the sectures and user that the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty:</li> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective material are ownorkmashing under normal use. The warranty period shall commence from the date of acceptance by the end-user and to the days upon the date of acceptance by the end-user and to the days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment threakdown/defects.</li> <li>d) The number of days where the equipment trains on the proper use and mainferance within five (5) calendar days upon notification for the equipment to end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment part in the equipment that is found to be defective or workmashing hall be added to the days upon notification of the equipment of days where the equipment to the end-user the equipment for the equipment</li></ul>		
<ul> <li>Max Measuring time: 120s</li> <li>Normal Measuring time: 30–45s</li> <li>MFM &amp; AFM</li> <li>Display Range: 0 ~ 999</li> <li>FM Mode: Automatic/ Manual</li> <li>AMF Technique: Pulsed Doppler Ultrasound</li> <li>MBCG</li> <li>MIR Measuring Range: 30 bpm ~ 240bpm</li> <li>Resolution: fbpm MR</li> <li>MIR Measuring Accuracy: ±2 bpm</li> <li>SPO2</li> <li>Measurement Range: 50%~100%</li> <li>Resolution: 1%</li> <li>Accuracy: ±3 bpm</li> <li>TEMP</li> <li>Measurement Range: 0~50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: 4.3 Curacy</li> <li>Temp</li> <li>Magement System: MFM-CS</li> <li>Base of the Contract</li> <li>HIS connection: HL7(GDT</li> <li>Requirements if awarded the Contract</li> <li>Completion Period: The delivery, installation, testing and</li> <li>commissioning of the equipment and its accessories, including the training of end-users and maintenance staff mush be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Z. Testing, Prior to acceptance, the end users and londuct a physical inspection and functionality test. The equipment mush be functioning and maintenance to the equipment to the hospital maintenance to the equipment.</li> <li>A. Warranty:</li> <li>A. Warranty certificate for two (2) years on parts and service. The supplier shall commission update the equipment.</li> <li>A. Warranty:</li> <li>A. Warranty certificate for two (2) years on parts and service. The supplier shall commission or parts and service. The supplier shall commence from the date of acceptance by the end-user atter testing and commission.</li> <li>D) Dreventive maintenance at least every six (6) months or according to the maintenance of the equipment to the equipment.</li> <li>A. Warranty:</li> <li>A. Warranty certificate for two (2) years on parts and service. The supplier shall commence from the date of acceptance by the end-user after testing and commission.</li> <li>D) Dreventive maintenance at least every six (6) months or according to the maintenance of the equipment to the equipment</li></ul>		
<ul> <li>Normal Measuring time: 30-45s</li> <li>MRM &amp; ARM</li> <li>Display Range: 0 - 999</li> <li>PM Mode: Automatic/ Manual</li> <li>AWT Exchinque: Pulsed Doppler Ultrasound</li> <li>AWT Exchinque: Pulsed Doppler Ultrasound</li> <li>MECG</li> <li>MRIK Measuring Range: 30 bpm ~ 240bpm</li> <li>Resolution: 1bpm MRR</li> <li>MIR Measuring Accuracy: ±2 bpm</li> <li>SPO2</li> <li>Measurement Range: 50%~100%</li> <li>Resolution: 0.1°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: ±3 bpm</li> <li>TEMP</li> <li>Measurement Range: 0-50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: ±0.3°C</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Transmission</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>HIS connection: HL//GDT</li> <li>Requirements if awarded the Contract</li> <li>Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Z. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical darge and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty:</li> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall commence from the date of acceptance by the end-user and commissioning.</li> <li>b) Preventive maintenance within five (5) calendar days upon nortification from the (2) years on parts and service. The supplier shall commend etimes of maintenance by the end-user after testing and commissioning.</li> <li>c) Orrective maintenance within five (5) calendar days upon nortification for the cquipment to three (3) years after the warranty period.</li> <li>d) The number of days where the equipment is unusable d</li></ul>		
<ul> <li>MFM &amp; AFM</li> <li>Display Range: 0 ~ 999</li> <li>FM Mode: Automatic/ Manual</li> <li>AWF Technique: Pulsed Doppler Ultrasound</li> <li>MECG</li> <li>MIR Measuring Range: 30 bpm ~ 240bpm</li> <li>Resolution: Tbpm MR</li> <li>MIR Measuring Accuracy: ± 2 bpm</li> <li>SFO2</li> <li>Measurement Range: 50%~100%</li> <li>Resolution: 1%</li> <li>Accuracy: ±3 bpm</li> <li>TEMF</li> <li>METEM</li> <li>METEMF</li> <li>Accuracy: ±3 bpm</li> <li>TEMF</li> <li>Data Management du USB</li> <li>Report Ethermet and Ethermet and USB Calendar days upon receipt of Notice to Proceed.</li> <li>Testing: Prior to acceptance, the end user shall conduct a physical inspection and maintenance staff must be delivery of the equipment.</li> <li>Marranty: ertificate for two (2) years on parts and service. The supplier shall provide a training on the proper use and maintenance staff within 3 days upon the delivery of the equipment.</li> <li>Warranty: ertificate for two (2) years on parts and service. The supplier shall commence fr</li></ul>		
<ul> <li>Display Range: 0 ~ 999</li> <li>FM Mode: Automatic / Manual</li> <li>AMF Technique: Polsed Doppler Ultrasound</li> <li>MECG</li> <li>MHR Measuring Range: 30 bpm ~ 240bpm</li> <li>Resolution: 1bpm MIR</li> <li>MHR Measuring Accuracy: ±2 bpm</li> <li>SFO2</li> <li>Measurement Range: 50%~100%</li> <li>Resolution: 1%</li> <li>Accuracy: ±3 bpm</li> <li>TEMP</li> <li>Measurement Range: 0-50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: ±0.3°C</li> <li>Data Transmission</li> <li>Data Transmission</li> <li>Data Transmission</li> <li>Data Transmission and USB</li> <li>Report Format: TRC</li> <li>Data Transmission of the equipment and its accessories, including the training of end-users and maintenance estaff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Z. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Z. Testing: Prior to acceptance, the end users and service. The supplier shall provide a training on the proper use and maintenance staff with 3 days upon the delivery of the equipment must be functioning and must have no physical diamage and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and maintenance staff within 3 days upon the delivery of the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commence from the date of acceptance by the end-user after testing and commence from the date of acceptance by the end-user after testing and commence from the date of acceptance by the end-user after testing and commence from the date of acceptance by the end-user after testing and commence from the date of acceptance by the end-user after testing and commence from the date of acceptance by the end-user after testing an</li></ul>	e e	
<ul> <li>- FM Mode: Automatic/ Manual</li> <li>- AMF Technique: Pulsed Doppler Ultrasound</li> <li>• MECG</li> <li>- MHR Measuring Range: 30 bpm ~ 240bpm</li> <li>- Resolution: 1bpm MHR</li> <li>- MIR Measuring Accuracy: ±2 bpm</li> <li>• SPO2</li> <li>- Measurement Range: 50%~100%</li> <li>- Resolution: 1%</li> <li>- Accuracy: ±3 bpm</li> <li>• TEMP</li> <li>- Measurement Range: 0~50°C</li> <li>- Resolution: 0.1°C</li> <li>- Accuracy: e1.3°C</li> <li>- Data Export: Ethernet and USB</li> <li>- Report Format: TRC</li> <li>- Data Management System: MFM-CS</li> <li>- Illi Connection: 11// /CDT</li> <li>Requirements if awarded the Contract</li> <li>- Completion Period: The delivery, installation, testing and</li> <li>commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and maintenance staff must be end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty:</li> <li>a) Warranty: certificate for two (2) years on parts and service. The supplier shall commerce from the date of acceptance by the end-user rafter testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the maintenance at least every six (s) fornomission raceording to the maintenance within five (5) calendar days upon entification from the end-user: and to the cospital equipment to the equipment of a careptance by the end-user rafter testing and commissioning.</li> <li>b) Preventive maintenance within five (5) calendar days upon notification for the equipment for the (5) calendar day</li></ul>		
<ul> <li>AMF Technique: Pulsed Doppler Ultrasound</li> <li>MECG <ul> <li>MIR Measuring Acareay: 30 bpm ~ 240bpm</li> <li>Resolution: 1bpm MHR</li> <li>MIR Measuring Accuracy: 22 bpm</li> </ul> </li> <li>SPO2 <ul> <li>Measurement Range: 50%~100%</li> <li>Resolution: 1%</li> <li>Accuracy: 33 bpm</li> </ul> </li> <li>TEMP <ul> <li>Measurement Range: 0~50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: 10.3°C</li> </ul> </li> <li>Data Transmission <ul> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Transmission</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Transmission</li> <li>Of the Guivery, installation, testing and</li> <li>commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Zresting: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>Training: The supplier shall provide a training on the proper use and maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty:</li> <li>Warranty:</li> <li>Warranty:</li> <li>Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>Orrective maintenance at least every sis (6) months or according to the mainfeance of the equipment the added to the warranty period.</li> <li>Orrective maintenance at least every sis (6) months or according to the mainfeance of the equipment transkip work complexity comprehensive preventive maintenance thin five (5) calendar days upon notification for the equipment breakdown/defects.</li> </ul> </li> <li>d) The number</li></ul>	- Display Range: 0 ~ 999	
<ul> <li>MECG <ul> <li>MHR Measuring Range: 30 bpm ~ 240bpm</li> <li>Resolution: 1bpm MHR</li> <li>MHR Measuring Accuracy: ±2 bpm</li> </ul> </li> <li>SPO2 <ul> <li>Measurement Range: 50%~100%</li> <li>Resolution: 1%</li> <li>Accuracy: ±3 bpm</li> </ul> </li> <li>TEMP <ul> <li>Measurement Range: 0~50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: t.0.3°C</li> </ul> </li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>HIS connection: HI/GDT</li> </ul> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>HIS connection: HI/GDT</li> <li>Requirements if awarded the Contract</li> <li>Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Z. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment fust that is found to be defective material or workmanship under normal use. The warranty period shall commerc from the date of acceptance by the end-user after testing and commission.</li> <li>D Preventive maintenance within five (5) calendar days upon normal use. The warranty period shall compress of most of according to the manafacturer's recommendations;</li> <li>O Corrective maintenance within five (5) calendar days upon normal use. The warranty period shall be added to the warranty period.</li> <li>Notarized undertaking that the supplier shall be added to the equipment is unusable due to decytome atherance within fiv</li>		
<ul> <li>HIR Measuring Range: 30 bpm ~ 240bpm</li> <li>Resolution: 1bpm MHR</li> <li>MIR Measuring Accuracy: ±2 bpm</li> <li>\$POZ</li> <li>Measurement Range: 50%~100%</li> <li>Resolution: 1%</li> <li>Accuracy: ±3 bpm</li> <li>TEMP</li> <li>Measurement Range: 0~50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: ±3 bpm</li> <li>Data Transmisson</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Transmisson</li> <li>Data Rapper Statement And PCC</li> <li>This connection: HL7/GDT</li> <li>Requirements if awarded the Contract</li> <li>Completion Period: The delivery, installation, testing and commissioning of the equipment and tas accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Z. Testing: Proi to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment to the social amaintenance taff must be functioning and must have no physical damage and defect.</li> <li>3.Training: The supplier shall provide a training on the proper use and maintenance to the equipment to the cand-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty:</li> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall cither repair or replace any item or part in the equipment that is found to be defective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment trainsioning.</li> <li>b) Preventive maintenance within five (5) calendar days upon notification of the equipment to end-user regarding support period.</li> <li>c) Orrective maintenance at least every six (6) months or according to the mainfaction from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the wa</li></ul>	- AMF Technique: Pulsed Doppler Ultrasound	
<ul> <li>Resolution: 1bpm MHR</li> <li>MHR Measuring Accuracy: ±2 bpm</li> <li>SPO2</li> <li>Measurement Range: 50%~100%</li> <li>Resolution: 1%</li> <li>Accuracy: ±3 bpm</li> <li>TEMP</li> <li>Measurement Range: 0~50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: ±0.3°C</li> <li>Data Transmission</li> <li>Data Kport: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>HIS connection: HI// GDT</li> <li>Requirements if awarded the Contract</li> <li>Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Z. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance at effective in material or workmaship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance within five (5) calendar days upon normal use. The warranty period shall commence from the date of acceptance by the end-user shalt to the equipment that is found to be defective in material or workmaship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance within five (5) calendar days upon notification of the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification of the maun</li></ul>	• MECG	
<ul> <li>MHR Measuring Accuracy: ±2 bpm</li> <li>SP02</li> <li>Measurement Range: 50%~100%</li> <li>Resolution: 1%</li> <li>Accuracy: ±3 bpm</li> <li>TEMP</li> <li>Measurement Range: 0~50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: ±0.3°C</li> <li>Data Transmisson</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>HIS connection: HL7/GDT</li> <li>Requirements if awarded the Contract</li> <li>Completion Period: The delivery, installation, testing and commissioning of the equipment table conjunct with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Testing: Prior to acceptance, the equipment must be functioning and must have no physical damage and defect.</li> <li>Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance at fir with 0 caleivary of the equipment.</li> <li>Warranty certificate for two (2) years on parts and service. The supplier shall commendations;</li> <li>Orrective maintenance at ele stervy six (6) months or according to the end-user and service. The supplier shall commence from the date of acceptance by the end-user ating on the acceptance by the end-user ating and envirtual to the end-user ating the normal use. The warranty period shall commence from the date of acceptance by the end-user ating and commissioning.</li> <li>D) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>Corrective maintenance within five (5) calendar days upon notification of the equipment to the explore shall be added to the warranty period.</li> <li>The supplier shall specify post-warranty comprehensive preventive maintenance within five (5) calendar days upon notification of the equipment to the days upon notification of the equipment to the equipment to</li></ul>	- MHR Measuring Range: 30 bpm ~ 240bpm	
<ul> <li>SP02 <ul> <li>Measurement Range: 50%~100%</li> <li>Resolution: 1%</li> <li>Accuracy: ±3 Bpm</li> </ul> </li> <li>TEMP <ul> <li>Measurement Range: 0~50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: ±0.3°C</li> </ul> </li> <li>Data Transmisison <ul> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Transmision of the capityment and the scensories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Testing: Prior to acceptance, the end user shall conduct a physical aniantenance staff within 3 days upon the delivery of the equipment.</li> <li>Warranty: <ul> <li>Warranty:</li> <li>Warranty:</li> <li>Warranty:</li> <li>Warranty:</li> <li>Orrective maintenance within five (5) calendar days upon receipt of Notice to replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>Preventive maintenance within five (5) calendar days upon notification for the equipment breakdown/defects.</li> <li>() Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>() The number of days where the equipment is unusable due to equipment defect/faults shall be added to the warranty period.</li> <li>() The supplier shall specify post-warranty comprehensive preventive maintenance within five (5) calendar days upon notification of the equipment for three (3) years after the warranty period.</li> <li>() The number of days where the equipment to end-user afts of the equipment to the end-user after testing and prices of major s</li></ul></li></ul></li></ul>	- Resolution: 1bpm MHR	
<ul> <li>Measurement Range: 50%~100%</li> <li>Resolution: 1%</li> <li>Accuracy: ±3 bpm</li> <li>TEMP</li> <li>Measurement Range: 0~50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: ±0.3°C</li> <li>Data Transmission</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>Hill connection: HL7/GDT</li> <li>Requirements if awarded the Contract</li> <li>Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Testing: Proto to acceptance, the end user shall conduct a physical inspection and functionality test. The equipments and service. The supplier shall provide a training on the proper use and maintenance staff within 3 days upon the delivery of the equipment.</li> <li>Warranty certificate for two (2) years on parts and service. The supplier shall provide a training on the hequipment.</li> <li>Warranty: <ul> <li>Warranty certificate for two (2) years on parts and service. The supplier shall commence from the date of acceptance by the end-user any tem or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>Corrective maintenance within five (5) calendar days upon notification of the equipment to frame deglecify losity shall be added to the warranty period.</li> <li>Marranty certification for workmanship subgroup notification of the equipment to the warranty period.</li> <li>The supplier shall specify post-warranty operiod shall be added to the warranty period.</li> <li>The supplier shall be added to the warranty period.</li> <li>Notarized undertaking that the supplier shall conduct the ne</li></ul></li></ul>	- MHR Measuring Accuracy: ±2 bpm	
<ul> <li>Resolution: 1%</li> <li>Accuracy: ±3 bpm</li> <li>TEMP</li> <li>Measurement Range: 0~50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: ±0.3°C</li> <li>Data Transmisison</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>HIS connection: HL7/GDT</li> <li>Requirements if awarded the Contract</li> <li>1. Completion Period: The delivery, installation, testing and</li> <li>commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3.Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-user and to the hospital maintenance of the equipment to the end-user and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty:</li> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall ether repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance within five (5) calendar days upon second to the equipment shall specify post-warranty comprehensive preventive maintenance to staff cubic to species of major spare parts of the equipment for the re(3) years after the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance torts including its and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the</li></ul>	• SP02	
<ul> <li>Resolution: 1%</li> <li>Accuracy: ±3 bpm</li> <li>TEMP</li> <li>Measurement Range: 0~50°C</li> <li>Resolution: 0.1°C</li> <li>Accuracy: ±0.3°C</li> <li>Data Transmisison</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>HIS connection: HL7/GDT</li> <li>Requirements if awarded the Contract</li> <li>1. Completion Period: The delivery, installation, testing and</li> <li>commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3.Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-user and to the hospital maintenance of the equipment to the end-user and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty:</li> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall ether repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance within five (5) calendar days upon second to the equipment shall specify post-warranty comprehensive preventive maintenance to staff cubic to species of major spare parts of the equipment for the re(3) years after the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance torts including its and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the</li></ul>	- Measurement Range: 50%~100%	
<ul> <li>• TEMP <ul> <li>• Measurement Range: 0~50°C</li> <li>• Resolutoin: 0.1°C</li> <li>• Accuracy: ±0.3°C</li> </ul> </li> <li>• Data Transmisison <ul> <li>• Data Export: Ethernet and USB</li> <li>• Report Format: TRC</li> <li>• Data Management System: MFM-CS</li> <li>• HIS connection: HL7/GDT</li> </ul> </li> <li>Requirements if awarded the Contract</li> <li>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3.Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty: <ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations; <ul> <li>c) Corrective maintenance at least every six (6) months or according to the manufacturer's necemmendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon softification from the end-user regarding equipment breakdown/defects.</li> </ul> </li> <li>d) The number of days where the equipment is unusable due to equipment of the equipment breakdown/defects.</li> <li>d) The number of day</li></ul></li></ul>		
<ul> <li>• TEMP <ul> <li>• Measurement Range: 0~50°C</li> <li>• Resolutoin: 0.1°C</li> <li>• Accuracy: ±0.3°C</li> </ul> </li> <li>• Data Transmisison <ul> <li>• Data Export: Ethernet and USB</li> <li>• Report Format: TRC</li> <li>• Data Management System: MFM-CS</li> <li>• HIS connection: HL7/GDT</li> </ul> </li> <li>Requirements if awarded the Contract</li> <li>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3.Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty: <ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations; <ul> <li>c) Corrective maintenance at least every six (6) months or according to the manufacturer's necemmendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon softification from the end-user regarding equipment breakdown/defects.</li> </ul> </li> <li>d) The number of days where the equipment is unusable due to equipment of the equipment breakdown/defects.</li> <li>d) The number of day</li></ul></li></ul>	- Accuracy: ±3 bpm	
<ul> <li>Resolutoin: 0.1°C</li> <li>Accuracy: :0.3°C</li> <li>Data Transmisison</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>HIS connection: HL7/GDT</li> <li>Requirements if awarded the Contract</li> <li>Completion Period: The delivery, installation, testing and</li> <li>commissioning of the equipment and its accessories, including the training</li> <li>of end-users and maintenance staff must be completed with 90 calendar</li> <li>days upon receipt of Notice to Proceed.</li> <li>Z. Testing: Prior to acceptance, the end user shall conduct a physical</li> <li>inspection and functionality test. The equipment must be functioning and</li> <li>must have no physical damage and defect.</li> <li>3.Training: The supplier shall provide a training on the proper use and</li> <li>maintenance of the equipment to the end-users and to the hospital</li> <li>maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty:</li> <li>a) Warranty certificate for two (2) years on parts and service. The</li> <li>supplier shall either repair or replace any item or part in the equipment</li> <li>that is found to be defective in material or workmanship under normal</li> <li>use. The warranty period shall commence from the date of acceptance by</li> <li>the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to</li> <li>equipment defects/faults shall be added to the warranty period.</li> <li>s. Notarized undertaking that the supplier shall conduct the necessary</li> <li>corrective maintenance within five (5) calendar days upon</li> <li>notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to</li> <li>equipment defects/faults shall be added to the warranty period.</li> <li>S. Notarized undertaking that the supplier shall conduct the necessary</li> <li>corrective maintenance within five (5) calendar days</li></ul>		
<ul> <li>Resolutoin: 0.1°C</li> <li>Accuracy: ±0.3°C</li> <li>Data Transmisson</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>HIS connection: HL7/GDT</li> <li>Requirements if awarded the Contract</li> <li>1. Completion Period: The delivery, installation, testing and</li> <li>commissioning of the equipment and its accessories, including the training</li> <li>of end-users and maintenance staff must be completed with 90 calendar</li> <li>days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical</li> <li>inspection and functionality test. The equipment must be functioning and</li> <li>must have no physical damage and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and</li> <li>maintenance of the equipment to the end-users and to the hospital</li> <li>maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty:</li> <li>a) Warranty certificate for two (2) years on parts and service. The</li> <li>supplier shall either repair or replace any item or part in the equipment</li> <li>that is found to be defective in material or workmanship under normal</li> <li>use. The warranty period shall commence from the date of acceptance by</li> <li>the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to</li> <li>equipment defects/faults shall be added to the warranty period.</li> <li>o. Notarized undertaking that the supplier shall conduct the necessary</li> <li>corrective maintenance within five (5) calendar days upon</li> <li>notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to</li> <li>equipment for three (3) years after the warranty period.</li> <li>Notarized undertaking that the supplier shall conduct the necessary</li> <li>corrective maintenance within five (5) calendar days upon n</li></ul>		
<ul> <li>Accuracy: ±0.3°C</li> <li>Data Transmisison</li> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>HIS connection: HL7/GDT</li> </ul> Requirements if awarded the Contract <ol> <li>Completion Period: The delivery, installation, testing and</li> <li>commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>Warranty: <ul> <li>Warranty certificate for two (2) years on parts and service. The supplier shall end row rokmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user regarding equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user regarding equipment breakdown/defects.</li> <li>The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>Notarized undertaking that the supplier shall protices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment the the number of days where the equipment breakdown/defects.</li> </ul></li></ol>		
<ul> <li>Data Transmission <ul> <li>Data Export: Ethernet and USB</li> <li>Report: Ethernet and USB</li> <li>Report: Format: TRC</li> <li>Data Management System: MFM-CS <ul> <li>HIS connection: HL7/CDT</li> </ul> </li> <li>Requirements if awarded the Contract</li> </ul> </li> <li>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3.Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>c) Notarized undertaking that the supplier shall periods of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>S. Notarized undertaking that the supplier shall conduct th necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the<!--</td--><td></td><td></td></li></ul>		
<ul> <li>Data Export: Ethernet and USB</li> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>HIS connection: HL7/GDT</li> </ul> Requirements if awarded the Contract <ol> <li>Completion Period: The delivery, installation, testing and</li> <li>commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar</li> <li>days upon receipt of Notice to Proceed.</li> <li>Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>Warranty: <ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall eriotive in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the maufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> </ul> </li> <li>d) The number of days where the equipment is unusable due to equipment of three (3) years after the warranty period.</li> <li>e) The supplier shall be added to the warranty period.</li> <li>e) Nearized functional is and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>S. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment that the number of days where the equipment tis unusable due to defective material or workmanship, shall be added to the</li> </ol>	5	
<ul> <li>Report Format: TRC</li> <li>Data Management System: MFM-CS</li> <li>HIS connection: HL7/GDT</li> <li>Requirements if awarded the Contract</li> <li>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty: <ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> </ul> </li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>c) Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall sociely enventive maintenance oxis including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier</li></ul>		
<ul> <li>Data Management System: MFM-CS</li> <li>HIS connection: HL7/GDT</li> <li>Requirements if awarded the Contract</li> <li>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty:</li> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance at least every six (6) months or according to the manufacturer's necommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the</li></ul>		
<ul> <li>- HIS connection: HL7/GDT</li> <li>Requirements if awarded the Contract</li> <li>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty:</li> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance within five (5) calendar days upon notification of the equipment the the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment the the unuber of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>	•	
Requirements if awarded the Contract         1. Completion Period: The delivery, installation, testing and         commissioning of the equipment and its accessories, including the training         of end-users and maintenance staff must be completed with 90 calendar         days upon receipt of Notice to Proceed.         2. Testing: Prior to acceptance, the end user shall conduct a physical         inspection and functionality test. The equipment must be functioning and         must have no physical damage and defect.         3.Training: The supplier shall provide a training on the proper use and         maintenance of the equipment to the end-users and to the hospital         maintenance staff within 3 days upon the delivery of the equipment.         4. Warranty:         a) Warranty certificate for two (2) years on parts and service. The         supplier shall either repair or replace any item or part in the equipment         that is found to be defective in material or workmaship under normal         use. The warranty period shall commence from the date of acceptance by         the end-user after testing and commissioning.         b) Preventive maintenance within five (5) calendar days upon         notification from the end-user regarding equipment breakdown/defects.         d) The number of days where the equipment is unusable due to         equipment defects/faults shall be added to the warranty period.         e) The supplier shall specify		
<ul> <li>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty: <ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> </ul> </li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>f. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment tis the unber of days where the equipment is unusable due to defective material or workmanship, shall be added to the upplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment</li></ul>		
<ul> <li>commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the end-users and to the proper use and maintenance of the equipment to the end-users and to the hospital maintenance of the equipment to the equipment.</li> <li>4. Warranty: <ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the maunfacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> </ul> </li> <li>d) The number of days where the equipment is unusable due to equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance with</li></ul>		
of end-users and maintenance staff must be completed with 90 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3.Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment. 4. Warranty: a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations; c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects. d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period. e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the		
<ul> <li>days upon receipt of Notice to Proceed.</li> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty: <ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> </ul> </li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance within five (5) calendar days upon notification of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to the equipment of the equipment for the eduipment for the eduipment from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to the equipment breakdown from the end-use</li></ul>		
<ul> <li>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty: <ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> </ul> </li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance within five (5) calendar days upon notification of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to the equipment is unusable due to defective material or workmanship, shall be added</li></ul>	-	
<ul> <li>inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</li> <li>3.Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty: <ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> </ul> </li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to equipment for three (3) years after the warranty period.</li> </ul>		
<ul> <li>must have no physical damage and defect.</li> <li>3.Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty: <ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> </ul> </li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>		
<ul> <li>3.Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.</li> <li>4. Warranty: <ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> </ul> </li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>		
maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment. 4. Warranty: a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations; c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects. d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period. e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the		
maintenance staff within 3 days upon the delivery of the equipment. 4. Warranty: a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations; c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects. d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period. e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the		
<ul> <li>4. Warranty:</li> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>		
<ul> <li>a) Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to the manufacture of the equipment breakdown from the end-user. The undertaking shall</li> </ul>		
<ul> <li>supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to the</li> </ul>		
<ul> <li>that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to the defective material or workmanship, shall be added to the</li> </ul>		
<ul> <li>use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>		
<ul> <li>the end-user after testing and commissioning.</li> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>	•	
<ul> <li>b) Preventive maintenance at least every six (6) months or according to the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>		
<ul> <li>the manufacturer's recommendations;</li> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>		
<ul> <li>c) Corrective maintenance within five (5) calendar days upon notification from the end-user regarding equipment breakdown/defects.</li> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>		
notification from the end-user regarding equipment breakdown/defects. d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period. e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the	,	
<ul> <li>d) The number of days where the equipment is unusable due to equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>		
<ul> <li>equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>	notification from the end-user regarding equipment breakdown/defects.	
<ul> <li>equipment defects/faults shall be added to the warranty period.</li> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>		
<ul> <li>e) The supplier shall specify post-warranty comprehensive preventive maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>		
<ul> <li>maintenance costs including list and prices of major spare parts of the equipment for three (3) years after the warranty period.</li> <li>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the</li> </ul>		
equipment for three (3) years after the warranty period. 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the		
5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the		
corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the		
the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the		
include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the		
unusable due to defective material or workmanship, shall be added to the	the equipment breakdown from the end-user. The undertaking shall	
	include a statement that the number of days where the equipment is	
warranty period.	warranty period.	

6. Manuals: The supplier must provide the end-user one (1) hard and one	
(1) soft copy of the following:	
a) Service manual in English language	
b) Operation manual in English language	
7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.	
X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-X-	
Source of Fund: NEP 2025 (Short of Award) Recipients:	
Ospital ng Paranaque II - 1 unit	

For Sole Proprietorships)

(use Bidder's Official Letterhead)

# SPECIAL POWER OF ATTORNEY

I, \_\_\_\_\_\_, Filipino, of legal age, doing business under the trade name and style of "\_\_\_\_\_\_", duly organized and existing under Philippine laws, with principal office address at \_\_\_\_\_\_\_ hereby name, constitute, and appoint \_\_\_\_\_\_ (Name of Attorney-in-Fact) as my authorized representative and attorney-in-fact to do, execute, and perform any and all acts necessary to participate, submit bids, sign and execute documents and instruments, including the Bid Securing Declaration and/or to represent me in any and all bidding proceedings conducted by the Metro Manila Center for Health Development for the Bid Project \_\_\_\_\_ (Indicate Bid Project Title and No.):

I hereby grant, unto my said attorney-in-fact, full power and authority, to do, execute and perform all acts necessary or proper to render effective the power above-stated, as fully and effectively as I might or could lawfully do if personally present, and hereby ratifying and confirming all that my said attorney-in-fact shall do with full power of substitution and hereby further confirms all that said representative shall lawfully do or cause to be done by virtue hereof.

IN WITNESS WHEREOF, I have hereunto affixed my signature on this \_\_\_\_\_ day of

\_\_\_\_\_, 20 at \_\_\_\_\_.

Affiant/Principal

Attorney-in-Fact

Signed in the Presence of:

Witness

Witness

# NOTE: PLS. USE THIS FORM ONLY IF THE REGISTERED PROPRIETOR OPTS TO

AUTHORIZE ANOTHER PERSON TO REPRESENT HER/HIM TO DO, EXECUTE, AND PERFORM ANY AND ALL ACTS NECESSARY IN ORDER TO PARTICIPATE, SUBMIT BIDS, SIGN AND EXECUTE DOCUMENTS PERTAINING TO THE BID PROJECT.)

**For FORMED Joint Venture - Incorporated**) (use Bidder's Official Letterhead) REPUBLIC OF THE PHILIPPINES) ) S.S.

# SECRETARY'S CERTIFICATE

I, (Na	me of the Corpo	orate Secretary), the Corporate Secretary of the
(Name of th	e Corporation), a	a corporation duly organized and existing under
	<b>L</b> , , ,	Philippines with principal office at
		f the Corporation) (the "Corporation"), after
having been duly sworn according	`	reby certify that at the meeting of the Board of
<i>e i e</i>		held on (Date of the meeting)
at	_ (Place of the r	meeting) at which a quorum was present and
acted throughout, the following reso	olutions were una	animously approved and adopted through Board
Resolution No (Indicate Board	Resolution No.), S	Series of 20:
"RESOLVED, that	()	Name of the Joint Venture) is authorized to
participate in the bidding	of	(Project ID-No.)
of the	Metro Manila C	Center for Health Development ("MMCHD" or
the "Procuring Entity"); and that in	1 connection with	h the said bidding, the following is/are hereby
appointed and designated as the du	y authorized repr	resentative/s of the JV, granted with full power
and authority to do, execute and pe	erform any and a	all acts necessary and/or to represent the JV to
participate in the bidding of the abo	ve-mentioned pro	oject which includes to sign for and in behalf of
the JV all bid documents, submit th	e bid, and to sign	n contracts, agreements, instruments, statements,
reports, and other documents pertain	ing to the biddin	ng including the ensuing contract with MMCHD
and all other documents, as may be	required:	
Name	Position	Specimen Signature
		-

acting	in	this	manner1:	
--------	----	------	----------	--

1 Any one (1) of the above signatories

2 All of the above signatories

3. Any <u>(state the number)</u> of the above signatories (in case the Board opts to have joint signing from designated representatives, i.e. any 2 jointly signing out of 4)

RESOLVED FURTHER THAT, \_\_\_\_\_ (Name of the JV):

(1) the Submits itself to the jurisdiction of the Philippine government and waives its right to question the jurisdiction of the Philippine courts; and

(2) Shall neither seek nor obtain writs of injunction or prohibition or restraining order against the MMCHD or any other agency in connection with this project to prevent and restrain the bidding procedures related thereto, the negotiating of the award of a contract to a successful bidder, and the carrying out of the awarded contract.

RESOLVED FINALLY, that the foregoing authorities shall remain in full force and effect and binding on the Corporation until notice in writing is received by MMCHD, revoking, amending, or otherwise modifying the same."

The undersigned further certifies that the foregoing resolutions have not been revoked, amended, or otherwise modified, and remain valid and subsisting.

The foregoing excerpts of the minutes of the Board meeting are true and correct and in accordance with the corporate records under my custody and are consistent with the Articles of Incorporation and By-laws of the Corporation.

IN WITNESS WHEREOF, I have hereunto affixed my signature on this \_\_\_\_\_ day of \_\_\_\_\_, 20 at \_\_\_\_\_.

Corporate Secretary

 SUBSCRIBED AND SWORN to before me, this
 day of
 , 20 at \_\_\_\_\_\_, affiant exhibiting to me his/her Competent Evidence of Identity \_\_\_\_\_\_
 issued on

\_\_\_\_\_ at \_\_\_\_\_. Doc. No. \_\_\_\_\_ Page No. \_\_\_\_\_ Book No. \_\_\_\_\_ Series of \_\_\_\_\_\_

For Corporations, Partnerships, or Cooperatives) (use Bidder's Official Letterhead) REPUBLIC OF THE PHILIPPINES) ) S.S.

## SECRETARY'S CERTIFICATE

(Name of the Corporate Secretary), the Corporate Secretary of the I. \_ (Name of the Corporation), a corporation duly organized and existing under and laws of the Philippines with principal office by virtue of the (Address of the Corporation) (the "Corporation"), after having at been duly sworn according to law, does hereby certify that at the meeting of the Board of Directors of the said Corporation duly convened and held on \_\_\_\_\_ (Date of the meeting) at (Place of the meeting) at which a quorum was present and acted throughout, the following resolutions were unanimously approved and adopted through Board Resolution No. \_\_\_\_ (Indicate Board Resolution No.), Series of 20\_\_:

"RESOLVED, that the Corporation is hereby authorized to participate in the bidding of \_\_\_\_\_\_\_ (Name of the Project and Project ID No.) of the Metro Manila Center For Health Development ("MMCHD" or the "Procuring Entity") and if awarded the project shall enter into contract with MMCHD;

RESOLVED, that in connection with the said bidding, the following is/are hereby appointed and designated as the duly authorized representative/s of the Corporation, granted with full power and authority to do, execute and perform any and all acts necessary and/or to represent the Corporation to

participate in the bidding of the above-mentioned project which includes to sign for and in behalf of the Corporation all bid documents, submit the bid, and to sign contracts, agreements, instruments, statements, reports, and other documents pertaining to the bidding including the ensuing contract with MMCHD and all other documents, as may be required:

Name	Position	Specimen Signature

acting in this manner1:

1. Any one (1) of the above signatories

2 All of the above signatories

3. Any <u>(state the number)</u> of the above signatories (in case the Board opts to have joint signing from designated representatives, i.e. any 2 jointly signing out of 4)

# RESOLVED FURTHER THAT, \_\_\_\_\_ (Name of the Corporation):

(1) the Submits itself to the jurisdiction of the Philippine government and waives its right to question the jurisdiction of the Philippine courts; and

(2) Shall neither seek nor obtain writs of injunction or prohibition or restraining order against the MMCHD or any other agency in connection with this project to prevent and restrain the bidding procedures related thereto, the negotiating of the award of a contract to a successful bidder, and the carrying out of the awarded contract.

RESOLVED FINALLY, that the foregoing authorities shall remain in full force and effect and binding on the Corporation until notice in writing is received by MMCHD, revoking, amending, or otherwise modifying the same."

The undersigned further certifies that the foregoing resolutions have not been revoked, amended, or otherwise modified, and remain valid and subsisting.

The foregoing excerpts of the minutes of the Board meeting are true and correct and in accordance with the corporate records under my custody and are consistent with the Articles of Incorporation and By-laws of the Corporation.

IN WITNESS WHEREOF, I have hereunto affixed my signature on this _	day of _	, 20
at		

Corporate Secretary

 SUBSCRIBED AND SWORN to before me, this
 day of
 , 20 at \_\_\_\_\_\_
 , affiant exhibiting to me his/her Competent Evidence of Identity \_\_\_\_\_\_
 issued on

al	·•
Doc. No	
Page No	
Book No.	
Series of	

### **Omnibus Sworn Statement (Revised)**

[shall be submitted with the Bid]

REPUBLIC OF THE PHILIPPINES ) CITY/MUNICIPALITY OF \_\_\_\_\_ ) S.S.

#### AFFIDAVIT

I, [Name of Affiant], of legal age, [Civil Status], [Nationality], and residing at [Address of Affiant], after having been duly sworn in accordance with law, do hereby depose and state that:

1. [Select one, delete the other:]

[*If a sole proprietorship:*] I am the sole proprietor or authorized representative of [Name of Bidder] with office address at [address of Bidder];

[If a partnership, corporation, cooperative, or joint venture:] I am the duly authorized and designated representative of [Name of Bidder] with office address at [address of Bidder];

2. [Select one, delete the other:]

[If a sole proprietorship:] As the owner and sole proprietor, or authorized representative of [Name of Bidder], I have full power and authority to do, execute and perform any and all acts necessary to participate, submit the bid, and to sign and execute the ensuing contract for [Name of the Project] of the [Name of the Procuring Entity], as shown in the attached duly notarized Special Power of Attorney;

[If a partnership, corporation, cooperative, or joint venture:] I am granted full power and authority to do, execute and perform any and all acts necessary to participate, submit the bid, and to sign and execute the ensuing contract for [Name of the Project] of the [Name of the Procuring Entity], as shown in the attached [state title of attached document showing proof of authorization (e.g., duly notarized Secretary's Certificate, Board/Partnership Resolution, or Special Power of Attorney, whichever is applicable;)];

- 3. [Name of Bidder] is not "blacklisted" or barred from bidding by the Government of the Philippines or any of its agencies, offices, corporations, or Local Government Units, foreign government/foreign or international financing institution whose blacklisting rules have been recognized by the Government Procurement Policy Board, <u>by itself or by relation</u>, <u>membership, association, affiliation, or controlling interest with another blacklisted person or entity as defined and provided for in the Uniform Guidelines on Blacklisting;</u>
- 4. Each of the documents submitted in satisfaction of the bidding requirements is an authentic copy of the original, complete, and all statements and information provided therein are true and correct;
- 5. [Name of Bidder] is authorizing the Head of the Procuring Entity or its duly authorized representative(s) to verify all the documents submitted;
- 6. [Select one, delete the rest:]

[If a sole proprietorship:] The owner or sole proprietor is not related to the Head of the Procuring Entity, Procurement Agent if engaged, members of the Bids and Awards

Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

[If a partnership or cooperative:] None of the officers and members of [Name of Bidder] is related to the Head of the Procuring Entity, Procurement Agent if engaged, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

[If a corporation or joint venture:] None of the officers, directors, and controlling stockholders of [Name of Bidder] is related to the Head of the Procuring Entity, Procurement Agent if engaged, members of the Bids and Awards Committee (BAC), the Technical Working Group, and the BAC Secretariat, the head of the Project Management Office or the end-user unit, and the project consultants by consanguinity or affinity up to the third civil degree;

- 7. [Name of Bidder] complies with existing labor laws and standards; and
- 8. *[Name of Bidder]* is aware of and has undertaken the responsibilities as a Bidder in compliance with the Philippine Bidding Documents, which includes:
  - a. Carefully examining all of the Bidding Documents;
  - b. Acknowledging all conditions, local or otherwise, affecting the implementation of the Contract;
  - Making an estimate of the facilities available and needed for the contract to be bid, if any; and
  - d. Inquiring or securing Supplemental/Bid Bulletin(s) issued for the [Name of the Project].
- [Name of Bidder] did not give or pay directly or indirectly, any commission, amount, fee, or any form of consideration, pecuniary or otherwise, to any person or official, personnel or representative of the government in relation to any procurement project or activity.
- 10. In case advance payment was made or given, failure to perform or deliver any of the obligations and undertakings in the contract shall be sufficient grounds to constitute criminal liability for Swindling (Estafa) or the commission of fraud with unfaithfulness or abuse of confidence through misappropriating or converting any payment received by a person or entity under an obligation involving the duty to deliver certain goods or services, to the prejudice of the public and the government of the Philippines pursuant to Article 315 of Act No. 3815 s. 1930, as amended, or the Revised Penal Code.
- IN WITNESS WHEREOF, I have hereunto set my hand this \_\_ day of \_\_\_, 20\_\_ at \_\_\_\_, Philippines.

[Insert NAME OF BIDDER OR ITS AUTHORIZED REPRESENTATIVE] [Insert signatory's legal capacity] Affiant

[Jurat] [Format shall be based on the latest Rules on Notarial Practice]