



**SUPPLEMENTAL/ BID BULLETIN NO. 1**

**IB No. 2025-031**  
**Procurement of Pregnancy Tracking Kits**  
**(EARLY PROCUREMENT ACTIVITY)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Query during Pre-bidding Conference:		
Technical Specification	Query	Response of the End User Unit
<b>Item No. 2: Portable Fetal Doppler</b>	<b>Item No. 2: Portable Fetal Doppler with CMDN</b>	<b>Item No. 2: Portable Fetal Doppler with CMDN</b>

Furthermore, this is to inform bidders that the bidding will be moved on the **29th of November, 2024** instead of November 27, 2024 due to lack of quorum

Bidders are advised to use the **following attached forms and submit them together with all required documents for the submission of bids on the 29th day of November 2024, 9:00 AM:**

This Supplemental/Bid Bulletin No. 1 shall be integral to the Bidding Documents. All other provisions indicated in the bidding documents not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 20<sup>th</sup> day of November 2024 in MMCHD

Approved by:

  
**JEREMIAS FRANCIS Y. CHAN, MD**  
Licensing Officer V / BAC Chairperson

Republic of the Philippines  
 Department of Health  
 Metro Manila Center for Health Development

**TECHNICAL SPECIFICATIONS**

Item	<b>Pregnancy Tracking Kits</b>	Qty./Unit	400 pcs
Name of Manufacturer:		Country of Origin (if applicable)	
Brand:		Model: (if applicable)	
<b>ABC: P 2,540,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p><b>Item no. 1:</b> Pregnancy Tracking Bag or OB Bag or Community Bag (sling Bag)</p> <p>Specifications:</p> <p>Fabric: Durable Water Resistant Nylon or Polyester</p> <p>Lining: Easy to Clean, possibly antimicrobial lining</p> <p>Size: Generally around 15 to 20 inches in height and 10 to 15 inches in width</p> <p>Weight: Lightweight, typically between 1 to 2.5 pounds when empty</p> <p>Spacious and Multiple zippered pockets and adjustable sling strap</p> <p>Branding: Option for customization or branding (imprinted or permanent sticker of DOH MMCHD logo)</p> <p><b>Item No. 2: Portable Fetal Doppler</b></p> <p>Specifications;</p> <p>Frequency: range from 2.0 MHz to 3.0 MHz for fetal heart rate detection</p> <p>Fetal Heart Rate; (FHR) Range: 50 to 210 beats per minute (bpm)</p> <p>Accuracy: <math>\pm</math> 1bpm for heart rate measurements</p> <p>Type: LCD or LED display</p> <p>Size: Sufficiently sized for easy readability, usually around 2-4 inches</p> <p>FHR visualization: continuous or averaged heart rate display</p> <p>Battery Level Indicator: For monitoring power status</p> <p>Detection Modes: Audio output with adjustable volume, some models may include visual display of fetal heart sounds.</p>			

Recording Capability: Some models may allow recording of FHR for monitoring trends

Auto shut off: Automatic shutdown after a period of inactivity (e.g. 1-5)

Power source: Battery operated (usually AA, AAA, or rechargeable lithium)

Battery Life: Typically 6 to 10 hours of continuous use

Durable, lightweight, and medical-grade plastic

Size: compact, usually around 20-30 cm in length

Weight: lightweight, typically under 500 grams

**Additional Requirement: Portable Fetal Doppler with CMDN**

### **Item no. 3: Digital Pressure Monitor**

Specifications;

Display: LCH Digital display

Measurement range: Pressure 0-299mmHg

Measurement method: Oscillometric method

Accuracy: Pressure +3mmHg

Power source: 4 "AA" batteries 1.5V or AC Adapter

Cuff dimensions; Approximately 145mm x 466mm/standard size

Cuff material: nylon , polyester

### **Item No. 4; Pulse Oximeter**

Specifications:

Durable, medical-grade plastic

Measurement Parameters:

SpO2 Measurement Range: 70% to 100 % (with accuracy of  $\pm 2\%$  to 80% to 100%)

Pulse Rate Measurement Range: 30 to 250 beats per minute (bpm) with accuracy of  $\pm 2$  bpm  $\pm 2\%$ (whichever is greater)

Display:

Type: LED or OLED Display

Size: Clear and easy to read; typically around 1.5 to 2 inches

Orientation: adjustable for convenience (landscape or portrait)

Backlight: optional feature for low light conditions

Operational feature:

Power Source: Battery Operated (usually AAA or lithium batteries)

Battery Life: 20 to 30 hours of continuous use, with low battery indicator

Auto shut off: after a set period of inactivity e. 10-30 seconds) to conserve battery

Dimension and Weight:

Size: compact and lightweight (typically around 5-70 grams)

Dimensions : approximately 6 to 10 cm in length: 3 to 5 cm in width

Performance Specification:

Response Time: Quick Measurement (usually within 5-10 seconds)

Item No. 5 Stethoscope

Type: Dual-head (diaphragm and bell)

Material: High quality stainless steel or aluminum for durability and acoustic performance

Diaphragm size:

Adult: 4 cm to 5 cm diameter

Pediatric: 3 cm diameter (if applicable)

Bell size: approximately 2.5cm to 3cm in diameter

Material: Flexible, latex-free PVC or similar materials

Length: typically 22 to 30 inches (56 to 76 cm)

Type: soft sealing or molded for comfort and a snug fit

Adjustibility: adjustable for individual ear canal sizes

Material: Rubber or silicone for comfort and sound isolation

Lightweight design for ease of use and portability, usually around 150 to 300 grams

Item no. 6: Digital Thermometer

Specifications:

Measurement Range: 32°C to 42.9°C (89.6°F to 109,2°F)

<p>Use for oral, rectal and axillary</p> <p>Weight: lightweight, range from 20-50grams with battery installed</p> <p>Accuracy: <math>\pm 0.1^{\circ}\text{C}(\pm 0.2^{\circ}\text{F})</math> within the normal body temperature range</p> <p>Size: compact, typically around 15-20 cm in length;</p> <p>Average Read Time: 6 second (rectal), 80 seconds(oral), 2 minutes axillary or underarm</p> <p>Power: 0.15mW (working)</p> <p>Battery: AG3 1.5V ordinary button battery</p> <p>Alarm: Thermometer beeps after temperature taken</p> <p>Item no. 7: Clipboard and Pen with filing case</p> <p>Specifications:</p> <p>Clip: Metal or strong plastic clip for secure grip</p> <p>Standard size: (A4 or letter size) approximately 9" x 12" (25 cm x 33 cm)</p> <p>Filing Case; Durable fabric or plastic, around 10" x 13" (25cm x 33 cm) or larger, multiple compartments or pockets, zipper or snap closure, handle or strap for easy carry</p> <p>Branding: option for customization or branding (imprinted or permanent sticker of DOH MMCHD Logo)</p> <p>Delivery Period: - 60 to 90calendar days from receipt of approved NTP</p> <p>Delivery Place: DOH MMCHD Pasig Warehouse</p>	
<p><b>B. Upon delivery, the following shall be complied with:</b></p> <p><b>Labeling instruction:</b></p> <p>A, On each small and bigger carton/box the following should be imprinted or stickered with a non removable or permanent sticker or label that is biding, and with residue and tearing, if removed:</p> <p><b>Philippine Government Property Department of Health NOT FOR SALE</b></p>	
<p><b>c. Additional requirement the Lowest/Single Calculated Bid (L/SCB) as part of post qualification:</b></p> <p>1. You are requested to submit within (5) five days upon receipt of this notice three (3) copies of all documents needed for Post Qualification of the following documents:</p> <ol style="list-style-type: none"> <li>a. Eligibility Documents <ol style="list-style-type: none"> <li>i. (Mayor's Permit (latest annual and quarterly)</li> <li>ii. SEC/DTI Registration,</li> <li>iii. Tax Clearance)</li> </ol> </li> </ol>	

- b. Certificate of Registration from BIR
  - c. Income Tax Returns – latest payment
  - d. Bid Bulletin
  - e. Product Sample /Brochure
  - f. Authority from the Manufacturer to Distribute the Product
  - g. License to Operate
  - h. And other documents stated in BDS
2. One (1) original sample of the manufacturer’s product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon contract award. The prototype of the labeling instruction must be part of the sample submitted; however, the technical specifications of the labeling instruction of the product must be complied with upon delivery.

## *Section VI. Schedule of Requirements*

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Quantity	Total ABC (Php)	Delivery Site	Delivered, Weeks/Months
	<i>Pregnancy Tracking kit</i>	<b>400 pcs</b>	P 2,540,000.00	DOH-MMCHD Pasig Warehouse	Sixty (60) to Ninety (90) Calendar days After receipt of NTP.
<b>1</b>	<i>Pregnancy Tracking Bag or OB Bag or Community Bag (Sling type)</i>	<b>400pcs</b>	200,000.00		
<b>2</b>	<i>Portable Fetal Doppler</i>	<b>400 units</b>	1,200,000.00		
<b>3</b>	<i>Digital Blood Pressure Monitor</i>	<b>400 pcs</b>	640,000.00		
<b>4</b>	<i>Pulse Oximeter</i>	<b>400 pcs</b>	160,000.00		
<b>5</b>	<i>Stethoscope</i>	<b>400 pcs</b>	160,000.00		
<b>6</b>	<i>Digital Thermometer</i>	<b>400 pcs</b>	120,000.00		
<b>7</b>	<i>Clipboard and Pen with filing case</i>	<b>400 pcs</b>	60,000.00		

---

Signature over Printed Name

[date of signing]

In the capacity of:

Duly authorized to sign bid for and on behalf of:

[title or other appropriate designation]

(Name of Company)

[Complete office address]

[Contact No.]

[Fax No.]

[Email Address]