



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB NO. 2024-004E
PROCUREMENT OF 2 UNITS CARTRIDGE-BASED NUCLEIC ACID
AMPLIFICATION TEST MACHINE (SHORT OF AWARD)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a fore cited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
ORIGINAL TECHNICAL SPECIFICATIONS	AMENDED
Inclusions: Desktop PC (At least Windows 10 OS), Barcode Scanner, Uninterruptible Power Supply (UPS), and Printer	Inclusions: Desktop PC (At least Windows 10 OS, At least i5, 8GB RAM, 200GB SSD for OS and 500GB Hard Drive for Storage), Barcode Scanner, Uninterruptible Power Supply (UPS), and LaserJet 3 in 1 Printer - RETAIN
Additional: Test: for TB Starter Kit: At least 12 Test kits, 50 tests per kit	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on November 20, 2023, 10:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned.

Issued this 14th day of November, 2023 in MMCHD.

Approved by:

SGD
PRETCHELL P. TOLENTINO, MD, MCHM
Director III / BAC Chairperson

Section VII. Technical Specifications

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	CARTRIDGE-BASED NUCLEIC ACID AMPLIFICATION TEST MACHINE	Qty./Unit	2 UNITS
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 5,000,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>TECHNICAL SPECIFICATIONS: Optics: Dye detection limit < 1Nm Reaction Time Thermal Controls - Solid state heater and forced-air cooling at each site - Reaction chamber thermistors should be calibrated to ± 1.0°C - Up to 4 independently-controlled reaction sites Cartridges - Single-use disposable cartridges - Polypropylene construction Performance Parameters - Heating ramp rates (max.): 10°C/sec from 50°C to 95°C - Cooling ramp rates (max.): 2.5°C/sec from 95°C to 50°C - Temperature duration accuracy: ± 1.0 sec from programmed time - Temperature accuracy: ± 1.0°C from 60°C to 95°C - Melt curve programmable ramp rates: 0.01°C/sec to 1.0°C/sec Dimensions: Manufacturer's Standard Power Supply: 220V, 60Hz Inclusions: Desktop PC (At least Windows 10 OS), Barcode Scanner, Uninterruptible Power Supply (UPS), and Printer Add: Test: for TB Starter Kit: At least 12 Test kits, 50 tests per kit</p>			
<p>Requirements if awarded the Contract 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 90 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment. 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p>			

5. **Notarized undertaking** that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

7. With "**DOH-MMCHD HFEP**"(Government Property not for sale) sticker in each unit

x-x

Source of Fund: NEP 2024 (Short of Award)



Name of Company: _____
Address: _____
Signature Over Printed Name : _____
Telephone/Fax Number : _____
Email: _____

Section VI. Schedule of Requirements

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Description	Quantity	Delivery Site	Delivery Period
CARTRIDGE-BASED NUCLEIC ACID AMPLIFICATION TEST MACHINE	2 units	Health Facilities in The City of Marikina	90 calendar days upon receipt of the Notice to Proceed

Name of Company: _____

Address: _____

Signature Over Printed Name : _____

Telephone/Fax Number : _____

Email: _____