



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

IB 2021 – 001E to 011E
Bid Bulletin No. 1
March 5, 2021

This Bid Bulletin will form an integral part of the bidding document for the following:

IB No.	Specifications Based on Purchase Request by End - User	Additional/Revision to the Specifications
<p>IB 2021 – 001E</p>	<p>Procurement of 5 units Fetal Monitor under Health Facilities in the National Capital Region (HFEP) ABC: P 800,000.00</p> <p>Technical Specifications: Dual pulsed Doppler FHR Ultrasound Frequency: 1 MHz Range: 30-240 bpm, 50-210 bpm With auto-detection of dual fetal movement With array-type thermal printer With automatic printing function Approx. 7-inch. TFT color LCD Display 3-channel display: FHR I, FHR II, UC With Clinical event marking Up to 150 patient data saving, 450 hrs (3hrs/patient) With USB data saving Dimension (H x L x D): approx. 80 x 330 x 280 mm Can be connected to a central monitoring system through PC interface (optional) Includes the ff. accessories: 2 units Ultrasound Doppler probe, 1 unit event marker jack, 3 units probe belt, 1 unit UC probe, 1 unit ultrasound gel, 2 units probe holder, Battery</p> <p>Additional Requirements:</p> <ul style="list-style-type: none"> • Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate or its equivalent • Delivery period: 30-60 Calendar days • User's manual in English language <ul style="list-style-type: none"> - Operations Manual 2 copies - Service Manual 2 copies • Certification that the supplier has the capability for corrective and preventive maintenance of the unit • Certification from the manufacturer that the unit offered is brand new and not discontinued model • Demonstration at least two (2) end-users <p>Warranty Certificate:</p> <ul style="list-style-type: none"> • Minimum of two (2) years Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the Manufacturer & Bidder. • Certification on the availability of spare parts for the next 5 year <p>With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit</p> <p>Recipients: Malabon RHU - 4 Justice Jose Abad Santos General Hospital – 1</p>	<p>Dimension (H x L x D): approx. 80 x 330 x 280 mm or Manufacturer's Standard</p> <p>Delivery Area: Direct to the Recipient</p>
<p>IB 2021 – 002E</p>	<p>Procurement of 1 Unit Mortuary Refrigerator (6 Corpses) Under Health Facility Enhancement Program (Justice Jose Abad Santos General Hospital) ABC: P 850,000.00</p> <p>Technical Specifications: Capacity/ Layers: 6 Corpses/ 6 Layers Voltage: 220 V/ 50 HZ Temperature Inside the Box: 1-5 °C Quantity of Compressor Unit: 18°C Refrigeration system for each chamber is individually controlled and the temperature can be adjusted without interference.</p>	

	<p>Internal and External shells of the Refrigerator are made of high quality stainless steel materials with high corrosion materials.</p> <p>Additional Requirements:</p> <ul style="list-style-type: none"> • Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate or its equivalent. • Certification from the manufacturer that the unit offered is brand new and not discontinued model. • Delivery period: 30-60 Calendar days • User's manual in English language <ul style="list-style-type: none"> - Operations Manual 2 copies - Service Manual 2 copies • Certification that the supplier has the capability for corrective and preventive maintenance of the unit. • Certification from the manufacturer that the unit offered is brand new and not discontinued model. • Training/ Demonstration of at least 2 users in operation and basic maintenance shall be provided. • Certification that the supplier will be responsible for the notification, transportation to site, delivery, installation and testing on the site (hospital / health facility) and expenses for such will be on the account of the supplier. <p>Warranty Certificate:</p> <ul style="list-style-type: none"> • Minimum of one (1) year Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the Manufacturer & Bidder. • Availability of local service center for out of warranty repair and preventive maintenance services of the unit. <p>With "DOH-MMCHD HFEP"(Government Property not for sale) sticker attached to the equipment.</p>	<ul style="list-style-type: none"> • Delivery period: 90 - 120 Calendar days <p>Delivery Area: Direct to the Recipient</p>
<p>IB 2021 – 003E</p>	<p>Procurement of 8 sets Laryngoscope Complete Set Under Health Facilities Enhancement Program (Novaliches District Hospital) ABC: P 560,000.00</p> <p>Comprehensive Fiberoptic Laryngoscope Set (Adult/ Pedia/ Newborn/ Neonate) Description: Handles made of solid aluminum alloy or Stainless, pistol gripped ergonomic designed of "up and away" lifting technique Handles are fully autoclavable and covered with lifetime guarantee</p> <p>1 Regular Handle 1 Pen Handle 1 Stubby Handle Blades: Durable and designed Dual Lighted blades with UV and LED technology for optimized high definition viewing of the area With Storage bag Macintosh Blade (Mac 1, 2, 3, 4) Miller Blades (0, 1, 2, 3) Battery Operated</p> <p>Additional Requirements: Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate or its equivalent Delivery Period: 30-60 Calendar days User's Manual in English Language >Operation Manual 2 copies >Service Manual 2 copies Certification from the manufacturer that the unit offered is brand new and not a discontinued model Demonstration of at least 2 users in operation and basic maintenance shall be provided. Certification that the supplier will be responsible for the notification, transportation to site, delivery, installation and testing on the site (hospital / health facility) and expenses for such will be on the account of the supplier.</p> <p>Warranty Certificate:</p>	<p>made of solid aluminum alloy or Stainless, pistol gripped ergonomic designed of "up and away" lifting technique</p> <p>Handles are fully autoclavable and covered with 1 to 2 years or higher guarantee</p> <p>Dual Lighted blades with LED and HALOGEN technology for optimized high definition viewing of the area</p>

	<ul style="list-style-type: none"> • Minimum of one (1) year Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the Manufacturer & Bidder. <p>Certificate for parts and craftsmanship upon delivery, inspection, and acceptance dully issued by the Manufacturer and Bidder With "DOH-MMCHD HFEP" (GOVERNMENT PROPERTY NOT FOR SALE) sticker in each unit</p>	<p>Delivery Area: Direct to the Recipient</p>
<p>IB 2021 – 004E</p>	<p>Procurement of 3 sets Cardiac Monitor Under Health Facilities Enhancement Program (Novaliches District Hospital) ABC: P 660,000.00</p> <p>Description: ECG, SpO2, NIBP, 2TEMP, PR, RESP 12.1" color TFT LCD screen Touchscreen 8 waveform display Powerful calculation (Drug Dose Oxygenation, Ventilation) Pacemaker detection ST & arrhythmia analysis OxyCRGs screen Wired/Wireless CMS, support HL7 protocol to HIS SpO2 pulse-tone modulation (Pitch Tone) MEWS (Modified Early Warning Score) Graphical & tabular trend review (120 hours) Rechargeable Lithium-Ion Battery (2600 mAh) EtCO2 Connection ready Package includes: 5 lead ECG patient cable NIBP hose NIBP cuffs: Adult (27–35cm) and Child (17–25cm) Adult Reusable Finger SpO2 sensor Skin & ESO/REC Temp sensors Local Trolley with basket</p> <p>Additional Requirements: >Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate or its equivalent. >A certification that the unit being offered must be brand new and not a discontinued model. >Delivery Period: 30-90 Calendar days >Certification to provide user's Manual in English Language. >Operation Manual 2 copies >Service Manual 2 copies >Certification that the supplier has the capability for corrective and preventive maintenance of the unit. >Training/ Demonstration of at least 2 users in operation and basic maintenance shall be provided. >Certification that the supplier will be responsible for the notification, transportation to site, delivery, installation and testing on the site (hospital / health facility) and expenses for such will be on the account of the supplier. Warranty Certificate: • Minimum of two (2) years Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the Manufacturer & Bidder. Certificate for parts and craftsmanship upon delivery, inspection, and acceptance dully issued by the Manufacturer and Bidder With "DOH-MMCHD HFEP" (GOVERNMENT PROPERTY NOT FOR SALE) sticker in each unit</p>	<p>Powerful calculation (Drug Dose Oxygenation, Ventilation) or optional</p> <p>Capable for central monitoring system (CMS)</p> <p>MEWS (Modified Early Warning Score) or Equivalent</p> <p>Rechargeable Lithium-Ion Battery (2550 mAh or higher) up to 2 hours monitoring or higher</p> <p>NIBP cuffs: Adult (27 - 35cm) and Child (17-25cm)</p> <p>Delivery Area: Direct to the Recipient</p>
<p>IB 2021 – 005E</p>	<p>Procurement of 1 Unit Ultrasound Machine Under Health Facilities in the City of Muntinlupa (HFEP) ABC: P 630,000.00</p> <p>Technical Specifications</p> <ul style="list-style-type: none"> • 15" LCD monitor on an articulating arm • Full size keyboard with user-defined keys • Automatic Optimization (AO) 	

	<ul style="list-style-type: none"> • Speckle Reduction Imaging (SRI) • 3D/4D Imaging • TruScan Architecture • SmartScan • ComfortScan • 3 Active Probe Ports • Integrated HDD & DVD-R/W • On-board peripheral storage • Frequency: 50/60 Hz • Voltage: 100-120 Vac or 220-240 Vac <p>Patient data can be seamlessly exported to a USB</p> <ul style="list-style-type: none"> • Height (Adjustable) <p>Monitor:</p> <ul style="list-style-type: none"> • Type: TFT LCD o Resolution: 1024 x 768 pixels o Recording Area: 800 x 800 • Adjustable: Yes, Tilt / Swivel • Brightness / Contrast: Digital Adjustment <p>Additional Requirements:</p> <ul style="list-style-type: none"> • Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate or its equivalent • Certification from the manufacturer that the unit offered is brand new and not a discontinued model. • Training/ Demonstration of at least 2 users in operation and basic maintenance shall be provided • User's manual in English language of the following: <ul style="list-style-type: none"> - Operations Manual (1 hard copy and 1 soft copy) - Service Manual (1 hard copy and 1 soft copy) • Certification that the supplier will be responsible for the notification, transportation and, delivery to MMCHD/health facility and expenses for such will be on the account of the supplier. • Minimum of two (2) years Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the Manufacturer & Bidder • Delivery period: 30-60 Calendar days <p>With "DOH-MMCHD HFEP"(Government Property not for sale) sticker</p>	<ul style="list-style-type: none"> • TruScan Architecture or its equivalent • SmartScan or its equivalent • ComfortScan or its equivalent • 3 Active Probe Ports (Manufacturer's Standard)
<p>IB 2021 – 006E</p>	<p>Procurement of 1 Unit of Operating Light LED Light Under Health Facilities Enhancement Program (Gat Andres Bonifacio Memorial Medical Center) ABC: P 2,000,000.00</p> <p>(Dual Ceiling Type) Minimum colour rendering index of 96% (Ra) at all dimming steps and temperature steps Maximum power consumption of main light & satellite 60W respectively at 100% light output Technology for compensation of ageing process of LED, for a 100% light power during the lifetime of the light Green light for endoscopic surgery Light head hermitically sealed with protection class IP 54 suitable for wet and spray disinfection Average of air turbulence intensity Total irradiance at 100,000 lx: <325W/m2 Service life of LED lamp should be 40,000 hours or higher Electrical data: External power supply units Light: Cool Light with no infrared components Optimum illumination in area and depth Shadow free with the ideal ratio of luminaire to light-field diameter Natural color rendering index with white light (Ra of 96 at all light levels, R9 of 95 in the red region) Can be regulated from 22,000 lux to 160,000 Hygiene: Flow-optimized luminaires housing for laminar airflow ceiling Degree of protection IP 54 as per IEC 144 Suitable for wipe disinfection Color Temperature Selection: Individually adjustable Three levels from 3,800 K to 4,800 K</p>	

	<p>High contrast light for different types of tissue</p> <p>Control Elements: Simple and clear structure of control elements Sensor keys on the adjustable arm with hygiene compatible glass surface All functions are quickly and accurately accessible Disinfectant resistant surfaces easy to clean Adjustable illuminance Adjustable color temperature endoscopy mode</p> <p>Lighting: Colour rendering index for red colours R9=99 for natural and detailed rendition of tissue structure Switch-over for endoscopic illumination (green light) Stepless dimming from 20% to 100% or in 5 steps from 30% to 100% Diameter of shadow less light field adjusted via centrally positioned sterile handle 215mm to 290mm With Illumination depth of 600 mm</p> <p>Installation: Installation, testing and certification of equipment operation shall be performed by qualified and factory-trained service engineers</p> <p>Additional Requirements: Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate or its equivalent Delivery Period: 30-90 Calendar days Certification that the supplier has a capability for corrective and preventive maintenance of the unit. User's Manual in English Language >Operation Manual 2 copies >Service Manual 2 copies Certification that the supplier has the capability for corrective and preventive maintenance of the unit Training/ Demonstration of at least 2 users in operation and basic maintenance shall be provided. Certification that the supplier will be responsible for the notification, transportation to site, delivery, installation and testing on the site (hospital / health facility) and expenses for such will be on the account of the supplier. Certification that the supplier shall provide applications training for users.</p> <p>Warranty Certificate: • Minimum of one (2) years Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the Manufacturer & Bidder. Certificate for parts and craftsmanship upon delivery, inspection, and acceptance dully issued by the Manufacturer and Bidder</p> <p>With "DOH-MMCHD HFEP" (GOVERNMENT PROPERTY NOT FOR SALE) sticker in each unit</p>	
<p>IB 2021 – 007E</p>	<p>Procurement of 1 Unit Defibrillator Under Health Facilities Enhancement Program (Navotas City Hospital) ABC: P 735,000.00</p> <p>Size with paddle include: not more than 330mm x 260mm x 360 mm LCD Dimension: 7 inches diagonal or at least 150 mm x 90 mm; Color TFT with atleast 800 x 480 pixels resolution. Truncated exponential biphasic (e-cube) waveform (200J) Shock delivery via Multifunction defibrillator electrode pads; External paddle with pediatric Internal paddle. Shock advisory sensitivity and specificity meets the Association for the Advancement of Medical Instrumentation (AAMI) DF-80 guidelines. Guides the user through the protocol via multifunction defib electrode pads. With non-invasive ECG monitoring and SpO2 monitoring Continuous ECG strip Real-time Battery capacity: when new, minimum of 100 shock deliveries Lithium polymer external battery pack. Battery type: 14.8V3.1Ah (Rechargeable) Charge Time: Adapter less than 5 second to 200 Joules, Battery less than 7 seconds to 200 Joules</p>	<p>Size with paddle include: not more than 330mm x 260mm x 360 mm or manufacturer's standard Color TFT with at least 640 x 480 pixels resolution. pwedi po 640 x 480 pixels (VGA) Truncated exponential biphasic (e-cube) waveform (200J) or its equivalent Shock advisory sensitivity and specificity meets the Association for the Advancement of Medical Instrumentation (AAMI) DF-80 guidelines or Manufacturer's Standard</p> <p>Lithium polymer or Lithium ion external battery pack. Battery type: 10.8v (Rechargeable)</p>

	<p>Weight not more than 5kg with external paddle</p> <p>Additional Requirements: >Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate or its equivalent >A certification that the unit being offered must be brand new and not a discontinued model >Delivery period: 15-30 Calendar days >Certification to provide user's Manual / Catalogue in English Language. - Operations Manual / Catalogue 2 copies - Service Manual / Catalogue in English Language Training/ Demonstration of at least 2 users in operation and basic maintenance shall be provided >Certification that the supplier has a capability for corrective and preventive maintenance of the unit. >Certification that the supplier will be responsible for the notification, transportation to site, delivery, installation and testing on the site (hospital/health facility) and expenses for such will be on the account of the supplier.</p> <p>Warranty Certificate: >Minimum of two (2) years Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the manufacturer & bidder.</p> <p>With "DOH-MMCHD HFEP" (GOVERNMENT PROPERTY NOT FOR SALE) sticker attached to the equipment.</p>	<p>Weight minimum of 5kg or 5kg to 7kg with external paddle</p> <p>Delivery period: 30 - 60 Calendar days</p>
<p>IB 2021 – 008E</p>	<p>Procurement of 28 Units Dental Unit, Motorized with Accessories Under Health Facilities in the National Capital Region (HFEP) ABC: P 5,040,000.00</p> <p>Technical Specifications:</p> <ul style="list-style-type: none"> - Fully motorized with pre-set functions - Touch control chair - Slim ergonomic design upholstery - LED light with sensor switch - Double side armrest - Colored glass/ ceramic bowl, rotating detachable cuspidor bowl (with plastic strainer) - Two (2) pcs. 3 way syringe - Three (3) pcs handpiece attachments (2 hi-speed adaptors & 1 low-speed adaptor) - Auto-cup-filler & bowl flush (water flushing control valve, cup-filler with control valve) - Saliva ejector-built-in motor pump type, does not use air or water - Clean water system - LED X-ray film viewer (Peri-apical X-ray viewer) - Multi-functional foot control - Water tank for distilled water for handpiece - Two (2) hi-speed handpiece - One (1) low-speed complete with straight handpiece & Contra-angle - LED light cure unit (external) - Tissue box with tray and cup dispenser - Silent type and oil-free dental air compressor - Saddle-type and backrest dentist stool <p>Additional Requirements:</p> <ul style="list-style-type: none"> *Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate. • A certification that the unit being offered must be brand new and not a discontinued model. • Delivery period: 30-60 Calendar days • Certification to provide user's manual/catalogue in English language <ul style="list-style-type: none"> - Operations Manual/Catalogue 2 copies - Service Manual/Catalogue 2 copies * Certification that the supplier has the capability for corrective and preventive maintenance of the unit. 	

	<p>* Certification that the supplier will be responsible for the notification, transportation and, delivery to MMCHD/health facility and expenses for such will be on the account of the supplier</p> <ul style="list-style-type: none"> • Training/ Demonstration of at least 2 users in operation and basic maintenance shall be provided. <p>Warranty Certificate:</p> <ul style="list-style-type: none"> • Minimum of one (1) year Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the Manufacturer & Bidder. <p>With "DOH-MMCHD HFEP" (GOVERNMENT PROPERTY NOT FOR SALE) sticker in each unit</p> <p>Recipients: Navotas RHU - 4 Pateros RHU - 2 Marikina RHU - 3 San Juan RHU - 2 Manila RHU - 16 Las Pinas RHU - 1</p>	
<p>IB 2021 – 009E</p>	<p>Procurement of 1 Unit Thermal Scanner Under Facilities Enhancement Program (Valenzuela City Emergency Hospital) ABC: P 645,000.00</p> <ul style="list-style-type: none"> > Floor mounted > Dual Vision Human body temperature measurement > Camera-model DTC 300 features > measuring distance:0.5-3cm high accuracy +/-0.3 degree celcius > with optional black body > measurement range:0-60 degree celcius > thermal resolution: 384x 288 > visible light resolution: 2 megapixels > delivery includes: dual camera, black body, tripods > computer and software warranty: 1 year from the date of acceptance > Training included for proper operation of the equipment provided after sales service. <p>Additional Requirements:</p> <ul style="list-style-type: none"> • Current and Valid Certificate of Manufacturer’s compliance with ISO and/or CE Certificate. • A certification that the unit being offered must be brand new and not a discontinued model. • Delivery period: 30-60 Calendar days • Certification to provide user’s Manual in English Language. <ul style="list-style-type: none"> - Operations Manual 2 copies - Service Manual 2 copies • Certification that the supplier has a capability for corrective and preventive maintenance of the unit. • Certification that the supplier will be responsible for the notification, transportation to site, delivery, installation and testing on the site (hospital / health facility) and expenses for such will be on the account of the supplier. • Certification that the supplier shall provide applications training for users. <p>Warranty Certificate:</p> <ul style="list-style-type: none"> • Minimum of one (2) years Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the Manufacturer & Bidder. • Availability of local service center for out of warranty repair and preventive maintenance services of the unit. <p>With "DOH-MMCHD HFEP" (Government Property not for sale) sticker attached to the equipment.</p>	<p>Warranty Certificate:</p> <ul style="list-style-type: none"> • Minimum of two (2) years Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the Manufacturer & Bidder.

<p>IB 2021 – 010E</p>	<p>Procurement of 25 Unit Fetal Doppler Under Health Facilities in the National Capital Region (HFEP) ABC: P 1,125,000.00</p> <p>Technical Specifications:</p> <ul style="list-style-type: none"> ● Power Supply: 220 VAC, 60Hz ● Type: Table top (heavy duty) ● Material: Made of hard plastic/ ABS Plastic <p>Standard Features:</p> <ul style="list-style-type: none"> ● LCD Display ● Fetal Heart Rate ● Built-in loudspeaker: with noise reduction ● Probe holder <p>Standard Accessories:</p> <ul style="list-style-type: none"> ● 1 conductive cream ● 1 ultrasonic probe <p>Additional Requirements:</p> <ul style="list-style-type: none"> • Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate or its equivalent • Delivery period: 15-30 Calendar days • User's manual in English language <ul style="list-style-type: none"> - Operations Manual 2 copies - Service Manual 2 copies • Training/ Demonstration of at least 2 users in operation and basic maintenance shall be provided • Certification that the supplier will be responsible for the notification, transportation to site, delivery, installation and testing on the site (hospital / health facility) and expenses for such will be on the account of the supplier. • Certification from the manufacturer that the unit offered is brand new and not discontinued model <p>Warranty Certificate:</p> <ul style="list-style-type: none"> • Minimum of one (1) year Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the Manufacturer & Bidder. • Availability of local service center for out of warranty repair and preventive maintenance services of the unit. <p>With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit</p> <p>Recipients: Valenzuela City Emergency Hospital - 3 Malabon - 4 Marikina - 6 Las Piñas - 8 San Juan - 4</p>	<ul style="list-style-type: none"> ● LCD or LED Display ● Built-in loudspeaker: with or with no noise reduction <p>Delivery period: 30 to 60 Calendar days c/o end user</p>
<p>IB 2021 – 011E</p>	<p>Procurement of 17 Units Autoclave Under Health Facilities in the National Capital Region (HFEP) ABC: P 850,000.00</p> <p>Technical Specifications:</p> <ul style="list-style-type: none"> ● Chamber: Stainless steel horizontal ● Chamber Capacity: 20 Liters ● Power Supply: 220 V - 3 Phase, 60 Hz ● Heater: 3 Kilo watts or higher ● Safety Devices: <ul style="list-style-type: none"> > (low water) Cut off switch > Safety release valve > Emergency exhaust valve > Door safety lock > Pressure gauge with pressure control switch > Steam trap > Timer with alarm > Standard accessories > Stainless steel tray (1) piece <p>Additional Requirements</p> <ul style="list-style-type: none"> ● Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate or its equivalent. ● User's Manual in English Language <ul style="list-style-type: none"> > Operation Manual 2 copies > Service manual 2 copies 	

<ul style="list-style-type: none"> • Certification from the manufacturer that the unit offered is brand new and not a discontinued model. • Training/ Demonstration of at least 2 users in operation and basic maintenance shall be provided • Delivery period: 15-30 Calendar days <p>Warranty Certificate:</p> <ul style="list-style-type: none"> • Minimum of one (1) year Warranty Certificate for parts and craftsmanship upon delivery, inspection and acceptance duly issued by the Manufacturer & Bidder <p>With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit</p> <p>Recipients: Caloocan RHU - 1 Navotas RHU - 4 Valenzuela RHU - 8 Marikina RHU - 3 San Juan RHU - 1</p>	<ul style="list-style-type: none"> • Delivery period: 30 - 60 Calendar days
---	--

General instructions:

1. Supplier/s who opted to avail of the bidding documents but did not attend the Pre-Bidding Conference is/are not exempted from the terms and conditions stipulated in the bid bulletin;
2. Philippine Standard Time (PST) is the basis of time for the conduct of Public Bidding, for which the Biometrics located at Building I shall be used as reference for the PST; The BAC will not accept late submission of Bidding Documents.
3. Person/s not included in the Letter of Authority by the company will not be entertained including their bidding documents. **Only One (1) representatives** is allowed inside the bidding room.
4. **Place your bidding documents in an assigned envelope. Observe and comply with the following: a) properly labeled, b) with table of contents, c) properly arranged, with tabbing (marker) and d) fastened properly.** Unarranged and unfastened Bidding Documents will not be read. Bidders who submitted Unarranged and unfastened Bidding Documents will be disqualified to join the Public Bidding. Bidders must submit:
 - I. One (1) original
 - II. Two (2) duplicate copies
5. **Financial Proposal must be placed inside the YELLOW ENVELOPE.** Non-compliance with this instruction will be rated "FAILED" by the BAC.
6. Bidding Quotation must be in **WORDS AND IN FIGURES**. Non-compliance with this instruction will be a ground for disqualification from the Public Bidding;
7. Only Xerox copy of official receipt will be placed inside the bidding envelope for **bid security of cash or managers check; (DOH- NCRO – BTR)**
8. **The following Bidding Documents must be notarized (noncompliance with this instruction will be rated as "FAILED BID"):**
 - A. Statement of all on going gov't and private contracts including contracts not yet started
 - B. Statement of Single Largest Contract (SLCC)
 - C. Letter of Authority
 - D. Omnibus Sworn Statement
 - E. Bid Securing Declaration
9. NFCC shall be accompanied by Comparative Financial Statement of CY 2019 and CY 2018.
10. GPPB Cir: No. 07-2017, 4.1.1 states that
"for all procurement projects advertised and/or posted after the effectivity of this Circular, bidders may still submit their Class A eligibility documents required to be uploaded and maintained current and updated registered in the PHILGEPS under Platinum category, their Certificate of Registration and Membership in lieu of their uploaded file of Class A documents, or a combination thereof., the Certificate of PHILGEPS Registration (Platinum Membership) shall remain as post qualification requirement to be submitted in accordance with Sec 34.2 of the 2016 Revised IRR of RA 9184
11. The declared Lowest Calculated and Responsive Bidder to whom the contract shall be awarded after due deliberation and resolution must receive the Contract within seven (7) calendar days from the approval thereof. Failure to comply shall be a ground for cancellation thereof.

12. Post Qualification Requirements:

- If Platinum Certificate is submitted during the Bid Opening, the following business licenses should indicate “certified true copy”:
 - Mayors permit
 - SEC/DTI/CDA
 - Tax Clearance
- 2019, 2018 Comparative FS and EFPS Files
- Six months Tax Returns
- Other related documents required by the Technical Working Group

13. Bidding Date: March 17, 2021

Time of bidding Docs submission: ON OR BEFORE 9:00 AM (closing time)
Time of Opening of Bids: 10:00 AM

IMPORTANT REMINDERS WHICH WILL FORM PART OF THE BID BULLETIN

For the submission of bid documents for the Public Bidding:

1. Observe and comply with proper labelling and tabbing of documents
2. Include table of contents
3. The remarks “COMPLY” should be indicated line by line (corresponding to each parameter of the specifications)

EXAMPLE:

Technical Specifications		
Item	Specifications	Statement of Compliance
1	Procurement of 10,001 Boxes of Paracetamol 500 mg/tablet, 100s per box ABC: P50,000,000.00	COMPLY
	Unit delivery, the following must be considered: Shelf Life: Supplier must be fresh contained stock with a total shelf life of not less than eighteen (18) months from the date of delivery.	COMPLY
	Packaging: Instructions: Standard packaging of the manufacturer	COMPLY
	Labelling Instructions: - Each bottle and box should indicate: - Manufacturing Date - Expiration Date - Each bottle, box and carton should legibly (printed/ stamped) with: - “DOH-NCIP Philippine Government Property” and EOE, Bid	COMPLY
	Area: Place of Delivery: Pasig Warehouse, Calancho Ave Pasig City	COMPLY

4. Use the standard format for Statement of Single Largest Completed Contract and On-Going Contract/s (e.g with name, signature, date)
5. Separate the Statement of Single Largest Completed Contract and Statement of On-Going Contract/s
6. Notarial date should not be earlier than the date of signing on bid documents required to be sworn (such as but not limited to Statement of Single Largest Completed Contract and Statement of On-Going Contract/s, etc)
7. The Bidder must submit a computation of its Net Financial Contracting Capacity (NFCC) which must be at least equal to the ABC to be bid, calculated as follows:

$NFCC = [(Current\ assets\ minus\ current\ liabilities) \times 15]$ minus value of all outstanding or uncompleted portions of the projects under ongoing contracts, including awarded contracts yet to be started coinciding with the contract to be bid.

8. Use the 2016 version of Omnibus Sworn Statement as per Section 25.3 of the 2016 RIRR
9. Copy of the BIR Form 2303 indicating the correct classification / category of the nature of business.
10. Website: ncroffice.doh.gov.ph

For guidance and information of all concerned.


JOSE MARI C. CASTRO, MD, CFP, MPH, MPA
 BAC Chairperson