



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 014E

Procurement of Supply Delivery, and Commissioning of Five (5) units Brand New Ambulance (Type 1) for the use of Facilities in the Mandaluyong City Under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for Health Development

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Addition and clarification to provisions/specifications in the Bidding Documents:

- | |
|---|
| <ul style="list-style-type: none">• Exterior to include Markings: To be provided by the end-user (AO 2020-0011 Annex D) |
|---|

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12 July 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 5th day of July, 2021 in MMCHD.

Approved by:

A handwritten signature in blue ink, appearing to read "J. Chan", is written over the printed name of the signatory.

JEREMIAS FRANCIS Y. CHAN, MD, MPH
BAC Chairperson



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 015E

Procurement of Supply Delivery, and Commissioning of Eight (8) units Brand New Ambulance (Type 1) for the use of seven (7) Facilities of Mandaluyong City Under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for Health Development

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Addition and clarification to provisions/specifications in the Bidding Documents:

- | |
|---|
| <ul style="list-style-type: none">• Exterior to include Markings: To be provided by the end-user (AO 2020-0011 Annex D) |
|---|

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12 July 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 5th day of July, 2021 in MMCHD.

Approved by:

A handwritten signature in blue ink, appearing to read "J. Chan", is written over the printed name of the chairperson.

JEREMIAS FRANCIS Y. CHAN, MD, MPH
BAC Chairperson



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 016E

Procurement of 16 units Patient Monitor for OSPITAL NG MAYNILA MEDICAL CENTER Under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for Health Development

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Addition and clarification to provisions/specifications in the Bidding Documents:	
<ul style="list-style-type: none">• Delivery Schedule: 60 to 90 Calendar Days• ECG Cable: Manufacturer's standard• SpO2 Finger Sensor: Manufacturer's standard	
Revision to the specifications in the Bidding Documents	
FROM	TO
Rechargeable Lithium-Ion Battery: 10.8V, 2550mAh, (good for 8 hours continuous use)	Rechargeable Lithium-Ion Battery: Manufacturer's standard, (4-8 hours continuous use or Manufacturer's standard)
NIBP Cuff: Adult 27cm - 35cm	NIBP Cuff: Manufacturer's standard
EVA Pad brand	EVA Pad brand (Remove)
EVA Pad – Anti Vibration Pad	EVA Pad – Anti Vibration Pad (Remove)

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12 July 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 5th day of July, 2021 in MMCHD.

Approved by:

JEREMIAS FRANCIS Y. CHAN, MD, MPH
BAC Chairperson



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 017E

Procurement of Supply Delivery, Installation, Testing and Commissioning of various Medical Equipments for CALOOCAN CITY MEDICAL CENTER (CCMC) Under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for Health Development (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Addition and clarification to provisions/specifications in the Bidding Documents:

- Delivery Schedule: 60 to 90 Calendar Days
- 1 UNIT COLONOSCOPE – compatible with the existing available light source (Light source brand: Pentax EC-380LKP)

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12 July 2021, 9:00 AM

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 5th day of July, 2021 in MMCHD.

Approved by:

A handwritten signature in blue ink, appearing to read "J. Chan", is written over the printed name below.

JEREMIAS FRANCIS Y. CHAN, MD, MPH
BAC Chairperson



Republic of the Philippines
 Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 018E

**Procurement of various Medical Equipments for NOVALICHES DISTRICT HOSPITAL
 Under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for
 Health Development (LOT BIDDING)**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Addition and clarification to provisions/specifications in the Bidding Documents:	
<ul style="list-style-type: none"> • Delivery Date: According to DOH Standards 	
Revision to the specifications in the Bidding Documents	
ITEM A. 10 UNITS CARDIAC MONITORS	
FROM	TO
Viewing angle: ±15	Viewing angle: Manufacturer's standard
ECG size (sensitivity): 4.0, 2.0, 1.0, 0.5, 0.25 cm/mV or Auto Capable of Basic Arrhythmia and Analysis	ECG size (sensitivity): Manufacturer's standard
Etheral Port:	Ethernet Port (Typo Error)
VGA Port: Connecting an External/ Secondary Display	VGA Port: Connecting on External/Secondary Display or its equivalent
Mechanical Vibration: .30 grams RMS Frequency Range: 5-500 Hz Acceleration Amplitude: 5-350 Hz: 0.0002 g2/Hz	Mechanical shock and vibration: Manufacturer's standard
Standard Accessories: <ul style="list-style-type: none"> • 1 pc - 3 Lead ECG Trunk cable and grabber set 	Standard Accessories: <ul style="list-style-type: none"> • 1 pc - 3 Lead ECG Trunk cable and grabber set or its equivalent
ITEM B. 1 UNIT CENTRAL MONITORING SYSTEM	
Addition and clarification to provisions/specifications in the Bidding Documents	
Additional Specifications of Accessories: <ul style="list-style-type: none"> • CPU- System input power: 100 - 240 VAC, 50/60 Hz • OS: Windows 7 embedded Video display: supports 1920x1080 native resolution or higher • Network: Ethernet speaker: Audio alarm range 45dB 85dB (+/-3dB) • Monitor-- 24" Led Monitor LCD response time 5 ms LCD resolution Fixed at 1920 x 1080 Program of Central monitor: <ul style="list-style-type: none"> • Displays waveforms, numeric data, and alarms from all connected patient monitors in network • An optional printer provides printouts of waveforms and trend data • Supports displays at 1920 x 1080 native resolution • HL7 data export over the hospital LAN to the HL7 interface server or EHR serve 	
ITEM C. 5 Units Pulse Oximeter	

Revision and clarification to the specifications in the Bidding Documents	
FROM	TO
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • High temperature stability up to $\pm 0.5^{\circ}\text{C}$ • Machine aluminum heating units • Low voltage 220-240/115v • Maximum Temperature up to 150°C • Led display time and temperature • Automatic calibration temperature sensor • Fixed safety circuit • Error code display 	<p>Technical Specification</p> <ul style="list-style-type: none"> • Oxygen Saturation (SpO2) Range: 0% to 100% • SpO2 Measurement Accuracy: Manufacturer's standard • Pulse Rate Range: 18 to 300 bpm (or higher) • Pulse Rate Accuracy: Manufacturer's standard • Display <ul style="list-style-type: none"> - Pulse Strength: LED, Bar graph, tn-color segments - Alarm Indicator: LED, bi-color - Alarm Silenced: LED, amber - Numeric Displays: 3-digit, 7-segment LED's, green - Low Battery: LED, amber • Saturation Accuracy (Arms)a: 70-100% <p>NO MOTION Adult! Pediatrics Neonates Reusable: Finger clip +-2 digits +-3 digits</p> <ul style="list-style-type: none"> • Pulse Rate Accuracy: 18 - 300 bpm (no motion) 40 - 240 (motion) <p>NO MOTION Adult! Pediatrics Neonates Reusable: Finger clip +-3 digits +-3 digits Disposable: 6000 Series +-3 digits +-3 digits</p> <p>MOTION Adult! Pediatrics Neonates Reusable: Finger clip +-5 digits +-5 digits</p> <p>LOW PERFUSION Adult! Pediatrics Neonates Reusable: Finger clip +-3 digits +-3 digits Disposable: 6000 Series +-3 digits +-3 digits</p> <p>Memory: 70 hours (assuming continuous operation)</p> <ul style="list-style-type: none"> • Temperature <ul style="list-style-type: none"> - Operating: 0° - 40° C - Storage: -30° - $+50^{\circ}\text{C}$ • Humidity <ul style="list-style-type: none"> - Operating: 10 - 90% non-condensing - Storage: 10 - 95% non-condensing • Altitude <ul style="list-style-type: none"> - Operating: Up to 12,000 meters (40,000 feet) - Hyperbaric Pressure: Up to 4 atmospheres • Power Requirements <ul style="list-style-type: none"> - Mains: 100-240 VAC, 50-60 Hz <p>Accessories: Reusable Finger Probe Sensor (Adult) – 1 pc</p>

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12 July 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 5th day of July, 2021 in MMCHD.

Approved by:

A handwritten signature in blue ink, appearing to read 'J. Chan', is written over the printed name below.

JEREMIAS FRANCIS Y. CHAN, MD, MPH
BAC Chairperson



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 019E

Procurement of various Medical Equipments for VALENZUELA CITY EMERGENCY HOSPITAL Under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for Health Development (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Addition and clarification to provisions/specifications in the Bidding Documents:
ITEM B. 1 UNIT MORTUARY REFRIGERATOR (4 CORPSE) Back Up Power UPS Specification: <ul style="list-style-type: none">• Input: 220 – 240v 50/60 Hz• Output: 220 – 240v 50/60 Hz• Capacity: 2000VA/1600W
ITEM C. 1 UNIT SLIT LAMP Slit Lamp UPS: <ul style="list-style-type: none">• Output power capacity: 650VA Slit Lamp Voltage Requirement <ul style="list-style-type: none">• Power: 220 – 240V 50/60Hz• 400VA
ITEM D. 3 UNITS PORTABLE TYPE STERILIZER <ul style="list-style-type: none">• Oven Type

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12 July 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 5th day of July, 2021 in MMCHD.

Approved by:

JEREMIAS FRANCIS Y. CHAN, MD, MPH
BAC Chairperson



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 020E

Procurement of various Medical Equipments for Ospital ng Sampaloc under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for Health Development (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Addition and Revision to provisions/specifications in the Bidding Documents:	
ITEM F. 1 SET NASOPHARYNGOSCOPE 2.7MM	
FROM	TO
Battery: Lithium Ion	Battery: Lithium-Ion Rechargeable Battery with Charger

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12 July 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 5th day of July, 2021 in MMCHD.

Approved by:

JEREMIAS FRANCIS Y. CHAN, MD, MPH
BAC Chairperson



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 021E

Procurement of various Medical Equipments for OSPITAL NG MALABON under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for Health Development (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Addition and clarification to provisions/specifications in the Bidding Documents:

- | |
|--|
| <ul style="list-style-type: none">• No changes as stated in the Technical Specifications |
|--|

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12 July 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 5th day of July, 2021 in MMCHD.

Approved by:

A handwritten signature in blue ink, appearing to read "J. Chan", is written over the printed name of the chairperson.

JEREMIAS FRANCIS Y. CHAN, MD, MPH
BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 022E

Procurement of various Medical Equipments for the use of OSPITAL NG TONDO under the Health Facilities Enhancement Program (HFEP) of DOH-Metro Manila Center for Health Development (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Addition and clarification to provisions/specifications in the Bidding Documents:	
ITEM I. 1 UNIT REAGENT REFRIGERATOR	
FROM	TO
<ul style="list-style-type: none">Digital Display Temperature	<ul style="list-style-type: none">Digital Display Temperature Thermometer
<ul style="list-style-type: none">With Fluorescent Lamp	<ul style="list-style-type: none">Fluorescent Lamp/LED Lamp
ITEM J. 5 UNITS SUCTION MACHINE HEAVYDUTY	
<ul style="list-style-type: none">Reservoir Material: Glass	
FROM	TO
<ul style="list-style-type: none">Reservoir Capacity: 500 mL/pc	<ul style="list-style-type: none">Reservoir Capacity: 2500 mL/pc
ITEM M. 10 UNITS PORTABLE UV LIGHT STERILIZER	
<ul style="list-style-type: none">UV Lamp: Manufacturer's Standard	
FROM	TO
<ul style="list-style-type: none">Voltage: 110V/220V/50HZ or Customized	<ul style="list-style-type: none">Voltage: Manufacturer's Standard (Compatible to Philippine Standard Voltage)
<ul style="list-style-type: none">Power: 40W	<ul style="list-style-type: none">Power: 60W
<ul style="list-style-type: none">Time Setting: 3 modes settings with remote control	<ul style="list-style-type: none">Time Setting: Manufacturer's Standard
ITEM O. 1 UNIT DENTAL UNIT MOTORIZED WITH ACESSORIES	
<ul style="list-style-type: none">Silent type and oil-free dental air compressor pressure: 72.5 psi (5 bar)	


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12 July 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 5th day of July, 2021 in MMCHD.

Approved by:

A handwritten signature in blue ink, appearing to read 'J. Chan', is written over the printed name.

JEREMIAS FRANCIS Y. CHAN, MD, MPH
BAC Chairperson



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 023E

Procurement of various Medical Equipments for GAT ANDRES BONIFACIO MEMORIAL MEDICAL CENTER Under the Health Facilities Enhancement Program (HFEP) of Metro Manila Center for Health Development (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Addition and clarification to provisions/specifications in the Bidding Documents:

- | |
|--|
| <ul style="list-style-type: none">• No changes as stated in the Technical Specifications |
|--|

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12 July 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 5th day of July, 2021 in MMCHD.

Approved by:

A handwritten signature in blue ink, appearing to read "J. Chan", is written over the printed name of the chairperson.

JEREMIAS FRANCIS Y. CHAN, MD, MPH
BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2021 – 024E

Procurement of various Medical Equipments for OSPITAL NG MUNTINLUPA Under the Health Facilities Enhancement Program (HFEP) of Metro Manila Center for Health Development (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Addition and clarification to provisions/specifications in the Bidding Documents:

- | |
|---|
| <ul style="list-style-type: none">• ULPA: Ultra Low Particulate Air |
|---|


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 12 July 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 5th day of July, 2021 in MMCHD.

Approved by:


JEREMIAS FRANCIS Y. CHAN, MD, MPH
BAC Chairperson