



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

**SUPPLEMENTAL/ BID BULLETIN NO. 1**

IB 2021 – 036E

**Procurement of Supply and Commissioning of Brand New Medical Equipment for the use of Health Facilities in City of Taguig Under the Health Facilities Enhancement Program (HFEP) of Metro Manila Center for Health Development**

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
<b>Item 1</b>	
<b>3 Unit MEDICINE CABINET</b>	
<ul style="list-style-type: none"> <li>• Stainless Steel</li> <li>• Dimension: Manufacturer’s Standard</li> </ul>	
<b>Item 2</b>	
<b>3 Unit INSTRUMENT CABINET</b>	
<ul style="list-style-type: none"> <li>• Stainless Steel</li> </ul>	
<b>Item 3</b>	
<b>3 Set HEMOGLOBINOMETER</b>	
<ul style="list-style-type: none"> <li>• Test tube size: Manufacturer’s Standard</li> <li>• Quantity: Manufacturer’s Standard</li> <li>• With Test Tube Rack included</li> </ul>	
<b>Item 4</b>	
<b>3 Unit WEIGHING SCALE WITH HEIGHT MEASURING STICK</b>	
<ul style="list-style-type: none"> <li>• WEIGHING SCALE WITH HEIGHT MEASURING STICK - For Adult</li> </ul>	
<b>Item 5</b>	
<b>3 Set GENERATOR SET, 10KVA</b>	
<ul style="list-style-type: none"> <li>• No changes as stated in the technical specifications</li> </ul>	
<b>Item 6</b>	
<b>3 Set DENTAL INSTRUMENT SET</b>	
<b>From</b>	<b>TO</b>
<ul style="list-style-type: none"> <li>• Perisoteal Elevator Stainless Steel</li> <li>• Bone Ronguer Stainless Steel</li> </ul>	<ul style="list-style-type: none"> <li>• Periosteal Elevator Stainless Steel</li> <li>• Bone Rongeur Stainless Steel</li> </ul>
<b>Item 7</b>	
<b>3 Unit EXAMINING TABLE</b>	
<ul style="list-style-type: none"> <li>• Table Type: Hamilton Type</li> <li>• Adjustable, can be reclined (bed, head part, leg part)</li> </ul>	
<b>Item 8</b>	
<b>3 Set DRESSING SET</b>	
<ul style="list-style-type: none"> <li>• Inclusion of Kidney Basin, Stainless Steel</li> </ul>	
<b>Item 9</b>	
<b>6 Unit FETAL DOPPLER</b>	
<ul style="list-style-type: none"> <li>• Technical Specifications:</li> <li>• Table top with trolley (durable)</li> </ul>	

<ul style="list-style-type: none"> <li>Standard Features:</li> <li>Display: Manufacturer's Standard</li> </ul>	
<b>From</b>	<b>To</b>
Standard Accessorires:	Standard Accessories:
<b>Item 10</b>	
<b>3 Set EENT DIAGNOSTIC SET</b>	
<ul style="list-style-type: none"> <li>Otoscope: different sizes (pedia, adult, neonate) for inclusion</li> </ul>	
<b>From</b>	<b>TO</b>
Ophthalmoscope	Ophthalmoscope
Post-natal mirror	Post-nasal mirror
<b>Item 11</b>	
<b>6 Set CERVICAL INSPECTION SET/VAGINAL SPECULUM SET</b>	
<b>From</b>	<b>To</b>
1 Needle holder long 7" or 8"	1 Needle holder long at minimum of 8"
<b>Item 12</b>	
<b>3 Unit INFANT WEIGHING SCALE, DIGITAL</b>	
<ul style="list-style-type: none"> <li>Medical Grade</li> </ul>	
<b>From</b>	<b>To</b>
Power Supply: 220 VAC, 60 Hz	Power Supply: Battery Operated or Direct Power Outlet Source 220VAC, 60 Hz, Single Phase
<b>Item 13</b>	
<b>3 Set AUTOCLAVE 20L</b>	
<b>From</b>	<b>To</b>
Power Supply: 220 V - 3 Phase, 60 Hz	Power supply: 220 V – Single Phase, 60 Hz
<b>Item 14</b>	
<b>3 Set DIFFERENTIAL COUNTER</b>	
<ul style="list-style-type: none"> <li>FUN = Function button</li> <li>Includes: additional 2 pcs of rechargeable AA batteries w/ compatible charger</li> </ul>	
<b>From</b>	<b>To</b>
Can operate on battery power (2 AA batteries)	Can operate on 2 AA rechargeable battery power
<b>Item 15</b>	
<b>6 Unit REAGENT REFRIGERATOR</b>	
<ul style="list-style-type: none"> <li>Reagent Refrigerator – Freezer</li> </ul>	
<b>Item 16</b>	
<b>3 Set CLINICAL CENTRIFUGE, 12 PLACER</b>	
<ul style="list-style-type: none"> <li>Adaptor compatible with Philippine Standard socket</li> </ul>	
<b>From</b>	<b>To</b>
Maximum speed: 3700 rpm	Maximum Speed: 3500 rpm to 5000 rpm
<b>Item 17</b>	
<b>6 Set MINOR SURGICAL SET OR CUT DOWN SET</b>	
<ul style="list-style-type: none"> <li>Surgical or cut down set – all instruments must be stainless steel</li> <li>1 pc Mayo scissors</li> </ul>	
<b>Item 18</b>	
<b>3 Set BED, MECHANICAL 2 CRANKS WITH MATTRESS</b>	
<ul style="list-style-type: none"> <li>2 cranks mechanism</li> <li>3 cranks mechanism is for bed height adjustment</li> <li>Bed specification - Manufacturer's Standard with knee raise</li> <li>With built-in retractable table or table tray with wheels for dining with overbed table included</li> </ul>	

<b>Item 19</b>	
<b>3 Unit INSTRUMENT TABLE</b>	
<ul style="list-style-type: none"> <li>Standard Features: With 4 caster (with lock) or better</li> </ul>	
<b>From</b>	<b>To</b>
Material: Stainless steel or equivalent	Material: Stainless steel
Stainless Steel Tray or equivalent	Stainless Steel Tray
<b>Item 20</b>	
<b>6 Unit MICROSCOPE BINOCULAR</b>	
<ul style="list-style-type: none"> <li>Stage has a single specimen holder with clips</li> <li>LED Bulb specifications:</li> <li>Provision for extra LED Bulb</li> </ul>	
<b>Item 21</b>	
<b>3 Set HEMATOCRIT CENTRIFUGE 12 PLACER</b>	
<ul style="list-style-type: none"> <li>With Automatic voltage regulator or Uninterrupted power supply: compatible with the equipment</li> </ul>	
<b>FROM</b>	<b>TO</b>
Speed Range 0 to 12,000 rpm	Speed Range: 0 to 15,000 rpm
Time Range: 0 to 60 minutes	Time Range: 0 to 60 minutes or higher

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 07 September 2021, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For guidance and information of all concerned

Issued this 1<sup>st</sup> day of September, 2021 in MMCHD.

Approved by:



**JEREMIAS FRANCIS Y. CHAN, MD, MPH**  
BAC Chairperson