



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 039E

PROCUREMENT OF VARIOUS BIOMEDICAL REFRIGERATOR (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
Item No. 1 – 21 UNITS BIOREFRIGERATOR Temperature Range: -60 ~ -86°C	Temperature Range: - 60°C ~ 20°C
Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 2 – 2 UNITS REAGENT REFRIGERATOR	
*No changes as stated in technical specifications.	

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 18th day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 9th day of April, 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III./ BAC Chairperson *ly*



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	BIOREFRIGERATOR	Qty./Unit	21 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 3,780,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> • Capacity: At least 340 L • Cabinet Type: Vertical • Cooling System: Direct Cooling • Temperature Range: - 60°C ~ 20°C • Humidity Range: 30~60% RH • Display: Digital, touch screen LCD or better • Material: Stainless Steel 304 • Insulation: Polyurethane (PU) foam • Power Supply: 220V, 50/60Hz • Dimensions: Manufacturer's Standard • Alarms: <ul style="list-style-type: none"> - High/Low Temperature - Power Failure - Abnormal Voltage - Sensor Error - Filter alarm - Low battery - Door ajar • Accessories: <ul style="list-style-type: none"> - Temperature recorder 			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Recipient: Health Facilities of Paranaque City		
<u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u> 1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.		



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 2	REAGENT REFRIGERATOR	Qty./Unit	2 Units	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 70,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
A. Technical Specifications: <ul style="list-style-type: none"> • Power Supply Input: 220V, 50/60Hz • Effective Capacity: at least 300 Liters • Polyurethane foam insulation • Hinged door, double-paned glass door • Compressor, at least 160 W • Stainless steel interior • Cyclical defrosting and evaporator temperature detection system • Digital display temperature • Temperature range of 2°C to 14°C • With fluorescent lamp • With 4 caster wheel swiveling with brake • Safety Features: <ul style="list-style-type: none"> - High and Low Temperature alarm system with audio-visual indicator - Temperature recorder - Door ajar alarm 				
B. DOCUMENTARY REQUIREMENTS: <ol style="list-style-type: none"> 1. Product brochure or technical data sheet(s) of the equipment showing the technical specifications in English Language. 				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>2. Valid and current Certificate of Compliance of manufacturer of the equipment with the latest version of ISO 13485: Quality Management System – Requirements for regulatory purposes in the name of the manufacturer. The Certificates must be issued by an independent Certifying Body/Agency.</p> <p>3. Notarized Certificate of the bidder:</p> <p>a. That the brand of the equipment has been in the local and/or international market for at least ten (10) years.</p> <p>b. That the equipment and its accessories are brand new, unused, not discontinued models and were not subjected to any product recall.</p> <p>4. Bidder’s valid and Current License to Operate (LTO) as a medical device distributor issued by the Philippine Food and Drug Administration. In case of expired LTO, the following must be submitted: i) Copy of expired LTO, ii) Application for renewal, iii) Official Receipt as proof of payment for the renewal of LTO.</p>		
<p><u>c. REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing</p>		

