



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 046E

### PROCUREMENT OF 1 UNIT BIPHASIC DEFIBRILLATOR

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
Display: At least 8 inches color display	Display: At least 7 inches color display
With event summary facility for recording (at least 55 events)	With event summary facility for recording (at least 55 events) or manufacturer's standard

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 18<sup>th</sup> day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 9<sup>th</sup> day of April, 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH  
Director III / BAC Chairperson *ly*



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. <b>1</b>	<b>BIPHASIC DEFIBRILLATOR</b>	Qty./Unit	<b>1 Unit</b>
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: <b>1,200,000.00</b>			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<b>A. Technical Specifications:</b>  <ul style="list-style-type: none"> <li>● Waveform: Biphasic</li> <li>● Capable to deliver shocks from 1 Joule to 300 Joules</li> <li>● Monitors ECG through paddles, pads and monitoring electrodes</li> <li>● Defibrillate through pads and paddles</li> <li>● With at least 50 mm strip printer</li> <li>● Charging time of less than 5 seconds for 200 Joules</li> <li>● Display: At least 7 inches color display</li> <li>● Monitoring Parameters               <ul style="list-style-type: none"> <li>- ECG</li> <li>- SpO2</li> <li>- NIBP</li> <li>- etCO2</li> </ul> </li> <li>● With adult and pediatric paddles</li> <li>● With event summary facility for recording (at least 55 events) or manufacturer's standard)</li> <li>● Battery usage for at least 5 hours monitoring</li> <li>● With facility for self-test/ check before usage</li> </ul>			



<p>and setup function</p> <ul style="list-style-type: none"> <li>● Power Supply Input: 220V, 60Hz</li> <li>● Accessories <ul style="list-style-type: none"> <li>- Patient cables</li> <li>- ECG rolls</li> <li>- Adult and Pedia SpO2 reusable sensor</li> <li>- Adult and Pedia NIBP cuff and hose</li> <li>- CO2 tubing</li> <li>- AED Multifunction Pads for adults</li> </ul> </li> </ul>		
<p><b><u>C. REQUIREMENTS IF AWARDED THE CONTRACT:</u></b></p> <p>1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.</p> <p>2; <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. <b>Training:</b> The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p>4. <b>Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall</p>		



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

<p>be added to the warranty period</p> <p><b>6. Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none"> <li>a) Service manual in English language</li> <li>b) Operation manual in English language</li> </ul> <p>7. With "<b>DOH-MMCHD HFEP</b>" (Government Property not for sale) sticker in each unit</p> <p>X-X</p> <p><b>Recipients:</b></p> <p>Ospital ng Paranaque I – 1 unit</p>		
<p><b><u>C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></b></p> <p>1. One (1) original sample of manufacturer’s product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>		