



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 048E

PROCUREMENT OF 1 SET ENDOSCOPY EQUIPMENT

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
Dimensions: Camera Console 31.8x10.02x38.7cm, Camera Head Cable to Console: 10.3 ft (3.15 m) sealed cable	Dimensions: Manufacturer's standard
Type of Light: Xenon	Type of Light: LED
Wattage: 300	Wattage: 100
Flow rate: 40 Liters	Flow rate: 30 Liters

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 18th day of April 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 9th day of April, 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS				
Item No. 1	ENDOSCOPY EQUIPMENT	Qty./Unit	1 Set	
Name of Manufacturer:			Country of Origin	
Brand:			Model: (if applicable)	
ABC: 2,000,000.00				
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE	
<p>A. Technical Specifications:</p> <ul style="list-style-type: none"> • A complete system for high-definition video endoscopic, laparoscopic and arthroscopic surgery • Endoscopic Video Camera <ul style="list-style-type: none"> - Dimensions: Manufacturer's standard - Enhancement: 16 levels (switchable) - Video Outputs: Digital/Analog: Two Digital Video Interface (DVI) / RGBHV, 1280 x 1024 (HD), 720p, 1080p, (HDTV) - Mounting: Eyepiece used with C-mount coupler; C mount camera head used with C-mount scopes • Light Source <ul style="list-style-type: none"> - Type of Light: LED - Wattage: 100 - Bulb life: At least 500 hours - Interface with visible LED Screen - Power Supply: 220V, 60Hz • Core Insufflator <ul style="list-style-type: none"> - Flow rate: 30 Liters - Power Supply: 220V, 60Hz - Dimensions: Manufacturer's Standard • Inclusions 				



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>- Video Monitor - Mobile cart</p>		
<p><u>B. REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2; Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment.</p> <p>4.Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period</p> <p>6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Service manual in English language b) Operation manual in English language</p> <p>7. With "DOH-MMCHD HFEP" (Government</p>		

