



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 132E

PROCUREMENT OF 1 UNIT BLOODBANK REFRIGERATOR

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Capacity: At least 500 liters, at least 260 bags (450ml/bag) AVR – 2kVA Data Storage: 32,000 data Report generation: Automatic PDF CSV report generation	
FROM	TO
Submission of Bids (Bid Opening): OCTOBER 19, 2022 AT 9:00 AM	Submission of Bids (Bid Opening): OCTOBER 24, 2022 AT 9:00 AM
<ul style="list-style-type: none"> • Safety - Safety door lock and controller setting password to prevent unauthorized access. - Built in back-up battery to display temperature and power alarm system without AC power. 	<ul style="list-style-type: none"> • Safety - Safety door lock - Built in back-up battery to display temperature and power alarm system without AC power. – removed
REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.	REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 – 60 calendar days upon receipt of Notice to Proceed.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 24th day of October 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 17th day of October 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson 4



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	BLOODBANK REFRIGERATOR	Qty./Unit	1 Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 750,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<p>TECHNICAL SPECIFICATIONS:</p> <ul style="list-style-type: none"> • Precise temperature control system <ul style="list-style-type: none"> - LED screen with 4°C (±1°C) adjustable platinum resistance temperature sensor • Safety <ul style="list-style-type: none"> - Safety door lock - Visual and audible alarm system: High/Low temperature, power failure, sensor error, door ajar. - Power failure protection: turn on delay of the cooling system after power failure. - Controller error protection: cooling system will remember the normal working cycle when controller or two sensor failed. • Refrigeration System <ul style="list-style-type: none"> - high quality compressor, cooling fan, dry filter and condensing system • Design <ul style="list-style-type: none"> - Adjustable shelves, firm glass doors, inner light, USB Temperature Data Logger <p>Capacity: At least 500 liters, at least 260 bags (450ml/bag)</p> <p>AVR – 2kVA</p> <p>Data Storage: 32,000 data</p> <p>Report generation:</p> <p>Automatic PDF CSV report generation</p>			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

**ADDITIONAL REQUIREMENT TO BE SUBMITTED BY THE
SINGLE/LOWEST CALCULATED BIDDER (SCB/LCB) AS PART OF
POST QUALIFICATION:**

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.