



## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 137E

### PROCUREMENT OF 1 UNIT PEDIATRIC DENTAL CHAIR WITH ACCESSORIES

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
<b>Submission of Bids (Bid Opening): OCTOBER 19, 2022 AT 9:00 AM</b>	<b>Submission of Bids (Bid Opening): OCTOBER 24, 2022 AT 9:00 AM</b>
• Special design for children	• Special design for children - retained
• All plastics are made of injection mold	• All plastics are made of injection mold - retained
• Normal operating light	• Normal operating light - retained
• Normal suction system	• Normal suction system - retained
- 2 hp Compressor with at least 25L capacity and 115 psi of max. pressure	2 hp Compressor with at least 25L capacity and 115 psi of max. pressure – silent type compressor
REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 calendar days upon receipt of Notice to Proceed	REQUIREMENTS IF AWARDED THE CONTRACT: 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 – 60 calendar days upon receipt of Notice to Proceed

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 24<sup>th</sup> day of October 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 17<sup>th</sup> day of October 2022 in MMCHD

Approved by:

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**

Director III / BAC Chairperson



**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	PEDIATRIC DENTAL CHAIR WITH ACCESSORIES	Qty./Unit	1 Unit
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 500,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<b>Technical Specifications:</b> <ul style="list-style-type: none"> <li>• Special design for children</li> <li>• All plastics are made of injection mold</li> <li>• Normal operating light</li> <li>• Normal suction system</li> <li>• Three-way syringe (One for hot and one for cold), 2 pcs</li> <li>• Built-in water purified system</li> <li>• Automatic water heating system</li> <li>• Built-in X-ray film viewer</li> <li>• With 2 pieces of dental stool (1 saddle type)</li> <li>• Inclusions               <ul style="list-style-type: none"> <li>- 1 unit of Intra Oral Camera with at least 17" LED monitor</li> <li>- 1 unit of Ultrasonic Scaler with LED (with 4 tips included)</li> <li>- 1 unit of Light Cure Machine</li> <li>- 3 pieces of Handpiece 4 holes or 2 holes</li> <li>- 2 hp Compressor with at least 25L capacity and 115 psi of max. pressure—silent type compressor</li> </ul> </li> </ul>			
<b>B. REQUIREMENTS IF AWARDED THE CONTRACT:</b> <ol style="list-style-type: none"> <li>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed with 30 – 60 calendar days upon receipt of Notice to Proceed</li> </ol>			



# METRO MANILA CENTER FOR HEALTH DEVELOPMENT

2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.

3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon the delivery of the equipment.

4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

7. With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit.

X-X

Source of Fund: SAA 2022-02-0531 (HFEP 2022)

Recipients:

Marikina City - 1 unit

**C. Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:**

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.