



## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 086

### PROCUREMENT OF 71 KITS QUALITATIVE RESPIRATOR TEST APPARATUS WITH REPLACEMENT SOLUTIONS

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
<b>Fit test kit:</b> Hood, Collar, 2 Nebulizers, 2 sets replacement for Nebulizer Inserts, Bitter Tasting Sensitivity Solution, Bitter Tasting Fit Test Solution, and Laminated User Instructions, FT-30	<b>Fit test kit:</b> Hood, Collar, 2 Nebulizers, 2 sets replacement for Nebulizer Inserts, 1 bottle Bitter Tasting Sensitivity Solution, 1 Bitter Tasting Fit Test Solution, and Laminated User Instructions, FT-30
<b>Replacement Solutions:</b> 1 case Bitter Tasting Sensitivity Solution (6 bottles per case), 1 case Bitter Tasting Fit Test Solution (6 bottles per case)	<b>Replacement Solutions:</b> 1 bottle Bitter Tasting Sensitivity Solution, 1 bottle Bitter Tasting Fit Test Solution
Delivery Period: Sixty (60) calendar days after receipt of notice to proceed	Delivery Period: Sixty to Ninety (60 to 90) calendar days after receipt of notice to proceed

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 13<sup>th</sup> day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 6<sup>th</sup> day of December 2022 in MMCHD

Approved by:

**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

<b>Republic of the Philippines</b> <b>Department of Health</b> <b>Metro Manila Center for Health Development</b>			
<b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	<b>PROCUREMENT OF 71 KITS QUALITATIVE RESPIRATOR TEST APPARATUS WITH REPLACEMENT SOLUTIONS</b>	Qty./Unit	71 KITS
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: P 1,988,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<u><b>SPECIFICATIONS:</b></u>  *Fit test kit for respirators  *Contents:  <b>Fit test kit:</b> Hood, Collar, 2 Nebulizers, 2 sets replacement for Nebulizer Inserts, 1 bottle Bitter Tasting Sensitivity Solution, 1 Bitter Tasting Fit Test Solution, and Laminated User Instructions, FT-30  <b>Replacement Solutions:</b> 1 bottle Bitter Tasting Sensitivity, 1 bottle Bitter Tasting Fit Test Solution  <b>Shelf Life:</b> Must be fresh from commercial stock with a total shelf life of 24 months from the date of manufacture but not less than 18 months from the date of delivery.  <b>Packaging Instructions:</b> Standard Packaging of the manufacturers as approved by			



the Philippine Food and Drug Authority (PFDA)

**Labelling Instructions:**

Primary or Standard labelling instructions as approved by the PFDA pursuant to Administrative Order No. 2016-0008.

**In addition to the labelling requirement of the PFDA:**

For the Secondary labelling: The following should be imprinted or stickered on each bottle using a permanent non-removable sticker/ label that is binding and will leave residue and ripping if removed:

**Philippine Government Property- Department of Health**

**NOT FOR SALE**

For the Tertiary labelling: The following should be imprinted or stickered on each small and bigger box/ corrugated carton, using a permanent non-removable sticker/ label that is binding and will leave residue and ripping if removed:

**Philippine Government Property- Department of Health**

**NOT FOR SALE**

Date of Manufacture:

Date of Expiry:

Batch/Lot No.:

**Delivery Period:** Sixty to Ninety (60 to 90) calendar days after receipt of notice to proceed

**Place of Delivery:** MMCHD Pasig Warehouse



**ADDITIONAL REQUIREMENT TO BE  
SUBMITTED BY THE SINGLE/LOWEST  
CALCULATED BIDDER (SCB/LCB) AS PART OF  
POST QUALIFICATION:**

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.