THE PHILIPS IN THE PRINT OF THE

Republic of the Philippines Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 116E PROCUREMENT OF 1 UNIT BIOSAFETY CABINET

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:						
FROM	ТО					
• Width: At least 4 feet	*Width: At least 4 feet – Retain					
• Front Window: Motorized, toughened glass or	*Front Window: Motorized or Manual,					
equivalent	toughened glass or equivalent					
• HEPA Filter: Two, 99.999% efficiency at	*HEPA Filter: Two, 99.995% efficiency at					
0.3um, with filter life indicator or better (ULPA	0.3um, with filter life indicator or better (ULPA					
filter)	filter)					
• Noise: ≤ 65dB	*Noise: ≤ 67dB					
• Tap: 1 Gas tap, 1 Water tap	*Tap: 1 Gas tap, 1 Water tap – Retain					
• LED Lamp: or its equivalent: At least 16W (1	*LED Lamp: or its equivalent: At least 14W -					
pc)	18W (1 pc) or Manufacturer Standard – Retain					
REQUIREMENTS IF AWARDED THE	REQUIREMENTS IF AWARDED THE					
CONTRACT:	CONTRACT:					
1. Completion Period: The delivery, installation,	1. Completion Period: The delivery, installation,					
testing and commissioning of the equipment and	testing and commissioning of the equipment and					
its accessories, including the training of end-	its accessories, including the training of end-					
users and maintenance staff must be completed	users and maintenance staff must be completed					
within 30 calendar days upon receipt of Notice to	within 30 to 60 calendar days upon receipt of					
Proceed.	Notice to Proceed.					

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 30^{th} day of August 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 22nd day of August 2022 in MMCHD

Approved by:

SGD.

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH

Director III / BAC Chairperson



Republic of the Philippines Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	BIOSAFETY CABINET	Qty./Unit	1 UNIT		
Name of Manufacturer:			Country of Origin		
Brand:	Model: (if applicable)				
ABC: 850,000.0	0				
	STATEMENT OF COMPLIANCE				
Technical Spec	ifications:				
• Class II A2 as p	oer NSF or Class II as per EN12469				
• Width: At least					
• External Size:	Manufacturer's Standard				
• Internal Size: N					
• Work Surface					
• Airflow Volum	e:				
I	Downflow - 65% - 70%				
1					
• Front Window	: Motorized, manual, toughened glass	or equivalent			
• HEPA Filter: To better (ULPA filt	wo, 99.995% efficiency at 0.3um, with	n filter life indicator or			
• Noise: ≤ 67dB					
• Illumination: ≥					
• Display: LCD D	isplay or better				
• Tap: 1 Gas tap,	1 Water tap				
• Power Supply:	220V, 60Hz				
• Alarms:					

THE PHILIPAGE STATES TO THE PH

Republic of the Philippines Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

- Abnormal airflow velocity
- filter replacement
- front window at unsafe height
- Motor: 220V, speed adjustable, high-efficiency and low power consumption
- Material:
- Work zone: Stainless Steel 304
- Main Body: Cold-rolled / Electrogalvanized steel with anti-bacterial powder coating or better
- UV Lamp: At least 30-40W
- LED Lamp: or its equivalent: At least 16W (1 pc)

REQUIREMENTS IF AWARDED THE CONTRACT:

- 1. **Completion Period:** The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 to 60 calendar days upon receipt of Notice to Proceed.
- **2. Testing:** Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.
- 3. **Training:** The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment
- 4. **Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.
- 5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.
- 6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:
 - a) Service manual in English language
 - b) Operation manual in English language
- 7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.

Recipient: Health Facilities of Muntinlupa City



Republic of the Philippines Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Additional	Requireme	ent to	be	submitted	by	the	Single/Lowest		
Calculated Bidder (SCB/LCB) as part of post qualification:									

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.