



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 120E

PROCUREMENT OF VARIOUS MEDICAL INSTRUMENTS (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
Item No. 1: 5 UNITS EENT DIAGNOSTIC SETS With hard casing (impact resistant)	With hard casing
Item No. 2: 6 SETS MINOR SURGICAL SET • Instrument tray with cover 18/8, stainless steel (410mm x 254mm x 64mm), 1pc • With Laser Mark "DOH-MMCHD" in each instrument	• Instrument tray with cover 18/8 (410mm x 254mm x 64mm), 1pc • Laser mark the word "DOH-MMCHD" in each instrument.
2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.	2. Prior to acceptance, the end user shall conduct a physical inspection. The equipment must have no physical damage and defect.
Item No. 3: 19 UNITS DRESSING SET • Surgical Scissor, Straight, stainless steel • Surgical Scissor, Curved, stainless steel • Bandage Scissor, stainless steel • Pick-up or Ovum Forcep, Stainless Steel • Mosquito Forcep, Stainless steel, 2pcs • Tissue Forceps with Teeth, Stainless Steel, 2pcs • Tissue Forceps without Teeth, Stainless Steel, 2pcs • Suture Removal Scissor, Stainless Steel	• Surgical Scissor, Straight, stainless steel, 5 ½", 1pc • Surgical Scissor, Curved, stainless steel, 5 ½", 1 pc • Bandage Scissor, stainless steel, 5 ½", 1 pc • Pick-up or Ovum Forcep, Stainless Steel, 7", 1pc • Mosquito Forcep, Stainless steel, 5", 1 pc • Tissue Forceps with Teeth, Stainless Steel, 5 ½", 2pcs • Tissue Forceps without Teeth, Stainless Steel, 5 ½", 2pcs • Suture Removal Scissor, Stainless Steel, 4 ¼", 1pc
2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.	2. Prior to acceptance, the end user shall conduct a physical inspection. The equipment must have no physical damage and defect.
Revision and clarification to provisions/specifications in the Bidding Documents:	
Add: • Kidney Basin, Stainless steel. 8", 1pc • Instrument Tray with Lid, Stainless steel, 12"x 8" x 2 ½", 1pc	
Item No. 4: 2 UNITS DELIVERY	



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

INSTRUMENT SET • 1 pc. Umbilican Scissors 4”	• 1 pc. Umbilical Scissors 4”
2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.	2. Testing: Prior to acceptance, the end user shall conduct a physical inspection. The equipment must have no physical damage and defect.
Item No. 5: 3 UNITS CERVICAL INSPECTION SET 2. Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.	2. Prior to acceptance, the end user shall conduct a physical inspection. The equipment must have no physical damage and defect.

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 30th day of August 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 22nd day of August 2022 in MMCHD

Approved by:

SGD.

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	EENT DIAGNOSTIC SETS	Qty./Unit	5 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 150,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> ● Otoscope with 2 levels of magnification ● Ophthalmoscope with different levels of magnification ● Stainless steel battery handles with brightness control ● Nasal speculum, stainless steel, pedia and adult ● Bent arm throat lamp ● Illuminated tongue depressor ● Post-nasal mirror 18 mm diameter, stainless steel ● Laryngeal mirror 22 mm diameter, stainless steel ● 4 pcs rechargeable battery with charger of 220 V 60 Hz, Li-ion or better ● Spare light bulb (LED) ● With hard casing 			
<p>REQUIREMENTS IF AWARDED THE CONTRACT:</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p>			



<p>3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>6- Manuale: Manuale: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <ul style="list-style-type: none">a) Service manual in English languageb) Operation manual in English language <p>7, With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit.</p> <p>X-X</p> <p>Recipient:</p> <p>Health Facilities in Muntinlupa City – 5 units</p>	
<p><u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 2	MINOR SURGICAL SET	Qty./Unit	6 Sets
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 120,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • Material: Stainless Steel • Instrument tray with cover 18/8, (410mm x 254mm x 64mm), 1pc • Mayo Scissors 5- 3/4", straight (14.5 cm), 1pc • Mosquito hemostatic forceps straight 5" (12.5 cm) 20mm jaw with full serrations 2pcs • Mosquito hemostatic forceps straight 4- 3/4", (12 cm) 20mm jaw with full serrations 2pcs • Needle holder, bulldog jaw, 6- 1/4', (16 cm), 1 pc • Surgical knife handle #3 (scalpel) 1 pc • Surgical knife handle #4 (scalpel) 1 pc • Tissue forceps 1 x 2 teeth, 5- 3/4" (14.5 cm) 2 pcs • Kidney Basin 8", 1 pc • Bandage scissors, 5 1/2", 1 pc • Laser mark the word "DOH-MMCHD" in each instrument. 			
<p>REQUIREMENTS IF AWARDED THE CONTRACT:</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Prior to acceptance, the end user shall conduct a physical inspection. The equipment must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for one (1) on craftsmanship. Attaching also the notarized undertaking that supplier shall either repair or replace any part in the equipment that is found to be defective in material or in workmanship under normal use within five (5) calendar days upon notification of the equipment breakdown from the end user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship shall be added to the warranty period. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning</p>			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

4. **Manuals:** Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

a) Sterilization and cleaning manual in English language

X-X

Recipient:

Health Facilities in Muntinlupa City – 6 units

Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 3	DRESSING SET	Qty./Unit	19 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 380,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • Surgical Scissor, Straight, stainless steel, 5 ½", 1pc • Surgical Scissor, Curved, stainless steel, 5 ½", 1 pc • Bandage Scissor, stainless steel, 5 ½", 1 pc • Pick-up or Ovum Forcep, Stainless Steel, 7", 1pc • Mosquito Forcep, Stainless steel, 5", 1 pc • Tissue Forceps with Teeth, Stainless Steel, 5 ½", 2pcs • Tissue Forceps without Teeth, Stainless Steel, 5 ½", 2pcs • Suture Removal Scissor, Stainless Steel, 4 ¼", 1pc • Kidney Basin, Stainless steel. 8", 1pc • Instrument Tray with Lid, Stainless steel, 12"x 8" x 2 ½", 1pc • Laser mark the word "DOH-MMCHD" in each instrument 			
<p>Documentary Requirements:</p> <ol style="list-style-type: none"> 1. Product brochure or technical data sheet(s) of the equipment showing the technical specifications in English Language. 2. Valid and current Certificate of Compliance of manufacturer of the equipment with the latest version of ISO 13485 or its equivalent. The Certificates must be issued by a independent Certifying Body/Agency. 3. Notarized Certificate of the bidder: <ol style="list-style-type: none"> a. That the brand of the equipment has been in the local and/or international market for at 			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>least ten (10) years.</p> <p>b. That the equipment and its accessories are brand new, unused, not discontinued models and were not subjected to any product recall.</p> <p>4. Bidder's and valid and current License to Operate (LTO) as a medical device Distributor issued by the Philippine Food and Drug Administration In case of expired LTO, the following must be submitted: 1) Copy of expired LTO, ii) application for renewal, iii) Official Receipt as proof of payment for the renewal of LTO.</p>	
<p><u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Prior to acceptance, the end user shall conduct a physical inspection. The equipment must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for one (1) on craftsmanship. the supplier shall either repair or replace any part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning</p> <p>4. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>5. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p style="padding-left: 40px;">a) Service manual in English language</p> <p>X-X</p> <p>Recipient:</p> <p>Marikina – 8 units</p> <p>Muntinlupa – 11 units</p>	
<p><u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 4	DELIVERY INSTRUMENT SET	Qty./Unit	2 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 120,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • 6 pcs. Sponge Holding Forcep 9 ½" straight • 6 pcs. Backhaus Towel Clamp 5 ¼" • 1 pc. Operating Scissors S/B 5 ½" straight • 1 pc. Braun Episiotomy Scissor 8 ½", Angled to side • 1 pc. Umbilical Scissors 4" • 2 pcs. Thumb Forceps 5 ½" Serrated • 2 pcs. Tissue Forceps (1x2T) 5 ½" • 6 pcs. Allis Tissue Forceps (5x6T) 6" • 6 pcs. Crile Forceps 5 ½" Straight • 1 pc Bozeman Uterine Dressing Forceps 10 ¼" (Double Curve) • 2 pcs Schroeder Braun Uterine Tenaculum 10 ¼" straight • 1 pc Lowa Membrane Puncturing Forceps 10 ¼" (Double Curve) • 1 pc Simpson Obstetrical Forceps 4" • 1 pc Mclean-Tucker-Luikart Obstetrical Forceps 15 ½" • 1 pc Kielland Obstetrical Forceps 17 ½" • 1 pc Piper Obstetrical Forceps 17 ½" • 1 pc Bill Traction Handle (for use with any Obstetrical Forceps) • 1 pc Martin Pelvimeter (Graduated in cm and inches) • 1 pc Mayo-Hegar Needle Holder 6" • 1 pc Mayo-Hegar Needle Holder 7" 			



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines

Department of Health

Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 5	CERVICAL INSPECTION SET	Qty./Unit	3 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 60,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> ● Material: Stainless Steel ● Vaginal Speculum (1pc S, 2pcs M, 1 pc L, size) ● 3 Ring Forceps – Long 3 pcs ● Vaginal Retractor, medium 2pcs ● Needle holder long 8", 1 pc ● Tissue Forceps Long 9", 1 pc ● Kidney Basin 8", 1pc ● Pick-up or Ovum Forcep, 7", 1 pc ● Instrument Tray with Lid, 12" x 8' x 2 ½", 1pc ● Laser Mark the word "DOH - MMCHD" in each instrument. 			
<p><u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Prior to acceptance, the end user shall conduct a physical inspection. The equipment must have no physical damage and defect.</p> <p>3. Warranty: Warranty certificate for one (1) on craftsmanship. Attaching also the notarized undertaking that supplier shall either repair or replace any part in the equipment that is found to be defective in material or in workmanship under normal use within</p>			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

five (5) calendar days upon notification of the equipment breakdown from the end user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship shall be added to the warranty period. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning

4. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

a) Sterilization and Cleaning manual in English language

X-X

Recipient: Health Facilities in Muntinlupa City – 3 sets

Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.