



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 122E

PROCUREMENT OF VARIOUS MEDICAL EQUIPMENT (LOT BIDDING)

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
Item No. 1: 5 UNITS FETAL DOPPLER *No changes as stipulated in technical specifications.	
Item No. 2: 2 UNITS FETAL MONITOR 15 roll thermal paper	
Item No. 3: 4 UNITS PULSE OXIMETER WITH ADULT AND PEDIA PROBE Spare Probe 1 for Adult 1 for Pedia	
FROM	TO
Item No. 2: 2 UNITS FETAL MONITOR up to 150 patient data saving	up to 150 patient data saving - removed
Item No. 3: 4 UNITS PULSE OXIMETER WITH ADULT AND PEDIA PROBE Operation: At least 4 hours	Operation: At least 4 hours - Retain
Charging Time: 300 min	Charging Time: 300 min - Retain
3. Training: The supplier shall provide a training/Demonstration on the proper use maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 day upon delivery of the equipment	3. Training: The supplier shall provide a training on the proper use maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 day upon delivery of the equipment

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 30th day of August 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 22nd day of August 2022 in MMCHD

Approved by:

SGD.

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH

Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	FETAL DOPPLER	Qty./Unit	5 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 225,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
Technical Specifications: <ul style="list-style-type: none"> ● Power Supply: Dual Power Operation (220VAC, 60 Hz and Battery should be rechargeable) ● Type: Table Top (heavy duty) ● Material: Made of hard plastic/ ABS Plastic ● Standard Features: <ul style="list-style-type: none"> - LCD Display - Fetal Heart Rate - Built-in loudspeaker: with noise reduction - Probe holder ● Standard Accessories: <ul style="list-style-type: none"> - 1 conductive gel - 1 ultrasonic probe 			
<u>REQUIREMENTS IF AWARDED THE CONTRACT:</u> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training/Demonstration: The supplier shall provide a</p>			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

training on the proper use maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 day upon delivery of the equipment.

4. **Warranty:** Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any item or part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.

5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service manual in English language
- b) Operation manual in English language

7. With "**DOH-MMCHD HFEP**" (Government Property not for sale) sticker in each unit.

x-x

Recipient:

Health Facilities in Muntinlupa City – 5 units

Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 2	FETAL MONITOR	Qty./Unit	2 Units
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 320,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
Technical Specifications: <ul style="list-style-type: none"> • Dual pulsed Doppler FHR • Ultrasound Frequency: 1 MHz • Range: 30-240 bpm, 50-210 bpm • With auto-detection of dual fetal movement • With array-type thermal printer • With automatic printing function • Approx. 7-inch. TFT color LCD Display • 3-channel display: FHR I, FHR II, UC • With Clinical event marking • Dimensions: Manufacturer's Standard • Can be connected to a central monitoring system through PC interface • Accessories <ul style="list-style-type: none"> - 2 units Ultrasound Doppler probe - 1 unit event marker jack - 3 units probe belt - 1 unit UC probe - 1 unit ultrasound gel - 2 units probe holder - Battery * 15 roll thermal paper 			
<u>REQUIREMENTS IF AWARDED THE CONTRACT:</u>			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p> <p>2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.</p> <p>3. Training: The supplier shall provide a training/Demonstration on the proper use maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 day upon delivery of the equipment</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall either repair or replace any part in the equipment that is found to be defective in material or in workmanship under normal use The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period</p> <p>6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:</p> <p>a) Service manual in English language</p> <p>b) Operation manual in English language</p> <p>7. With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit.</p> <p>X-X</p> <p>Recipient:</p> <p>Health Facilities in Muntinlupa City – 2 units</p>	
<p><u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer’s product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 3	PULSE OXIMETER WITH ADULT AND PEDIA PROBE	Qty./Unit	4 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 280,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> Dimensions: Manufacturer's Standard Type: Table top Power Supply: Dual Power Operation (220V, 60Hz or Auto-volt and Battery should be rechargeable) Battery Type: Manufacturer's Standard Built-in rechargeable battery Operation: At least 4 hours Charging Time: 300 min Protection Class: BF, Class II Conformity: MDD 93/42/EEC, CE0470 or equivalent Enhanced SpO2 Accuracy during motion and low perfusion Compact side configuration or with handle 72 hours data memory Adult and pediatric application Display: LED or better Audible and visible alarm Spare Probe 1 for Adult 1 for Pedia 			
REQUIREMENTS IF AWARDED THE CONTRACT: <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use maintenance of the equipment to the end-users and to the hospital 			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

<p>maintenance staff within 3 day upon delivery of the equipment</p> <p>4. Warranty: Warranty certificate for two (2) years on parts and service. the supplier shall either repair or replace any item part in the equipment that is found to be defective in material or in workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning</p> <p>5. Notarized undertaking that the supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment is unusable due to defective material or workmanship, shall be added to the warranty period.</p> <p>6. Manuals: The supplier must provide the end-user one (1) hard and one (1) soft.</p> <p>a) Service manual in English language</p> <p>b) Operation manual in English language</p> <p>7. With "DOH-MMCHD HFEP" (Government Property not for sale) sticker in each unit.</p> <p>X-X</p> <p>Recipient:</p> <p>Health Facilities in Muntinlupa City – 4 units</p>	
<p><u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></p> <p>1. One (1) original sample of manufacturer’s product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	