



Republic of the Philippines  
Department of Health

## METRO MANILA CENTER FOR HEALTH DEVELOPMENT

### SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 125E

### PROCUREMENT OF 32 UNITS PORTABLE UVC LIGHT STERILIZER

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

<b>Revision and clarification to provisions/specifications in the Bidding Documents:</b>	
<b>FROM</b>	<b>TO</b>
<b>Submission of Bids: SEPTEMBER 22, 2022</b>	<b>Submission of Bids: SEPTEMBER 21, 2022</b>
Power: At least 38W	Power: At least 32W


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 21<sup>st</sup> day of September 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 14<sup>th</sup> day of September 2022 in MMCHD

Approved by:

  
**ALELI ANNIE GRACE P. SUDIACAL, MD, MPH**  
Director III / BAC Chairperson 8



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

Republic of the Philippines Department of Health Metro Manila Center for Health Development  <b>TECHNICAL SPECIFICATIONS</b>			
Item No. 1	<b>PORTABLE UVC LIGHT STERILIZER</b>	Qty./Unit	<b>32 Units</b>
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: <b>480,000.00</b>			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
<b>Technical Specifications:</b>  Type: Lamp Power: At least 32W Voltage: 220V/60Hz or customized Lamp Lifespan: 8,000 to 12,000 hours Time Setting: 3-mode settings with remote control Dimensions: Manufacturer's Standard			
<b><u>REQUIREMENTS IF AWARDED THE CONTRACT:</u></b>  1. <b>Completion Period:</b> The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within <b>30</b> calendar days upon receipt of Notice to Proceed.  2. <b>Testing:</b> Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect.  3. <b>Warranty:</b> Warranty certificate for two (2) years on parts and service. The supplier shall repair or replace any part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning.  4. Notarized undertaking that supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the number of days where the equipment due to defective material or workmanship, shall be added to the warranty period.  5. <b>Manuals:</b> The supplier must provide the end-user one (1) hard and one (1) soft			



Republic of the Philippines  
Department of Health  
**METRO MANILA CENTER FOR HEALTH DEVELOPMENT**

<p>copy of the following:</p> <ul style="list-style-type: none"><li>a) Service Manual in English language</li><li>b) Operation manual in English language</li></ul> <p>6. With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit.</p> <p>x-x</p> <p><b>Recipient:</b></p> <p>Ospital ng Malabon – 2 units Ospital ng Tondo – 10 units Muntinlupa City – 16 units Marikina City - 4 units</p>	
<p><b><u>Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:</u></b></p> <p>1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.</p>	