



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 126E PROCUREMENT OF 7 UNITS ECG MACHINE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
FROM	TO
Submission of Bids: SEPTEMBER 22, 2022	Submission of Bids: SEPTEMBER 21, 2022
With graphic LCD to display 1 channel ECG waveform, setting menu, and operating conditions of the unit	With graphic LCD to display 3 channel ECG waveform, setting menu, and operating conditions of the unit


Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 21st day of September 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 14th day of September 2022 in MMCHD

Approved by:


ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson



Republic of the Philippines
Department of Health
METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines Department of Health Metro Manila Center for Health Development TECHNICAL SPECIFICATIONS			
Item No. 1	ECG MACHINE	Qty./Unit	7 Units
Name of Manufacturer:			Country of Origin
Brand:			Model: (if applicable)
ABC: 490,000.00			
PURCHASER'S SPECIFICATION			STATEMENT OF COMPLIANCE
Technical Specifications: <ul style="list-style-type: none"> • At least 80mm x 20 m paper size, roll type • Leads: Standard 12 leads • Automatic measurement & interpretation of standard ECG parameter • Manual/Auto/Arrhythmia Analysis mode or equivalent • Rhythm lead with histogram & trend graph for RR interval or equivalent • With graphic LCD to display 3 channel ECG waveform, setting menu, and operating conditions of the unit • Dual operation, AC & Li-ion battery supporting 2 hours of operation 			
<u>REQUIREMENTS IF AWARDED THE CONTRACT:</u> <ol style="list-style-type: none"> 1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed. 2. Testing: Prior to acceptance, the end user shall conduct a physical inspection and functionality test. The equipment must be functioning and must have no physical damage and defect. 3. Training: The supplier shall provide a training on the proper use and maintenance of the equipment to the end-users and to the hospital maintenance staff within 3 days upon delivery of the equipment. 4. Warranty: Warranty certificate for two (2) years on parts and service. The supplier shall repair or replace any part in the equipment that is found to be defective in material or workmanship under normal use. The warranty period shall commence from the date of acceptance by the end-user after testing and commissioning. 5. Notarized undertaking that supplier shall conduct the necessary corrective maintenance within five (5) calendar days upon notification of the equipment breakdown from the end-user. The undertaking shall include a statement that the 			



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

number of days where the equipment due to defective material or workmanship, shall be added to the warranty period.

6. **Manuals:** The supplier must provide the end-user one (1) hard and one (1) soft copy of the following:

- a) Service Manual in English language
- b) Operation manual in English language

7. With "DOH-MMCHD HFEP"(Government Property not for sale) sticker in each unit.

x-x

Recipient:

- Health Facilities in Muntinlupa City – 3 units
- Health Facilities in Paranaque City – 2 units
- Ospital ng Paranaque I – 1 unit
- Ospital ng Paranaque II – 1 unit

Additional Requirement to be submitted by the Single/Lowest Calculated Bidder (SCB/LCB) as part of post qualification:

1. One (1) original sample of manufacturer's product to be submitted and returned after evaluation. The sample submitted and approved during the evaluation shall be the same item to be delivered upon award of contract. Prototype of the labelling instruction must be part of the sample submitted however, the technical specifications of the labelling instruction of the product must be complied upon delivery.