



Republic of the Philippines
Department of Health

METRO MANILA CENTER FOR HEALTH DEVELOPMENT

SUPPLEMENTAL/ BID BULLETIN NO. 1

IB 2022 – 151E PROCUREMENT OF 1 UNIT ABG MACHINE

This Supplemental/Bid Bulletin No. 1 is being issued to revise provisions/specifications in the Bidding Documents for a forecited project:

Revision and clarification to provisions/specifications in the Bidding Documents:	
*No changes as stipulated in technical specifications.	
FROM	TO

Bidders are advised to use the following attached forms and submit together with all required documents for the submission of bids on 5th day of December 2022, 9:00 AM:

This Supplemental/Bid Bulletin No. 1 shall form an integral part of the Bidding Documents. All other provisions indicated in the bidding documents which are not affected by this Supplemental/Bid Bulletin No. 1 shall remain in effect.

For guidance and information of all concerned.

Issued this 25th day of November 2022 in MMCHD

Approved by:

ALELI ANNIE GRACE P. SUDIACAL, MD, MPH
Director III / BAC Chairperson *g*



METRO MANILA CENTER FOR HEALTH DEVELOPMENT

Republic of the Philippines
Department of Health
Metro Manila Center for Health Development

TECHNICAL SPECIFICATIONS

Item No. 1	ABG MACHINE	Qty./Unit	1 Unit
Name of Manufacturer:		Country of Origin	
Brand:		Model: (if applicable)	
ABC: 1,450,000.00			
PURCHASER'S SPECIFICATION		STATEMENT OF COMPLIANCE	
<p>Technical Specifications:</p> <ul style="list-style-type: none"> • Essential Measured parameters: pH, pCO₂, pO₂, tHb, Barometric Pressure, Na⁺, K⁺, Ca⁺⁺, Cl⁻. All these parameters should be measured simultaneously • Calculated parameters should include BE, BE ecf, HCO₃, Lactate, Anion Gap, SaO₂ etc • Sample volume-less than 100µl • Fast analysis time - less than 60 sec • Fully automatic liquid calibration of all parameters at user-defined intervals without the use of Gas calibrated reagents, external gases, tanks or regulators • Continuous reagent level monitoring with graphic display • Data display on well-illuminated, adequate size LCD color touch screen display • Data print out on built in graphic printer • Built in Auto Quality Control facility • Automatic result processing and test ordering • Automatic data archiving and customizable layout. Data backup with read/write CDROM drive 			
<p>REQUIREMENTS, if awarded the contract</p> <p>1. Completion Period: The delivery, installation, testing and commissioning of the equipment and its accessories, including the training of end-users and maintenance staff must be completed within 30 calendar days upon receipt of Notice to Proceed.</p>			

